

Certificate of Analysis

96-Well Region of Interest (ROI) and Background Plates

Product No. **4432364**
Lot No. **2509140**
Date of Manufacture **01Oct2025**
Expiration Date **26Sep2026**

TEST	SPECIFICATION	RESULT
Material Test		
Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
FAM™ Dye	518 ± 4 nm	518 nm
VIC™ Dye	548 ± 4 nm	549 nm
NED™ Dye	570 ± 4 nm	571 nm
ROX™ Dye	601 ± 5 nm	602 nm
LIZ™ Dye	649 ± 5 nm	650 nm

For Research Use Only. Not for use in diagnostic procedures.

**ISO13485
REGISTERED**

Singapore

RICKY LEONG

06 OCT 2025

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-06,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at
cofarequest@appliedbiosystems.com