

ANALYSIS CERTIFICATE

5-203-C4-0 AHN myTip LFT 100-1000 µl, clear, sterile
Lot-No.: P-192482
Expiry Date: 05/2022

Properties*

Tests	Test Method	Unit	Typical Value
MFR (2.16 kg/230°C), melt flow rate	ISO 1133	g/10 min.	40 - 60
Tensile modulus	ISO 527/2	MPa	1300 - 1900
Vicat softening temperature VST/A50	ISO 306	°C	130 - 160
Density (23°C)	ISO 1183	g/cm ³	0.9 – 0,95

*The property values shown are based on a limited number of tests.

Storage and Handling:

The product should be stored in dry conditions at temperature below 35°C and protected from UV-light. Improper storage can initiate degradation, which results in colour changes and can have negative effects on the physical properties of the products.

Safety:

AHN Biotechnologie GmbH Tips/Filter Tips are not classified as dangerous preparation.

Recycling:

The products are suitable for recycling using modern methods of shredding and cleaning.

AHN Biotechnologie GmbH

Quality Control Department

Uthleber Weg 14

D-99734 Nordhausen

Nordhausen, 2019-05-09

This document was sent online and is valid without signature