

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : SD216	Material Name : Levofloxacin LE 5 mcg	Lot No : 0000417190
AR No.: 040000982032	Date of Report : 2019-12-31	Exp. Date : 2021-11

TEST	SPECIFICATIONS	RESULTS
<u>Appearance</u> 1 Appearance	Filter paper discs of 6 mm diameter with printed "LE 5" on centre of each side of the disc.	Complies
<u>Cultural response</u> 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
<u>Escherichia coli ATCC 25922</u> 1 Zone of inhibition (mm)	29mm - 37mm	36mm
<u>Staphylococcus aureus ATCC 25923</u> 1 Zone of inhibition (mm)	25mm - 30mm	29mm
<u>Pseudomonas aeruginosa ATCC 27853</u> 1 Zone of inhibition (mm)	19mm - 26mm	25mm
<u>Staphylococcus aureus ATCC 29213</u> 1 Zone of inhibition (mm)*	23mm - 29mm	27mm
<u>Enterococcus faecalis ATCC 29212</u> 1 Zone of inhibition (mm)*	19mm - 25mm	23mm

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets.

HiMedia Laboratories Private Limited



23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Material Code : SD216	Material Name : Levofloxacin LE 5 mcg	Lot No : 0000417190
AR No.: 040000982032	Date of Report : 2019-12-31	Exp. Date : 2021-11

TEST	SPECIFICATIONS	RESULTS
------	----------------	---------

This document has been produced electronically and is valid


Sukanya Deshmukh**Microbiologist/Sr.Executive Microbiologist**
Ujjwala M. Kokate**Asst./Dy/QC Manager**
Dr. Santosh Kaul**Dy/QA Manager**