

Certificate of Analysis

Issue date	2025-11-04
Product	virotype® CSFV 2.0 RT-PCR Kit
Cat. Number(s)	VT281825
Lot number	F202500156E
Expiry date	2026-06-30, extended to 2026-12-31 (yyyy-mm-dd)
Storage temperature	-30°C to -15°C

Quality Control

Kit Component	Acceptance Criteria		CSFV	Internal Control	Result	Passed
	Ct Range FAM	Ct Range HEX	Ct FAM	Ct HEX		
Negative Control	No Ct	No Ct	No Ct	No Ct	negative	✓
Positive Control	25 - 30	25 - 30	27.55	27.96	positive	✓

Sample	Sample Type	CSFV Status	CSFV	Internal Control	Result	Passed
			Ct FAM	Ct HEX		
CSF 1	blood 1:1000	pos.	21.76	n.a.	positive	✓
CSF 2	blood	neg.	No Ct	27.19	negative	✓
CSF 1047	tonsil	pos.	21.88	29.91	positive	✓

n.a. = not applicable

Batch release thermal cycler: BioRad CF X96 (Bio-Rad Laboratories, Inc.)

If the new batch lot number ends with E, E indicates a prolongation of the expiry date.

Manufactured and tested at INDICAL BIOSCIENCE GmbH, Leipzig, Germany, and in compliance with our ISO 9001 certified quality management system. The product fulfills all INDICAL quality requirements.

Kristin Wappler, Quality Assurance

This certificate is a computer printout and therefore valid without signature.