

**EC Certificate**  
**Directive 93/42/EEC Annex II, excluding Section 4**  
**Full Quality Assurance System**  
**Medical Devices**

**Registration No.:** HD 60135869 0001

**Report No.:** 15096340 002

**Manufacturer:** Zhengyuan Technology Co., Ltd.  
No. 9 East Fengsi Road, Fengjing  
Industrial Park, Huyi District  
Xi'an  
710300 Shaanxi  
P.R. China

**Products:**

- Plasma Separators for Single Use
- Centrifugal Blood Processing Bowls for Single Use
- Apheresis Tubing Sets for Single Use
- Apheresis Needle Assemblies for Single Use
- Plasma Component Bags for Single Use
- Plasma Collection Bottles for Single Use

**Expiry Date:** 2024-01-24

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

**Effective Date:** 2020-01-14

**Date:** 2020-01-14



**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**  
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.