



# EU DECLARATION OF CONFORMITY

<b>Responsible Manufacturer:</b>	Laerdal Medical AS P.O. Box 377 Tanke Svilandsgate 30 4002 Stavanger Norway
<b>Single Registration Number:</b>	NO-MF-000000012
<b>Manufacturing site:</b>	Laerdal Medical (Suzhou) Co., Ltd. Building 18,19,20, No. 57 Huoju Road Science & Technology Industrial Park Suzhou, Jiangsu Province 215009 China
<b>Product Name:</b>	<b>Stifneck Select Extrication Collar</b> <b>Stifneck Pedi-Select Extrication Collar</b>
<b>Basic UDI-DI:</b>	0704543209932SV
<b>Intended Purpose:</b>	Intended to support cervical spine in a neutral position during transportation, either alone or in combination with other cervical and full body immobilization devices (CSIDs and FBSIDs).
<b>Product Options:</b>	980010 Stifneck Select Extrication Collar (US/AU) 98001005 Stifneck Select Extrication Collar(J) 98001033 Stifneck Select Extrication Collar 980012 Stifneck Select Green Extrication Collar (US/AU) 98001205 Stifneck Select Green Extrication Collar (J) 98001233 Stifneck Select Green Extrication Collar 980020 Stifneck Pedi-Select Extrication Collar (US/AU) 98002005 Stifneck Pedi-Select Extrication Collar(J) 98002033 Stifneck Pedi-Select Extrication Collar

to which this declaration relates is in conformity with the General Safety and Performance Requirements of EU Regulation 2017/745.

**Classification:**

The Stifneck Select Extrication Collars are class I according to rule 1 of Annex VIII of the EU Medical Device Regulation.

Conformity Assessment is based on the principles described in Article 52 of Regulation 2017/745.

**Conformity is declared in relation to common Specification(s):**

*No CS for Extrication Collars has been published at this time*

This EU Declaration of Conformity is issued under the sole responsibility of Laerdal Medical AS.

Adrienne Farrington  
Regulatory Affairs Specialist  
On behalf of CEO Alf Christian Dybdahl  
17<sup>th</sup> July 2024  
Stavanger Norway

DocuSigned by:  
*Adrienne Farrington*  
F5B74E6700DF43D...