

Commitment Letter


To Whom It May Concern

We, “TEK-BAL Tıbbi Cihazlar Sanayi ve Ticaret Ltd. Şti” having factories at “Maltepe Mah. Dokumacılar Sitesi No:36/A Zeytinburnu / İstanbul / Türkiye”, hereby confirm that the offered product “STR-23” fully complies with the following parameters:

1. Sterilization time accuracy: 1 min;
2. Temperature range: 105 - 134 °C;
3. Data recorder capabilities:
 - Date
 - Time
 - Temperature
 - Pressure

Name and Position

MEHMET YASAR – GENERAL MANAGER

 **TEK-BAL TIBBİ CİHAZLAR
SAN. TİC. LTD. ŞTİ.**
Maltepe Mah. Dokumacılar Sk. Sok.No:36/A
Zeytinburnu / İSTANBUL
Davutpaşa V.D:8330068497 Tic.Sicil: 339445-0
Mersis No.: 0833006849700019

06.08.2024