

PRODUCT SPECIFICATION

Email: info@whyasap.com

Product Series: ASAP Latex Powder Free Examination Gloves

Website: www.whyasap.com

Material : Natural Rubber Latex

Features : Latex Examination Glove, Powder Free, Non-Sterile, Ambidextrous, Palm / Finger Textured,

Beaded Cuff, Single Use Only

Color : Natural

Available Sizes : Extra Small, Small, Medium, Large, Extra Large
Packing Mode : 100 Pieces Per Box, 10 Boxes Per Carton

Shelf Life : 3 Years

Table 1: Tests and Specifications

Test	Particular		Specification	Standard Compliance
Dimension (mm)	Palm Width	Extra Small	75 ± 5	EN 455 Part 2
		Small	85 ± 5	
		Medium	95 ± 5	
		Large	105 ± 5	
		Extra Large	115 ± 5	
	Length		Min. 240	
	Thickness (Single wall)	Cuff (25 ± 5 from bead)*	0.08 ± 0.02	In-house
		Palm (center of palm)	0.10 ± 0.02	In-house
		Finger (13 ± 3 from tip)	0.12 ± 0.02	In-house
Physical	Force at Break (N)	Before Ageing	Min. 6	EN 455 Part 2
Properties*		After Ageing	Min. 6	
Powder Residue*	Powder free (mg/glove)		Max. 2	EN 455 Part 3
Physical Appearance	Beaded Cuff Palm / Finger Textured Ambidextrous			

^{*}Remark: Results applicable upon request

Tel: +6 03 5191 0166 Fax: +6 03 5191 0702 Website: www.whyasap.com Email: info@whyasap.com

Table 2: Inspection Level and Acceptance Quality Level (AQL)

Performance Requirements	Particular	Inspection Level	AQL	Standard Compliance
Freedom from Holes	Holes	G1	1.5	EN 455 Part 1
	Critical	G1	1.5	In-house
Visual	Major	G1	2.5	In-house
	Minor	G1	4.0	In-house
Dimension	Palm Width, Length, Thickness	S2	4.0	In-house
Physical Properties	Force at Break	Median Value of N=13		EN 455 Part 2
Powder Residue	Powder Free	N=5	N/A	EN 455 Part 3
Powder Content	Powdered	N=2	N/A	EN 455 Part 3

Handling and Storage

Store in a cool, dry and dust free facility or at room temperature. Exposed product should be shielded from humidity, direct sunlight, and fluorescent lighting.

Caution

This product contains substance which may cause an allergic reaction in some individuals. Do not expose these gloves to any person with a known or suspected sensitivity to latex. If allergic reaction occurs, discontinue use immediately and consult a physician.

Doc. No.: PS AS2802-E Rev. Date: 05.11.2024

Rev. No.: 0

