

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

Model / Type Ref.
Ref. De type

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

CERTIFICAT D'ESSAI OC

Medical LCD Monitor

D & T INC
(JANG-DONG, (DAEDEOK VALLEY))
26-121 GAJEONGBUK-RO
YUSEONG-GU
DAEJEON, 305-343 Korea

FORESEESON CUSTOM DISPLAYS INC
2210 E. WINSTON RD. ANAHEIM, CA 92806 USA

D & T INC
(JANG-DONG, (DAEDEOK VALLEY)) 26-121
GAJEONGBUK-RO YUSEONG-GU DAEJEON 305-343
KOREA

☐ Additional Information on page 2

LCD Monitor: 24 Vdc, 6.25A
Separate power supply: 100-240 V~, 2.5A(2.5A - 1.5A), 50-60 Hz



FS-L2602D(1), FS-L2603D(1), FS-L270(1)D(2), FS-P260(1)D(2)
See Page 2

National Differences specified in the CB Test Report.

☐ Additional Information on page 2

IEC 60601-1(ed.3)

E361872-A9-CB-1 issued on 2015-07-06



- ☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-08-10

Signature:

Jan-Erik Storgaard

Model Details:

FS-L2602D(1),FS-L2603D(1)

(1) Could be a blank or T with touch screen function; Option A

FS-L270(1)D(2)

(1) Could be a numeric(0 - 9); not related to product safety

(2) Could be a blank or T with touch screen function; Option B

FS-P260(1)D(2)

(1) Could be a numeric(0 - 9); related to product options;

0, 1, 2, 8, 9 : Option A = Metal enclosure;

3, 4, 5, 6, 7: Option B = Polymeric enclosure, 5 Vdc output port

(2) Could be a blank or T with touch screen function

Additional information (if necessary)
Information complémentaire (si nécessaire)


UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA



UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK



UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN



UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-08-10

Signature:



Jan-Erik Storgaard