



Product Service

CERTIFICATE

No. Z1 11 07 77786 001

Holder of Certificate: **Kartsana Medical GmbH**

Richardstraße 26
73054 Eisingen/Fils
GERMANY

Production Facility(ies):

26730

Certification Mark:



Product:

Stretchers
Medical stretcher for heavy load

Model(s):

Kartsana Roll In Trage "Jupiter"
(undercarriage TG-880 and medical stretcher MTR-120)

Parameters:

Length:	1965 mm
Width:	585 mm
Height of the stretcher:	230 mm
Weight of the stretcher:	22 kg (without mattress and retaining belt)
Weight of the undercarriage:	35 kg
Load:	250 kg

Tested according to: prEN 1865-3:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 71385769

Date, 2011-07-29 (Torsten Zimmer)

Page 1 of 1

