



CALIBRATION CERTIFICATE

Certificate No. AI_4365_GP10_1301
Customer Code. ACAC150
Test Method Gravimetric Method based on DIN/ISO 835
Item Calibrated Graduated Pipette
Catalogue No. 19-125.203.04
Batch No. / Serial No. 3.24
Item Calibrated On 20-03-2024

Calibrated Result

The Product was found to contain the following volume at **20°C**

Avg. Volume of Batch (ml)	9.99
Tolerance	±0.05
Certified Volume (ml)	10
Correction	+0.01

This product has been calibrated with following device:

Sartorius, S.No.: 0038203971 Range: 420 g LC: 0.001g

Thermometer : Greisinger, S.No.: TS01, Range: 15-35 °C, LC: 0.1 °C

The above product has been examined for compliance to DIN/ISO, Class A.
It has been found to satisfy the specified requirements.

The above calibration result is valid at the time of calibration under lab condition.

Calibrated By :

Date of Issue :

20-03-2024

This is an electronically generated Certificate

CALIBRATION CERTIFICATE