

EC Declaration of Conformity

Manufacturer:
Huaian Hening Medical Instruments Co.,Ltd.
NO.6, West Hongdou Road, Economic
Development Zone, Huaian,
Jiangsu, 22305 China

Authorized Representative:
Shanghai International Hoding Corp.GmbH(Europe)
Eiffestrasse 80, 20537 Hamburg, Germany
Tel:0049-40-2513175 Fax:0049-40-255726
Vat No.:DE166892350

We, the manufacturer, here with declare that the products
Sterile Blood lancets SIZE:Diameter of the pinpoint (unit: mm):0.30,0.40,0.45,0.50,0.63,0.70,0.80,0.90
Alcohol Pad SIZE:Diameter of the pinpoint (unit: mm):50*50mm (inner swab
size30*60mm,60*60mm)
Sterile Umbilical Cord Clamp

UMDNS-Code:10-440

Meet the provisions of Directive 93/42/EEC which apply to them

The medical device has been assigned to class IIa according to Annex IX of the Directive 93/42/EEC.

It bears the mark

CE 0197

The product concerned has been manufactured under a quality management system according to Annex V of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assured via assessment of the quality management system by the Notified Body

TUV Rheinland LGA Products GmbH
Tillystraße 2, 90431, Nürnberg, Germany
Certificate No.: DD 60082383 0001

Issue date:2023-03-19

Expiry date:2028-02-19

Following the procedure relating to the EC Declaration of Conformity set out in Annex V of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration and is only valid in connection with a batch specific Certificate of Compliance for all products concerned bearing the CE mark

The above mentioned declaration of conformity is exclusively under the responsibility of

Huaian Hening Medical Instruments Co.,Ltd.
NO.6, West Hongdou Road, Economic Development Zone, Huaian, Jiangsu, 223005 China

China Huaian 2023-04-12

Place, date

GM : Dazhengcan
Legally binding signature, Funct

