



# Calibration Certificate

Calibration certificate issued by an accredited calibration laboratory

## Tube Voltage

<b>Page of pages</b>	1(2)	<b>Serial Number</b>	CB3-24082127
<b>Certificate Number</b>	248A90721	<b>Object</b>	Cobia Smart R/F
<b>Date of Calibration</b>	2024-08-21		
<b>Date of Issue</b>	2024-08-21		
<b>Location</b>	RTI Group Headquarters, Mölndal		
<b>Radiation Quality</b>	Radiography (RTI)		
<b>Object</b>	kVp-, dose-, doserate- and time- meter		
<b>Manufacturer</b>	RTI Group		
<b>Man. part Number</b>	4560.000076		
<b>Calibrated By</b>	Sama Hussein		
<b>Customer</b>	RTI Group		
<b>Environment</b>	All climatic conditions are within RTI's limits for a reliable calibration environment, i.e. 18–25 deg C, 90–110 kPa, and <70 % air humidity.		
<b>Geometric Arrangement</b>	The detector was irradiated perpendicular to the entrance window. The point of reference is 10.0 mm below the top surface.		
<b>Method</b>	The method is described in the document MTB-010 (rev. I) Calibration method–Tube Potential, by RTI Group AB.		
<b>Traceability</b>	The calibration is performed by comparison against a reference high voltage divider system. The reference high voltage divider system is traceable through RISE Technical Research Institute of Sweden to national or international measurement standards.		
<b>Uncertainty</b>	The expanded uncertainty at reference conditions when calibrating is $\pm 0.56\%$ . The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95 %. The standard uncertainty of measurement has been determined in accordance with EAL Publication EA-4/02.		
<b>Evaluations</b>	The measured values are within the error limits specified by the manufacturer of the equipment under test.		

Authorized signature:

Sama Hussein



Ackred. nr. 2021  
Calibration  
ISO/IEC 17025

The calibration results refer exclusively to the object.  
This calibration certificate may not be circulated other than in full.  
Template version: 2024.3A

# Calibration Certificate

Calibration certificate issued by an accredited calibration laboratory.

## Tube Voltage

**Page of pages** 2(2)  
**Certificate Number** 248A90721  
**Serial Number** CB3-24082127  
**Date of Calibration** 2024-08-21

**Radiography (R1)** **SDD** 80 cm  
**Anode/Filter** W / 3.0 mm Al  
**HVL** 3,0 mm Al at 80 kV

Settings		Measured Data		Tolerance		Result		
Current (mA)	Time (ms)	Detector Filter	Reference kVp (kV)	Detector kVp (kV)	High (kV)	Low (kV)	Deviation (%)	Status Pass/Fail
100	100	Auto	40,43	40,20	41,43	39,43	-0,6	Pass
100	100	Auto	59,58	59,55	60,77	58,39	-0,1	Pass
100	100	Auto	59,58	59,70	60,77	58,39	0,2	Pass
100	100	Auto	99,46	99,50	101,45	97,47	0,0	Pass
100	100	Auto	119,77	119,50	122,17	117,37	-0,2	Pass
100	100	Auto	139,65	139,40	142,44	136,86	-0,2	Pass

### Pass/Fail Criteria

The pass and fail criteria are based on Case 2 as defined in *ILAC-G8:03/2009*. i.e. The calibration result is compared against manufacturer specification, see MTB-010\_bil.1\_B\_Object Uncertainty – Tube Potential. A deviation (excluding expanded uncertainty) less than manufacturer specification is reported as pass, and a larger deviation is reported as fail. Pass/Fail criteria for kV calibrations of the Cobia is  $\pm 2.0\%$ .

Reference Equipment	Ref Number	Type	Model
	11-RF005	Voltage divider	Sedecal Internal Divider
	19-5-111031	RTI X-ray lab 5:1, R/F	Sedecal, Mod. SHF 535