

Certificate of Analysis

Issue date	2024-12-13
Product	virotype® CSFV 2.0 RT-PCR Kit
Cat. Number(s)	VT281825
Lot number	F202400218
Expiry date	2025-12-31 (yyyy-mm-dd)
Storage temperature	-30°C to -15°C

Quality Control

Kit Component	Acceptance Criteria		CSFV	Internal Control	Result	Passed
	Ct Range FAM	Ct Range HEX	Ct FAM	Ct HEX		
Negative Control	No Ct	No Ct	No Ct	No Ct	negative	✓
Positive Control	25 - 30	25 - 30	27.55	28.23	positive	✓

Sample	Sample Type	CSFV Status	CSFV	Internal Control	Result	Passed
			Ct FAM	Ct HEX		
CSF 1	blood 1:1000	pos.	21.98	n.a.	positive	✓
CSF 2	blood	neg.	No Ct	27.30	negative	✓
CSF 1047	tonsil	pos.	22.37	29.93	positive	✓

n.a. = not applicable
Batch release thermal cycler: BioRad CFX96 (Bio-Rad Laboratories, Inc.)

Manufactured and tested at INDICAL BIOSCIENCE GmbH, Leipzig, Germany, and in compliance with our ISO 9001 certified quality management system. The product fulfills all INDICAL quality requirements.

Dr. Diana Kadner, Quality Assurance
This certificate is a computer printout and therefore valid without signature.