



广微测  
Gmicro Testing



201819000883

# GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## REPORT FOR ANALYSIS

**Report №.**

2025SP08337R01E

**Name of Sample**

Bactericidal air recirculator

**Applicant**

CARAVAN EXPRESS SRL

**Test Type**

Entrustment Test



**Address:** Building 66, No.100 , Xianlie Middle Road, Guangzhou, China

**Postcode:** 510070

**Tel:** +86 20 87137666

**Fax :** +86 20 87137668

**Website :** www.gddcm.com




GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS



Report №.) 2025SP08337R01E

Verification Code: LU98PTJR

Name of Sample	Bactericidal air recirculator	Test Type	Entrustment Test
Applicant	CARAVAN EXPRESS SRL	Address	Republic of MOLDOVA, Chisinau city, Decebal Boulevard 76, postal code MD-2038
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	One
Spec and Lot № of Sample	_____	State and Characteristic	Machine
Sample Received Date	2025-07-22	Test Completion Date	2025-08-08
Item Tested	Microbial removal test method		
Test Standard and Method	QB/T 2761-2024 Appendix E		
Test Conclusion	<p>The test data of the sample(s) is attached to the page(s) of this report.</p> <p style="text-align: right;">Issue Date: 2025-08-08</p> <div style="text-align: right;">               (Official Seal)              检验检测专用章         </div>		
Remarks			

Editor: Li Jiashu

Verifier: Chen Ling

Approver: Guo Weipeng



GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2025SP08337R01E

Microbial removal test method

1. Test method: QB/T 2761-2024 Appendix E
2. Test results:

Test organisms	Action time	Test times	Control group		Test Group		Killing rate (%)
			Number of bacterial colonies in initial space air (CFU/m <sup>3</sup> )	Number of bacteria in the air after sterilization (CFU/m <sup>3</sup> )	Number of bacterial colonies in initial space air (CFU/m <sup>3</sup> )	Number of bacteria in the air after sterilization (CFU/m <sup>3</sup> )	
<i>Acinetobacter baumannii</i> ATCC 19606	2 h	1	1.2 × 10 <sup>5</sup>	3.2 × 10 <sup>4</sup>	1.0 × 10 <sup>5</sup>	7	99.97
		2	1.0 × 10 <sup>5</sup>	3.0 × 10 <sup>4</sup>	1.1 × 10 <sup>5</sup>	14	99.96
		3	1.1 × 10 <sup>5</sup>	3.0 × 10 <sup>4</sup>	1.1 × 10 <sup>5</sup>	7	99.98
<i>Salmonella enteritidis</i> CMCC(B)50335	2 h	1	9.1 × 10 <sup>4</sup>	2.5 × 10 <sup>4</sup>	1.0 × 10 <sup>5</sup>	14	99.95
		2	1.1 × 10 <sup>5</sup>	3.2 × 10 <sup>4</sup>	1.0 × 10 <sup>5</sup>	14	99.95
		3	1.1 × 10 <sup>5</sup>	2.8 × 10 <sup>4</sup>	9.8 × 10 <sup>4</sup>	14	99.94

Test space: 30 m<sup>3</sup> testing chamber

Environment relative humidity: 62.2 %

Environment temperature: 24.1 °C

Note: The result of killing rate test has eliminated the effect of microbial natural decay in air.

(Blank Below)

分析  
专用



## Notice Items

1. The test report is invalid without the special stamp of testing and paging seal.
2. The test report is invalid without the signatures of the editor, the verifier and the approver.
3. The test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the test report cannot be reproduced, except in full.
5. Unless otherwise stated , the results shown in this test report refer only to the sample(s) tested.
6. Any dispute of the test report must be raised to the testing institution within 15 days after the test report is received, exceeding which the dispute will not be accepted.
7. For the sample(s) submitted for testing, the sample information is provided by the applicant, and this testing institution is not responsible for its authenticity.

