

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg, Mumbai - 400086

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity


Material Code : SD041	Material Name : Cefoxitin (Cephoxitin) CX 30 mcg	Lot No : 0000418175
AR No.: 040000984539	Date of Report : 2020-01-08	Exp. Date : 2021-04

TEST	SPECIFICATIONS	RESULTS
<u>Appearance</u> 1 Appearance	Filter paper discs of 6 mm diameter with printed "CX 30" on centre of each side of the disc.	Complies
<u>Cultural response</u> 1 Cultural response	Average diameter of zone of inhibition observed on Mueller Hinton Agar (M173) after 18 hours incubation at 35-37°C for standard cultures.	
<u>Escherichia coli ATCC 25922</u> 1 Zone of inhibition (mm)	23mm - 29mm	28mm
<u>Staphylococcus aureus ATCC 25923</u> 1 Zone of inhibition (mm)	23mm - 29mm	27mm

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid


Sukanya Deshmukh

Microbiologist/Sr.Executive Microbiologist

Ujjwala M. Kokate

Asst./Dy/QC Manager

Dr. Santosh Kaul

Dy/QA Manager