

Certificate of Analysis



POP-7 Polymer (384 Samples) 3500 Series

Product No. 4393708
Lot No. 2604369
Date of Manufacture 04JUN2026
Expiration Date 19JAN2027

TEST	RESULT
------	--------

VISUAL INSPECTION

NOTE The polymer is particulate free at room temperature, but upon refrigeration, urea may precipitate out of solution.

Clear, particulate free, viscous solution at room temperature	Pass
---	------

FLUORESCENCE TEST

The ratio of the background fluorescence intensity of the polymer to that of water is measured for each bulk lot to ensure conformance to manufacturing specifications

Fluorescence Ratio	Pass
--------------------	------

RESOLUTION TEST

Resolution is evaluated by measuring separations and widths of peaks in a multi-fragment DNA standard on a multi-capillary Genetic Analyzer

Resolution	Pass
------------	------

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Kimberly Boge

**Quality Assurance
Issued 08JUN2026**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com