



www.taidoc.com

泰博科技股份有限公司  
TaiDoc Technology Corp.

新北市24888五股區五工二路127號B1-7樓  
B1-7F., No.127, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24888, Taiwan

Tel : +886-2-6625-8188  
Fax : +886-2-6625-0288

## EC Declaration of Conformity

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, 24888 Wugu Dist, New Taipei City, TAIWAN

declare under our sole responsibility that the product

**Product Name** : Blood Glucose Monitoring System  
**Product Model** : TD-4227  
**Classification** : 98/79/EC (IVDD), Annex II, List B  
**Conformity Assessment Route** : 98/79/EC (IVD), Annex IV excluding section 4&6  
**EC Certificate Number** : V1 052126 0042 Rev.01  
**European Representative** : MedNet EC-REP GmbH  
Borkstraße 10, 48163 Münster , Germany  
**Notified Body (CE0123)** : TÜV SÜD Product Service GmbH  
Ridlerstraße 65, 80339 München, Germany  
**GMDN code** : 62537

to which this declaration relates is in conformity with the following standard(s) or other normative document(s) :

|                       |                                                                                                                                                 |
|-----------------------|-------------------------------------------------------------------------------------------------------------------------------------------------|
| ISO 13485:2016        | Medical devices - Quality management systems - Requirements for regulatory purposes                                                             |
| EN ISO 14971:2012     | Medical devices - Application of risk management to medical devices                                                                             |
| EN 61010-1:2001       | Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements.                                 |
| EN 62304:2006+A1:2015 | Medical device software -- Software life cycle processes                                                                                        |
| EN ISO 15197:2015     | In vitro diagnostic test systems —Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus              |
| EN ISO 23640:2015     | In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents                                                    |
| EN 15223-1:2016       | Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied. Part 1: General requirements         |
| EN ISO 18113-1:2011   | In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use |



|                           |                                                                                                                                                                                 |
|---------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| EN ISO 18113-3:2011       | In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use                                 |
| EN ISO 18113-2:2011       | In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for professional use                                    |
| EN ISO 18113-4:2011       | In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for self-testing                                        |
| EN ISO 18113-5:2011       | In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing                           |
| EN 13532:2002             | General requirements for in vitro diagnostic medical devices for self-testing                                                                                                   |
| EN 13612:2002             | Performance evaluation of in vitro diagnostic medical devices                                                                                                                   |
| EN 61326-1:2006           | Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements                                                                        |
| EN 61326-2-6 :2006        | Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment           |
| EN 61010-2-101:2002       | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment |
| EN 62366-1:2015           | Medical devices -- Application of usability engineering to medical                                                                                                              |
| EN 60601-1-6:2010+A1:2015 | Medical electrical equipment. General requirements for basic safety and essential performance. Collateral standard. Usability                                                   |
| EN ISO 17511:2003         | In vitro diagnostic medical devices. Measurement of quantities in biological samples. Metrological traceability of values assigned to calibrators and control materials         |
| EN 50581:2012             | Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances                                        |
| 2011/65/EU                | The restriction of the use of certain hazardous substances in electrical and electronic equipment.                                                                              |



www.taidoc.com

泰博科技股份有限公司  
TaiDoc Technology Corp.

新北市24888五股區五工二路127號61-7樓  
61-7F., No.127, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24888, Taiwan

Tel : +886-2-6625-8188  
Fax : +886-2-6625-0288

The object of the declaration described above is in conformity with Directive 98/79/EC, 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

2020. 7. 15:

Date of Issue

**Jim Jan**

Management Representative