

CERTIFICATE OF ANALYSIS

F-548L Phusion Flash High-Fidelity PCR Master Mix, 500 react.

Packaging Lot: 2923029

Expiry Date: 31.12.2024 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

Lot	Quantity	Description
2877131	5 × 1 mL	2X Phusion Flash Master Mix

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional testing	PCR amplification of 7.5 kb fragments from human genomic DNA and 20 kb fragment from lambda DNA and analysis on agarose gel after ethidium bromide staining.	Reactions produce specific PCR products	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

