

EU Declaration of Conformity

In accordance with EU Regulation 2017/746 of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices.

Manufacturer:	Ventana Medical Systems Inc. 1910 E Innovation Park Drive Tucson, AZ 85755, USA
Single Registration Number (SRN) Manufacturer:	US-MF-000016993
Authorized Representative:	Roche Diagnostics GmbH Sandhofer Strasse 116 68305 Mannheim Germany
Single Registration Number (SRN) Authorized Representative:	DE-AR-000006262

This declaration is issued under the sole responsibility of Ventana Medical Systems Inc.

Product Information

Part Number:		Product Name:	Basic UDI-DI:
09780041001 (alternative P/	′N 790-7176)	anti-CD8 (SP239) Rabbit	761333602693BJ
		Monoclonal Primary Antibody	
Intended Purpose:	ose: Anti-CD8 (SP239) Rabbit Monoclonal Primary Antibody is intended for laboratory use in the qualitative immunohistochemical detection of CD8 by light microscopy in sections of formalin-fixed, paraffin-embedded tissue stained on a BenchMark IHC/ISH instrument.		
Risk Class:	Class C		
Common Specifications:	Not applicable as no Common Specifications exist for the concerned device.		
Name, Address and Identification number of the Notified Body:	TÜV SÜD Product Service GmbH (No. 0123) Ridlerstraße 65 80339 MÜNCHEN Germany		

Conformity Assessment was established through the procedure described in Annex IX of EU Regulation 2017/746, including an assessment of the Technical Documentation as described in Annex IX, Chapter II. This declaration is supported by the following certificate(s):

EU Quality Management System Certificate: V12 096981 0003



Conformity of the product with EU Regulation 2017/746 and other applicable EU legislation has been established.

On behalf of Ventana Medical Systems Inc.

Place: Tucson, AZ 85755, USA

Date:

DocuSigned by:

Jeff Boone

Jeff Boone

Site Head of Quality Function

Place: Tucson, AZ 85755, USA

Date:

----- DocuSigned by:

Benjamin (urson

Benjamin Curson

Site Head of Regulatory Affairs Function