

**1. PRODUCT**

FULL REINFORCED SURGICAL GOWN

**2. PRODUCT CONFIRMITIES**

ISO 9001, ISO 13485, EN 13795, ISO 11135:2014, ISO 14644-3:2019, ISO 14791:2012, ISO 14937: 2009, ISO 10993-7:2008, EN 20811

**3. DESCRIPTION**

Nonwoven , 3 layers SMS 35 gsm, 3 gsm medical glue, 15 gsm white lamination, full reinforced on front and sleeves  
Resistance to penetration of liquids within 50-80cm H2O according to EN 20811. Velcro adjustment at the Neck, overlapping at the back by ties , Elastic Cuff Kinitted

Antiallergenic, AirPermeable, Non-Linting

Water Repellent , latex free, full ultrasonic seems seam (including armpits)

**4. PACKING**

The product is individually packaged. On the packaging is the inscription with the name and address of the manufacturer, the product model, description, dimensions, date of manufacture, validity date, lot number, etc.

**5. STORAGE**

0°C - + 30°C, Avoid Sunlight, Avoid Humidity

**6. Term of validity**

Sterile EO. 3 years from the date of production

Size	Lengh (cm)
<b>S (SMALL)</b>	min.120
<b>M (MEDIUM)</b>	min. 125
<b>L (LARGE)</b>	min. 135
<b>XL (XLARGE)</b>	min. 140
<b>2XL (XXLARGE)</b>	min. 150



### Ultrasonic Welding Seams



Ultrasonic welding is operated by machine, easy to control quality with high output. Make sure good quality. Avoid the problem of stitches coming loose, more firm and durable; Good ability of leak-proof and permeation-resistance;