

The management system of

# Mediplus (India) Limited

1261-1262, M.I.E., Part B, Bahadurgarh-124507, Haryana, India

has been assessed and certified as meeting the requirements of

## Directive 93/42/EEC

on medical devices, Annex II (excluding Section 4)

For the following products

**Sterile I. V. Cannula, Sterile I.V. Cannula with Safety Feature,  
Sterile 3-Way Stop Cock, Sterile 3-Way Stop Cock with Extension  
Tubing, Sterile Injection Stopper, Sterile Male-Female Luer Lock,  
Sterile Luer Lock, Sterile Needle Free Valve  
and Sterile Pressure Monitoring line/ Sterile Extension Tube**


Where the above scope includes class III medical device(s), a valid EC Design Examination Certificate according to Annex II (Section 4) is a mandatory requirement for each device in addition to this certificate to place that device on the market.

This certificate is valid from 09 April 2020 until 12 January 2023  
and remains valid subject to satisfactory surveillance audits.

Issue 2. Certified since 12 January 2015  
and first certified by SGS Belgium NV since 16 December 2019.

Certification is based on reports numbered IN/GUR 235607

Authorised by



**SGS Belgium NV, Notified Body 1639**

SGS House Noorderlaan 87 2030 Antwerp Belgium  
t +32 (0)3 545-48-48 f +32 (0)3 545-48-49 www.sgs.com

LPMD5007 - Certificate CE1639 Annex II-4\_EN rev. 02

Page 1 of 1

