

RIBO-prep nucleic acid extraction kit



For Professional Use Only

Instruction Manual

KEY TO SYMBOLS USED

	Catalogue number		Caution
	Batch code		Contains sufficient for <n> tests
	In vitro diagnostic medical device		Use-by Date
	Version		Consult instructions for use
	Temperature limit		GHS02: Flame
	Manufacturer		GHS05: Corrosion
	Date of manufacture		GHS07: Exclamation mark
	Authorized representative in the European Community		

1. INTENDED USE

RIBO-prep nucleic acid extraction kit is intended for extraction of total RNA/DNA from human biological material for subsequent testing by the nucleic acid amplification techniques (NAT):

- plasma of venous, umbilical cord blood,
- serum of venous, umbilical cord blood,
- whole venous, umbilical cord blood,
- leukocytes of venous, umbilical cord blood,
- cerebrospinal fluid (liquor),
- discharge of the conjunctiva,
- saliva,
- nasopharyngeal and oropharyngeal swabs,
- sputum,
- endotracheal aspirate,
- bronchial washings,
- bronchoalveolar lavage fluid,
- pleural fluid,
- transudates,
- amniotic liquid,
- breast milk,
- pus / liquid or dense necrotic content,
- vesicles and pustules content,
- hair follicles,
- nail plates,
- tissue (biopsy, surgical, autopsy) native material,
- paraffin-embedded tissue (biopsy, surgical, autopsy) material,
- urogenital mucous discharge (swab, scrape),
- urine,
- prostate secretion,
- feces / fecal swab,
- anal canal/rectal mucosal discharge,
- cultures of microorganisms, isolated by inoculation of human biological material.

Reagent kit can be used for extraction of total RNA/DNA from ticks, mosquitoes, washes from environmental objects, water; water sample concentrates (drinking, open sources, sewage) for carrying out prophylactic measures in order to prevent human diseases.

Indications and contra-indications for use of the reagent kit

RNA/DNA extraction is used in preanalytical stage of *in vitro* diagnostics by NAT.

2. PRINCIPLE OF NUCLEIC ACID EXTRACTION

Test samples are treated by **Solution for Lysis** to destruct cell membranes, viral envelopes and other biopolymer complexes and release nucleic acids and cellular components. The dissolved RNA/DNA precipitates after addition of the **Solution for Precipitation** and centrifugation, while the other components of the lysed clinical material remain in the solution and removed with subsequent washes. The final stage of extraction is dissolution of the pellet in elution buffer, the purified RNA/DNA is transferred into the solution.

The obtained nucleic acid sample is purified and free from inhibitors of amplification, which provides high analytical sensitivity of NAT assay.

3. CONTENT

RIBO-prep nucleic acid extraction kit is produced in 1 form:

variant 100, K2-9-Et-100-CE.

Variant 100 includes:

Reagent	Description	Volume, ml	Quantity
Solution for Lysis	clear liquid from colorless to blue grey colour ¹	30	1 vial
Solution for Precipitation	colorless clear liquid	40	1 vial
Washing Solution 3	colorless clear liquid	50	1 vial
Washing Solution 4	colorless clear liquid	20	1 vial
RNA-buffer	colorless clear liquid	1.2	8 tubes

Variant 100 is intended for RNA/DNA extraction from 100 samples, including controls. While working with nail plates and hair follicles **variant 100** is intended for RNA/DNA extraction from 50 samples, including controls.

4. ADDITIONAL REQUIREMENTS

- 1.5-ml disposable polypropylene screwed or tightly closed tubes.
- Screwing caps for tubes.
- Sterile DNase- and RNase-free pipette tips without filter (up to 100 and 200 µl).
- Sterile RNase-free pipette tips with aerosol filters (up to 100 µl, 200 µl).
- Tube racks for 1.5-ml tubes.
- Vortex mixer.
- Thermostat with working temperature for 25-100 °C.
- Desktop microcentrifuge.
- Vacuum aspirator with flask for removing supernatant.
- PCR box or Biological cabinet.
- Pipettes (adjustable).
- Refrigerator for 2-8 °C.
- Deep-freezer at the temperature from minus 24 to minus 16 °C.
- Reservoir for used tips.
- Disposable powder-free gloves and a laboratory coat.

5. GENERAL PRECAUTIONS



The user should always pay attention to the following:

- Use sterile DNase- and RNase-free pipette filter tips and use a new tip for every procedure.
- Store all extracted positive material (specimens, controls and amplicons) away from all other reagents and add it to the reaction mix in a distinctly separated facility.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable protective gloves and laboratory cloths, and protect eyes while samples and reagents handling. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use the kit if the internal packaging was damaged or its appearance was changed.
- Do not use the kit if the transportation and storage conditions according to the Instruction Manual were not observed.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local regulations.
- Samples should be considered potentially infectious and handled in a biological cabinet in compliance with appropriate biosafety practices.
- Clean and disinfect all sample or reagent spills using a disinfectant, such as 0.5% sodium hypochlorite or another suitable disinfectant.
- Avoid inhalation of vapors, samples and reagents contact with the skin, eyes, and mucous membranes. Harmful if swallowed. If these solutions come into contact, rinse the injured area immediately with water and seek medical advice if necessary.
- While observing the conditions of transportation, operation and storage, there are no risks of explosion and ignition.
- Safety Data Sheets (SDS) are available on request.
- The kit is intended for analysis of specified number of samples (see the section "Content").
- The kit is ready for use in accordance with the Instruction Manual. Use the kit strictly for intended purpose.
- Use of this product should be limited to personnel trained in DNA/RNA extraction techniques.
- Workflow in the laboratory must be one-directional, beginning in the Extraction Area and moving to the Amplification and Detection Area. Do not return samples, equipment and reagents to the area where the previous step was performed.

NOTE: **Solution for Lysis** has an unpleasant smell. Work with this solution should be performed in a biological cabinet.

 Danger	Contains substance: guanidine thiocyanate. H302: Harmful if swallowed. H312: Harmful in contact with skin. H314: Causes severe skin burns and eye damage. H318: Causes serious eye damage. H332: Harmful if inhaled. H412: Harmful to aquatic life with long lasting effects. EUH032: Contact with acids liberates very toxic gas. P260: Do not breathe vapours. P264: Wash your hands thoroughly after handling. P273: Avoid release to the environment. P303+P361+P353: IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. P501: Dispose of contents in accordance with national regulation.
--------------------	--

¹ If Solution for Lysis is stored at 2-8 °C, a crystalline precipitate may form.

 Warning	Contains substance: isopropanol H226: Flammable liquid and vapour. H319: Causes serious eye irritation. H336: May cause drowsiness or dizziness P210: Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking. P261: Avoid breathing vapours. P264: Wash your hand thoroughly after handling. P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses if present and easy to do – continue rinsing. P403+P233: Store in a well ventilated place. Keep container tightly closed. P501: Dispose of contents in accordance with national regulation.
	Isopropanol EC No 200-661-7 CAS No 67-63-0 H225: Highly flammable liquid and vapour. H319: Causes serious eye irritation. H336: May cause drowsiness or dizziness P210: Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking. P261: Avoid breathing vapours. P280: Wear protective gloves/protective clothing/eye protection/face protection. P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses if present and easy to do – continue rinsing. P403+P233: Store in a well ventilated place. Keep container tightly closed. P501: Dispose of contents in accordance with national regulation.
 Danger	

6. SAMPLING AND HANDLING

NOTE: See the information about the sampling, conditions of transportation and storage of the test material, the necessity and procedure of its pretreatment before DNA/RNA extraction in the *Instruction manual* for the PCR kit.

RIBO-prep nucleic acid extraction kit is recommended for **RNA and DNA** extraction and purification from:

- plasma of venous, umbilical cord blood,
- serum of venous, umbilical cord blood,
- whole venous, umbilical cord blood,
- leukocytes of venous, umbilical cord blood,
- cerebrospinal fluid (liquor),
- discharge of the conjunctiva,
- saliva,
- nasopharyngeal and oropharyngeal swabs,
- sputum,
- endotracheal aspirate,
- bronchial washings,
- bronchoalveolar lavage fluid,
- pleural fluid,
- transudates,
- amniotic liquid,
- breast milk,
- pus / liquid or dense necrotic content,
- vesicles and pustules content,
- hair follicles,
- nail plates,
- tissue (biopsy, surgical, autopsy) native material,
- paraffin-embedded tissue (biopsy, surgical, autopsy) material,
- urogenital mucous discharge (swab, scrape),
- urine,
- prostate secretion,
- feces / fecal swab,
- anal canal/rectal mucosal discharge,
- cultures of microorganisms, isolated by inoculation of human biological material.
- ticks, mosquitoes,
- washes from environmental objects,
- water; water sample concentrates (drinking, open sources, sewage).

Interfering substances and limitations of using test material samples

The information about potential interfering substances and limitations of using test material samples is specified in the *Instruction Manual* of the PCR kit.

Potential interfering substances

Endogenous and exogenous substances that may be present in the biological material used for the study were selected to assess potential interference (see Table 1). Model samples of various biological material without adding and with the addition of potentially interfering substances were tested. The concentration of each potentially interfering substance is listed in Table 1.

Table 1

Test material	Type of potential interferent	Potential interferent	Tested concentration in a sample	Interference presence
Plasma of venous, umbilical cord blood	Endogenous substances	Hemoglobin	5 g/l (upper limit of normal 1 g/l)	Not detected
		Triglycerides	37 mmol/l (upper limit of normal 3.7 mmol/l)	Not detected
		Bilirubin	210 µmol/l (upper limit of normal 21 µmol/l)	Not detected
		Protein	120 g/l (upper limit of normal 85 g/l)	Not detected
Serum of venous, umbilical cord blood, whole venous, umbilical cord blood, leukocytes of venous, umbilical cord blood	Endogenous substances	Bilirubin	210 µmol/l (upper limit of normal – 21 µmol/l)	Not detected
		Cholesterol	78 µmol/l (upper limit of normal – 7.8 µmol/l)	Not detected
		Triglycerides	37 mmol/l (upper limit of normal – 3.7 mmol/l)	Not detected
	Exogenous substances	Heparin	15 UI/ml	<u>Detected</u>
Cerebrospinal fluid (liquor)	Exogenous substances	Whole blood	Up to 4 %	Not detected
		Leukocytes	500 cells/mm ³	Not detected

Test material	Type of potential interferent	Potential interferent	Tested concentration in a sample	Interference presence
Discharge of the conjunctiva	Exogenous substances	Tetracycline	5 mg/ml	Not detected
Saliva, nasal and oropharyngeal swabs	Endogenous substances	Mucin	5 %	Not detected
	Exogenous substances	Lugol's solution with glycerin	0.5 %	Not detected
		Chlorhexidine bigluconate aqueous solution	2.5 %	Not detected
Sputum, endotracheal aspirate, bronchial washings, bronchoalveolar lavage fluid	Exogenous substances	Hypertonic sodium chloride solution	5 % (the drug volume to the volume of the test sample)	Not detected
		Amoxicyclin+ clavulanic acid	1125 µg/ml + 281 µg/ml	Not detected
		Cetirifaxone	1500 µg/ml	Not detected
		Rifampicin	9-17.5 µg/ml	Not detected
		Isoniazid	3-7 µg/ml	Not detected
		Ethambutol	2-5 µg/ml	Not detected
		Pyrazinamide	39 µg/ml	Not detected
		Ofloxacin	5.2 µg/ml	Not detected
		Kanamycin	22 µg/ml	Not detected
		Para-aminosalicylic acid (PAS)	75 µg/ml	Not detected
		Cycloserine	25-30 µg/ml	Not detected
	Endogenous substances	Blood	1:1	Not detected
Amniotic liquid	Exogenous substances	Cefazolin sodium salt	64 µg/ml	Not detected
Brest milk	Exogenous substances	"Miramistin"	16 % (the drug volume to the volume of the test sample)	Not detected
Pleural liquid, transudates	Endogenous substances	Blood	1:1	Not detected
	Exogenous substances	Metronidasol	16 % (the drug volume to the volume of the test sample)	Not detected
		Rifampicin	9-17.5 µg/ml	Not detected
		Isoniazid	3-7 µg/ml	Not detected
		Ethambutol	2-5 µg/ml	Not detected
		Pyrazinamide	39 µg/ml	Not detected
		Ofloxacin	5.2 µg/ml	Not detected
		Kanamycin	22 µg/ml	Not detected
		Para-aminosalicylic acid (PAS)	75 µg/ml	Not detected
		Cycloserine	25-30 µg/ml	Not detected
Pus / liquid or dense necrotic contents	Endogenous substances	Blood	1:1	Not detected
	Exogenous substances	Hemoglobin	250 g/l	Not detected
		Metronidasol	16 % (the drug volume to the volume of the test sample)	Not detected
		Rifampicin	9-17.5 µg/ml	Not detected
		Isoniazid	3-7 µg/ml	Not detected
		Ethambutol	2-5 µg/ml	Not detected
		Pyrazinamide	39 µg/ml	Not detected
		Ofloxacin	5.2 µg/ml	Not detected
		Kanamycin	22 µg/ml	Not detected
		Para-aminosalicylic acid (PAS)	75 µg/ml	Not detected
		Cycloserine	25-30 µg/ml	Not detected
Vesicles and pustules content	Exogenous substances	Iodine (Potassium iodide)	0.5%	Not detected
Hair follicles	Exogenous substances	"Hydrocortisone" ointment for external application	1 % (the drug weight to the volume of the test sample)	Not detected
Nail plates	Exogenous substances	Chlorhexidine bigluconate aqueous solution	2.5 %	Not detected
Tissue (biopsy, surgical, autopsy) native material	Exogenous substances	Vancomycin	1500 µg/ml	Not detected
Paraffin-embedded tissue (biopsy, surgical, autopsy) material	Exogenous substances	Formaline sour	—	<u>Detected</u>
		Buffered formaline	—	Not detected
Urogenital mucous discharge (swab, scrape)	Endogenous substances	Hemoglobin	260 µg/ml	Not detected
		Lactoferrin	5 µg/ml	
		Glycogen	120 mg/ml	
	Exogenous substances	Mucin	150 µg/ml	
		Clotrimazole + "Neomycin" + "Nystatin" + "Polymixin B"	16 % (the drug volume to the volume of the test sample)	
Urine	Endogenous substances	Albumin	500 mg/ml	Not detected
		Blood	1:1	Not detected
	Exogenous substances	Azithromycin	0.8 mg/ml	Not detected
		"Rifampicin"	9-17.5 µg/ml	Not detected
		"Isoniazid"	3-7 µg/ml	Not detected
		"Ethambutol"	2-5 µg/ml	Not detected

Test material	Type of potential interferent	Potential interferent	Tested concentration in a sample	Interference presence
		"Pyrazinamide"	39 µg/ml	Not detected
		"Ofloxacin"	5.2 µg/ml	Not detected
		"Kanamycin"	22 µg/ml	Not detected
		Para-aminosalicylic acid (PAS)	75 µg/ml	Not detected
Prostate gland secret	Endogenous substances	Fructose	10 mg/ml	Not detected
	Exogenous substances	Ibuprofen	300 µg/ml	Not detected
Feces/ fecal swab	Endogenous substances	Blood	1:1	Not detected
		Mucin (mucus)	Up to 3 % (the drug weight to the volume of the test sample)	Not detected
		Fecal fats	40%	Not detected
	Exogenous substances	Loperamide	Up to 5 mg/ml	Not detected
		Hydrocortisone	Up to 3 % (the drug weight to the volume of the test sample)	Not detected
		Rifampicin	9-17.5 µg/ml	Not detected
		Isoniazid	3-7 µg/ml	Not detected
		Ethambutol	2-5 µg/ml	Not detected
		Pyrazinamide	39 µg/ml	Not detected
		Ofloxacin	5.2 µg/ml	Not detected
		Kanamycin	22 µg/ml	Not detected
		Para-aminosalicylic acid (PAS)	75 µg/ml	Not detected
		Cycloserine	25-30 µg/ml	Not detected
Anal canal/ rectal mucosal discharge	Exogenous substances	Glycerin	20%	Not detected
Ticks, mosquitoes	Endogenous substances	Bilirubin	210 µmol/l	Not detected
		Cholesterol	78 mmol/l	Not detected
		Triglycerides	37 mmol/l	Not detected
		Hemoglobin	5 g/l	Not detected
		Protein	120 g/l	Not detected
Water, water samples concentrates (drinking, open sources, sewage)	Exogenous substances	Feces	5 %	Not detected

7. WORKING CONDITIONS

RIBO-prep nucleic acid extraction kit should be used at the temperature from 20 to 28 °C and relative humidity from 15 to 75 %.

8. PROTOCOL

8.1. RNA and DNA Extraction

It is allowed to use another types of biological material, volumes of investigated sample according to the *Instruction manual* for the used PCR kit.

NOTE: To work with RNA, it is necessary to use only disposable sterile plastic consumables that have a special marking RNase-free, DNase-free.

Procedure (for all types of biological material except nail plates and hair follicles)

1. Warm **Solution for Lysis** (if stored at 2-8 °C) at the temperature **65 °C** until crystals disappear.

NOTE: **Solution for lysis** has unpleasant odor. Work in the PCR box.

2. Take required number of 1.5-ml disposable tubes with tightly closable caps (including one tube for **Negative Control of Extraction (C-)** and one tube for **Positive Control of Extraction (PCE)** if they are provided for analysis). Mark the tubes.

3. Add **10 µl** of **Internal Control** (if it is provided for analysis) to each tube and then add **300 µl** of **Solution for Lysis**.

NOTE: It is allowed to change the volume of **Internal Control** according to the *Instruction manual* for the used PCR kit.

4. Add **100 µl** of prepared samples to the tubes with **Solution for Lysis** and **Internal Control** (if used) using pipette tips with aerosol filters. Add **100 µl** of **Negative Control** (if it is provided for analysis) to the tube labeled **C-**. Add **90 µl** of **Negative Control** and **10 µl** of **Positive Control** (if it is provided for analysis) to the tube labeled **PCE**.

NOTE: It is allowed to change the volume and dilution of **Negative Control (C-)** and **Positive Control of Extraction (PCE)** according to the *Instruction manual* for the used PCR kit.

5. Thoroughly mix the contents of the tubes by vortexing, then centrifuge tubes for 5 s to be sure there are no drops on the cap, and incubate them at **65 °C** for **5 min**.

6. Add **400 µl** of **Solution for Precipitation**, tightly close the tubes and mix them by vortexing.

7. Centrifuge all tubes for **5 min** at **13,400 rpm**.

8. Carefully remove the supernatant without disturbing the pellet using a vacuum aspirator and 200-µl tips. Use a new tip for each tube.

9. Add **500 µl** of **Washing Solution 3** to each tube, tightly close the tubes and turn them carefully upside down 3–5 times to wash the pellet. This procedure can be performed simultaneously for all the tubes: cover the tubes placed in a rack with a lid or another rack, press them, and turn the rack.

10. Centrifuge all the tubes at **13,400 rpm** for **1–2 min**.

11. Carefully remove the supernatant without disturbing the pellet using a vacuum aspirator and 10-µl tips. Use a new tip for each tube.

12. Add **200 µl** of **Washing Solution 4** to each tube, tightly close the tubes and turn them carefully upside down and back 3–5 times to wash the pellet.

13. Centrifuge all tubes at **13,400 rpm** for **1–2 min**.

14. Carefully remove the supernatant without disturbing the pellet using a vacuum aspirator and 10-µl tips. Use a new tip for each tube.

15. Incubate all tubes with open caps at **65 °C** for **5 min** (to dry the pellet).

16. Add **50 µl** of **RNA buffer** into each tube. Mix the tubes by vortex. Then incubate them at **65 °C** for **5 min** occasionally stirring by vortex.

NOTE: Increasing of the elution volume is allowed (for example up to 90 µl) according to the *Instruction manual* for the used PCR kit.

17. Centrifuge the tubes at **13,400 rpm** for **1 min**. The supernatant contains purified RNA and DNA. Samples are ready for reverse transcription and PCR.

Procedure (for nail plates and hair follicles)

1. Warm **Solution for Lysis** (if stored at 2-8 °C) at the temperature **65 °C** until crystals disappear.

2. Take required number of 1.5 ml disposable tubes with tightly closable caps (including one tube for **Negative Control of Extraction (C-)** and one tube for **Positive Control of Extraction (PCE)** if they are provided for analysis). Mark the tubes.

3. Add Internal Control (if it is provided for analysis) to each tube in a volume 1.5 times more than recommended volume for other types of biological material.

NOTE: It is allowed to change the volume of **Internal Control** according to the *Instruction manual* for the used reagents kit for carrying out the amplification.

4. Add to each tube per one examination, 2 samples of nail plates (about 2x10 mm in size) or 2-3 hair follicles with hair shaft no more than 1-2 cm long.

5. Add **100 µl** of **Negative Control** (if it is provided for analysis) to the tube labeled **C-**. Add **85 µl** of **Negative Control** and **15 µl** of **Positive Control** (if it is provided for analysis) to the tube labeled **PCE**.

NOTE: It is allowed to change the volume and dilution of **Negative Control (C-)** and **Positive Control of Extraction (PCE)** according to the *Instruction manual* for the used PCR kit.

6. Add **600 µl** of **Solution for Lysis** to the tubes. Make sure that test material is completely immersed into the solution. Tightly close the tubes.

7. Mix **C-** and **PCE** tubes thoroughly by vortexing, then centrifuge tubes for 5 s to be sure there are no drops on the cap.

8. Put the tubes with test and control samples to the thermostat and incubate them at **65 °C** for **20 min**.

9. After incubation mix thoroughly by vortexing, then centrifuge tubes for **2 min** at **13,000 rpm**.

10. Take necessary number of 1.5-ml disposable polypropylene screwed tubes (including controls of the extraction if they are provided for analysis). Mark the tubes.

11. Carefully remove the supernatant and add **400 µl** of it to the prepared tubes.

12. Continue the extraction according to the protocol for other biological materials (see above), starting with entry 5.

The purified RNA/DNA can be stored:

- at 2-8 °C for 24 h;
- at the temperature from minus 24 to minus 16 °C for 1 year.

8.2. Amplification

It's recommended to use AmpliSens® PCR kits and REVERTA-L reverse transcription reagents kit.

NOTE: Carry out the amplification according to the manufacturer instruction.

9. TROUBLESHOOTING

These troubleshooting rules may be helpful in explaining any questions that may arise.

False negatives with extraction product:

- Degradation of the nucleic acid contained in the sample. It is necessary to use a new sample. Store samples under appropriate conditions.
- Loss of nucleic acid pellet. Carefully discard the washing solution trying not to disturb the nucleic acid pellet.
- Degradation of the extracted nucleic acid. It is necessary to use DNase- and RNase-free plastic.

False positives with extraction product:

- Contamination during sample extraction. It is necessary to open one test tube at a time. Avoid spilling the contents of the test tube. Always change tips.
- Contamination of the reagents prepared for the step. It is necessary to repeat the test.
- Contamination of the extraction zone by amplicons. It is necessary to clean surfaces and instruments using aqueous detergents, wash lab coats, and replace test tubes and tips in use. Use different laboratory coats in different areas.

If you have any questions or encounter problems, please contact our Authorized Representative in the European Community.

10. TRANSPORTATION

RIBO-prep nucleic acid extraction kit should be transported at 2–25 °C for no longer than 5 days.

11. STABILITY AND STORAGE

All components of **RIBO-prep** nucleic acid extraction kit are to be stored at 2-8°C, when not in use. All components of **RIBO-prep** nucleic acid extraction kit are to be stable until the expiration date on the label. The shelf life of reagents before and after the first use is the same, unless otherwise stated.

12. REFERENCES

1. H. Zetzsche, H.-P. Klenk, M.J. Raupach, T. Kneibelsberger, B. Gemeinholzer. Comparison of methods and protocols for routine DNA extraction in the DNA Bank Network // DNA Bank Network. URL: <https://www.dnabank-network.org>. [date of the application: 2020/02/05]
2. Gubala V. et al. Point of care diagnostics: status and future //Analytical chemistry. – 2012. – T. 84. – №. 2. – C. 487-515.
3. Boom R. et al. Rapid and simple method for purification of nucleic acids //Journal of clinical microbiology. – 1990. – T. 28. – №. 3. – C. 495-503.
4. Boom R. et al. Improved silica-guanidiniumthiocyanate DNA isolation procedure based on selective binding of bovine alpha-casein to silica particles //Journal of Clinical Microbiology. – 1999. – T. 37. – №. 3. – C. 615-619.

13. QUALITY CONTROL

In compliance with Federal Budget Institute of Science "Central Research Institute for Epidemiology" ISO 13485-Certified Total Quality Management System, each lot of **RIBO-prep** nucleic acid extraction kit has been tested against predetermined specifications to ensure consistent product quality.

Please contact our Authorized representative in the European Community if side effects, undesirable reactions, facts and circumstances that pose a threat to the life and health of citizens and medical workers are identified during the use of the reagent kit.

List of Changes Made in the Instruction Manual

VER	Location of changes	Essence of changes
10.12.10	Cover page	The phrase "For Professional Use Only" was added
	Content	New sections "Working Conditions" and "Transportation" were added
	Stability and Storage	The "Explanation of Symbols" section was renamed to "Key to Symbols Used"
	Key to Symbols Used	The information about the shelf life of open reagents was added
	Content	The explanation of symbols was corrected
04.07.11 VV	Cover page, text	The color of Solution for Lysis was changed into blue
06.09.11 RT	8. Protocol 8.1. RNA/ DNA extraction	The volume of Washing Solution 3 was changed into 25 ml (for variant 50)
12.07.12 BM	5. General precautions	The reference «If Solution for Lysis is stored at 2-8 °C, a crystalline precipitate may form» was added
31.03.15 ME	5. General precautions, 14. Key to symbols used	The name of Institute was changed to Federal Budget Institute of Science "Central Research Institute for Epidemiology"
10.05.17 PM	Through the text 5. General precautions, 14. Key to symbols used	Procedure of extraction was corrected (heating at 65 °C for 5 min was added in Section 8.1, article 4).
27.03.18 PM	3. Content	Information about an unpleasant smell of Solution for Lysis and the necessity to work in a biological cabinet was added
09.04.20 KK	Through the text Footer	Information about hazards was corrected
21.10.20 KK	Footer, 3. Content	Correction according to the template
11.03.21 VA	—	Information about hazards was rewritten according to the Regulation 1272/2008/EC.
12.08.21 EM	Interfering substances and limitations of using test material samples	The color of the reagent was specified
	Principle of nucleic acid extraction	The text formatting was changed
	Through the text	The phrase "Not for use in the Russian Federation" was added
31.05.22 KK	1. Intended use	The information about variant 50, REF K2-9-Et-50-CE was deleted
	5. General precautions	The name, address and contact information for Authorized representative in the European Community was changed
	13. Quality control	The sections were added
18.07.24 HM	1. Intended use	Corrections according to the template
	3. Content	"Indications and contra-indications for use of the reagent kit" subsection was added
	4. Additional requirements	The phrase "for single use" was deleted
	5. General precautions	The Authorized representative in the European Community was specified for the contact in case of undesirable effects when using the reagent kit
	6. Sampling and handling	The intended use was specified. The list of biological material was expanded.
	8. Protocol	Information about the number of extraction reactions while working with nail plates and hair follicles was added
	12. References	The list of additional requirements was specified.
	Through the text	Information about hazards and precautionary statements list for Solution for lysis was rewritten according to the Regulation (EU) 2020/878
17.02.25 HM	5. General precautions	List of potentially interfering substances was expanded
		Working procedure was rewritten
		References were renewed
		Corrections according to the template
		Information about hazards for Washing Solution 4 was corrected

AmpliSens®



Ecoli Dx, s.r.o., Purkyňova 74/2
110 00 Praha 1, Czech Republic
Tel.: +420 325 209 912
Cell: +420 739 802 523



Federal Budget Institute of
Science "Central Research
Institute for Epidemiology"
3A Novogireevskaya Street
Moscow 111123 Russia