

# Certificate of Analysis

Issue date	2025-02-12
Product	intype® IC-RNA
Cat. Number(s)	IC289970
Lot number	F202500029
Expiry date	2027-02-28 (yyyy-mm-dd)
Storage temperature	-30°C to -15°C

## Quality Control

intype® IC-RNA as amplification control

Applied test system: real-time RT-PCR

Sample	Sample Type	Acceptance Criteria Ct Range <sub>HEX</sub>	Internal Control Ct <sub>HEX</sub>	Result	Passed
					✓
Sample 1	RNase-free water	24 – 29	26.40	positive	✓
Sample 2	buffer	24 – 29	26.87	positive	✓

Batch release thermal cycler: Stratagene™ Mx3005P Real-Time PCR System (Agilent Technologies, Inc.)

Manufactured and tested at INDICAL BIOSCIENCE GmbH, Leipzig, Germany, and in compliance with our ISO 9001 certified quality management system. The product fulfills all INDICAL quality requirements.

**Kristin Wappler**, Quality Assurance

This certificate is a computer printout and therefore valid without signature.