Intraoral Scanner

Instructions for Use

Original

Model 100

Model 001





CONTENTS

1.	About this guide	3
	1.1 Disclaimer	3
	1.2 License, trademarks and other rights	4
2.	Introduction and overview	6
	2.1 Intended use	6
	2.2 Device description and features	б
	2.3 Accessories and products used in combination	7
	2.4 Indications	8
	2.5 Precautions	8
	2.6 Compatibility information	10
	2.7 Data protection	11
	2.8 Further information	12
	2.9 Disposal	15
З.	Installation	16
	3.1 Unpacking the scanner	16
	3.2 Setting up the scanner	17
4.	Basic principles and user interface	23
5.	Scanning	28
	5.1 Accuracy	28
	5.2 Step-by-step	28
б.	Maintenance	32
	6.1 Handling	32
	6.2 Reprocessing	34
	6.3 Data backup	36
	6.4 Software updates	36
	6.4 Maintenance tools	37
	6.5 Changing fine fuses	40
	6.6 Service	41
7.	Troubleshooting	42
8.	Packing the scanner	45
9.	Distributors and service	47
10	. Technical data and label	48
11.	. Explanation of symbols	56
Ар	pendix A - Reprocessing of the handpiece sleeve	58
Ар	pendix B - Reprocessing of the Intraoral Scanner (without handpiece sleeve)	63
Ар	pendix C - Electromagnetic Compatibility Guideline	68

1. About this guide

<u>∧</u> Caution

The provided instructions contain important information on the safe and appropriate use of your scanner and its software. Make sure to have read and understood this document prior to installing and operating the device.

These Instructions for Use are valid for the product life cycle of your device and its software unless new instructions are published within this period. In case of questions, contact your local distributor. Please keep this document for future reference.

The term "scanner" refers to any and all Intraoral Scanner models, including their parts and accessories. The term "scanner unit" refers to the scanner component used to process scans. The term "handpiece" refers to the handpiece with its detachable sleeve component.

Electronic Instructions for Use

Instructions for Use for your scanner are provided in electronic form. They are stored on your scanner and also available through the eIFU website. If a paper version is required, the user may print these instructions himself/herself or request a paper version from the manufacturer. For further information, delivery time and contact details see eIFU website: https://ifu.dentalwings.com

1.1 Disclaimer

This device is designed for use by persons who have appropriate knowledge in dentistry and dental laboratory technology.

The user of this device is solely responsible for determining whether or not this device is suitable for a particular patient case and circumstances. The user is solely responsible for the correctness, completeness and adequacy of all data entered into this device and the used software. The user must check the correctness of the results and assess each individual case.

Dental Wings products must be used in accordance with their accompanying instructions for use. Improper use or handling of Dental Wings products will void the warranty, if any, given for Dental Wings products. If you require additional information on the proper use of Dental Wings products, contact your local distributor. The user is not allowed to modify Dental Wings products. DENTAL WINGS, ITS AFFILIATES OR DISTRIBUTORS DISCLAIM REPRESENTATIONS AND WARRANTIES OF ANY KIND, WHETHER EXPRESSED OR IMPLIED, WRITTEN OR ORAL, WITH RESPECT TO THE PRODUCTS, INCLUDING ANY WARRANTY OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, ERROR-FREE OPERATION OR NON-INFRINGEMENT, AND THE PRODUCTS ARE SOLD "AS IS".

Our maximum liability arising out of the products or their use, whether based upon warranty, contract, tort or otherwise, shall not exceed the actual payments received by us in respect of the purchase price thereof. In no event shall we be liable for special, incidental or consequential damages, including, but not limited to, loss of profits, loss of data or loss of use damages, arising hereunder or from the sale of the products.

Updates to hardware and software components are made regularly. Therefore, some of the instructions, specifications and pictures contained in these Instructions for Use may differ slightly from your particular situation. Dental Wings reserves the right to revise or make changes or improvements to this product or documentation at any time without obligation to notify any person of such revisions or improvements.

Third party products and product data

When scanning and designing cases, the user may work with third party products and/or product data. Dental Wings, its subsidiaries, or distribution partners disclaim any liability for damages in connection with or arising from such third party product data or products being used within this scope.

1.2 License, trademarks and other rights

Software license

Your scanner software is protected by a license agreement. The scanner software may be used or copied only in accordance with the terms of this agreement. It is illegal to copy or use the software on any medium except those permitted in the license agreement. The agreement is available through the main menu by selecting the *Configuration* icon, then *Agreement* from the context menu (see Basic principles and user interface, p. 23).

Availability

Some of the products and accessories mentioned in these Instructions for Use may not be available in all countries.

Third-party software

Your scanner software contains code from third-parties which is provided under specific licenses. License conditions are available within the software through the main menu by selecting the *Configuration* icon, then *About* from the context menu (*see Basic principles and user interface, p. 23*).

Trade names and trademarks

DENTAL WINGS and/or other trademarks and registered trademarks from Dental Wings mentioned herein are the trademarks or registered trademarks of Dental Wings. All other trademarks are the property of their respective owners.

Copyright

Dental Wings documents may not be reprinted or published, in whole or in part, without the written authorization of Dental Wings.

2. Introduction and overview

2.1 Intended use

The Intraoral Scanner is a dental 3D scanner intended to be used to digitally record topographical characteristics of teeth and surrounding tissues. The system produces 3D scans for use in computer-assisted design and manufacturing of dental restorations.

2.2 Device description and features

The Intraoral Scanner consists of a computer with embedded software, a monitor, and a handpiece equipped with optical sensors.

Variants

The Intraoral Scanner (Type: DW-IO) is delivered in two product variants (see *Instructions for Use*, p. 1):

- Model 001 (cart) Article No. 6000-1
- Model 100 (laptop) Article No. 6300-1

The delivery contains all components necessary to run the Intraoral Scanner as intended. For details, *see Parts identification*, *p. 18.*

D Notes

1. The laptop (scanner unit of Model 100) can be replaced by the scanner unit of Model 010 (portable).

2. Distributors may acquire laptops for Model 100, install the software, and validate the installation. For further details, *see Technical data and label, p. 48.*

<u>∧</u> Caution

Laptops must comply with national safety regulations and be validated for use with the Intraoral Scanner to ensure adequate performance. Refer to the information available for distributors.

Configurations

The device is delivered in the following configurations:

- With the distributor logo and color
- With a distributor-specific set of accessories
- With a distributor-specific software configuration
- With a country-specific power cord

Principles of operation

Sensors in the handpiece of the Intraoral Scanner provide data, which is matched to each other in the processing unit to create a digital impression.

Novel features

The Intraoral Scanner does not contain novel features in relation to conventional dentistry. It supports digital realization of conventional dental services in the area of restorative prosthetic.

2.3 Accessories and products used in combination

The use of the Intraoral Scanner requires accessories (cleaning and disinfection products, standard packaging material). Moreover, it can be used in combination with other products (e.g. CAD/CAM software).

Dental Wings has validated the specific products given in chapter *Technical data and label (see Replacement parts, p. 53)*, in *Appendix A - Reprocessing of the handpiece sleeve (p. 58)* and in *Appendix B - Reprocessing of the Intraoral Scanner (without handpiece sleeve) (p. 63)*. Others may be suitable as well if they meet the requirements below. However, the user has to validate their compatibility with the Intraoral Scanner.

The scanner is available in different configurations. For device variants and optional PC peripherals, see chapter *Device description and features (p. 6)*. For information on specific accessories, contact your local distributor.

Reprocessing

The Intraoral Scanner requires reprocessing. The particular use, specific requirements, as well as validated products for the reprocessing material and equipment are found in Appendices A and B.

Cleaning and disinfection products : The scanner requires reprocessing with specific detergents and disinfectants. They are specified in Appendices A and B.

The handpiece sleeve is a semicritical area and requires sterilization. The handpiece (without sleeve), the cable, the scanner unit, and the accessories are noncritical areas and require intermediate or low level disinfection.

Standard packaging material : Standard see-through sterilization pouches shall be used to prepare the handpiece sleeve for sterilization or to protect the handpiece when stored on the Intraoral Scanner. A size of 9 cm x 20 cm is required. In addition, a standard plastic bag may be used to store the handpiece when detached from the scanner. A minimum size of 25 cm x 40 cm is required.

Further use of 3D scans

3D scans from the Intraoral Scanner may be used with other software products. It is required that such products can handle open STL files. Dental Wings has validated the use of:

DWOS: 3D dental restorations design software

Straumann® CARES® Visual: 3D dental restorations design software

coDiagnostiX: Dental surgery treatment planning software

2.4 Indications

The scanner is an intraoral optical 3D scanning device intended to be used to digitally record topographical characteristics of teeth and surrounding tissues. The system produces 3D scans for use in computer-assisted design and manufacturing of dental restorations. The 3D scan is provided in a 3D geometry format. The device allows several scanner heads embedded in the handpiece to scan at once from multiple angles. It is intended to be used by dental professionals.

2.5 Precautions

The provided instructions contain important information on the safe and appropriate use of your scanner and its software. Make sure to have read and understood this document prior to installing and operating the device.

The sleeve must be assembled to the handpiece before scanning. If it is not assembled correctly, the device won't scan properly.

Laptops must comply with national safety regulations and be validated for use with the Intraoral Scanner to ensure adequate performance. Refer to the information available for distributors.

This scanner is capable of very precise scanning. To achieve scanning accuracy, comply with all the given instructions.

The Intraoral Scanner (cart model) is a protection class I device. To avoid risk of electric shock it shall only be connected to a power outlet with protective earthing.

The Intraoral Scanner (laptop model) is a protection class II device. Always use the power supply provided.

The scanner is suitable for use in laboratories, dental offices and equivalent environment. However, the protection against dust and water has not been specifically tested.

Install the scanner in accordance with the instructions. Only use the parts provided or specified. Before any maintenance action, unplug the scanner from the power outlet.

The handpiece must always be connected to the scanner through the handpiece connector. Make sure the handpiece is plugged as instructed before it comes into contact with patients.

Make sure the building and its electrical circuitry is grounded and the handpiece is never connected directly to the laptop when using Intraoral Scanner (laptop model).

Keep the original packaging of the scanner for safe transportation in case of service, repair and disposal. Unpack and pack your scanner following the instructions.

The Intraoral Scanner is only compatible with Dental Wings handpieces. Make sure you are connected to the internet when plugging a handpiece for the first time.

Reprocessing shall be done before the first use, then after each use, as well as prior to maintenance, service, repair, and disposal to prevent cross-contamination.

Always ensure a biunique relationship between patient, scan and treatment case.

Protection of the data entered into the software is a shared responsibility. Please refer to the license agreement for further details.

Backup your data at regular intervals. It is the responsibility of the user to perform and maintain data backups in order to prevent loss of data.

The user is responsible to assure health data de-identification, data backup and disaster recovery, and health data authenticity.

The user is responsible to configure the operating system of the Intraoral Scanner, and as applicable integrate it in its network, as required by its IT Policy. This includes set up of automatic logoff, audit controls, authorizations, configuration of security features, emergency access, malware detection/protection, node authentication, personal authentication, physical locks, integration of the Intraoral Scanner in the product life cycle roadmap, system and application hardening, health data storage confidentiality. In case logs for auditing or emergency access related to services provided by Dental Wings are required, contact your local distributor.

Certain functionalities and services require data transfer. Access to data, storage and transfer shall comply to national regulations on information security. For further details on encryption methods, type of data transferred, location of hosts, and storage refer to https://dentalwings.com/data-protection-faqs/.

Online updates are used to solve safety and/or information security issues. Such updates are mandatory. Such will be clarified by the information provided with the online update.

Protect your data against loss, unauthorized access and unauthorized use.

Model 001 (cart): Installing additional software may compromise system integrity, slow down performance, and create compatibility issues.

Model 100 (laptop): Install only software validated to be used in combination with the scanner.

Always handle the handpiece with great care as it contains fragile components. Dropping the handpiece and its sleeves on the floor may damage it permanently.

The handpiece heats up over normal body temperature, but it does not pose a hazard to patients due to short term exposure and small contact area.

Always keep a 60 mm distance between eyes and the handpiece.

Make sure the scanner is placed in a safe and stable position.

Never put any load on the scanner.

Make sure to comply with all applicable standards and technical regulations for any computer hardware used. Computer hardware is not intended to be used in close proximity (within a distance of 1.5 meters) to the patient. Do not use extension cords or multiple-socket extension cords to plug the hardware.

Do not dispose of this device as unsorted municipal waste. Collect separately.

2.6 Compatibility information

The device is compatible with the Intraoral Scanner software application version 3.0 and higher.

2.7 Data protection

<u>∧</u> Caution

Protection of the data entered into the software is a shared responsibility. Please refer to the license agreement for further details. *(see License, trademarks and other rights, p. 4)*

The user is responsible to configure the operating system of the Intraoral Scanner, and as applicable integrate it in its network, as required by its IT Policy. This includes set up of automatic logoff, audit controls, authorizations, configuration of security features, emergency access, malware detection/protection, node authentication, personal authentication, physical locks, integration of the Intraoral Scanner in the product life cycle roadmap, system and application hardening, health data storage confidentiality. In case logs for auditing or emergency access related to services provided by Dental Wings are required, contact your local distributor.

Access to the software is restricted by the licensing system.

To protect data against loss or unauthorized use, several security mechanisms have been implemented:

- Data encryption during storage and transfer through the internet
- An archiving system to backup patient and planning data
- An on-demand anonymization function for privacy protection

Bear in mind that the user is responsible for data protection on user side. Make sure to adequately protect such data.

<u>∧</u> Caution

The user is responsible to assure health data de-identification, data backup and disaster recovery, and health data authenticity.

Protect your data against loss, unauthorized access and unauthorized use.

- Secure your computer system by installing a malware scanner or firewall.
- Use a strong password to protect your computer system, storage media and presentation devices.
- Use data encryption to secure data on your computer system and storage media.
- Backup your data regularly to avoid loss of data.
- Use the anonymization function to protect patient personal data if required.

<u>∧</u> Caution

Certain functionalities and services require data transfer. Access to data, storage and transfer shall comply to national regulations on information security. For further details on encryption methods, type of data transferred, location of hosts, and storage refer to https://dentalwings.com/data-protection-faqs/.Protect your data against loss, unauthorized access and unauthorized use.

2.8 Further information

Dental Wings and its products are registered in the main markets. Please be aware that any serious incident that has occurred in relation to the Intraoral Scanner should be reported to your local distributor and the competent authority responsible for medical devices or data protection in your country. If you have any questions, please contact your local distributor.


Radiation information

The handpiece contains LEDs. They are rated *Exempt* according to applicable standards *(see Technical data and label, p. 48)*.

<u>∧</u> Caution

Always keep a 60 mm distance between eyes and the handpiece.

The Intraoral Scanner (cart model) contains a gesture recognition control, which includes an infrared laser projector. It is rated *Class 1* according to applicable standards *(see Technical data and label, p. 48)*.





<u>∧</u> Caution

Usage, control, adjustment or execution of the infrared laser projector other than specified below may result in hazardous exposure to radiation.

Usage of the gesture recognition module is specified in *Gesture control - Model 001 (cart) (p. 26)*. The gesture recognition module (SR300) contains an infrared laser projector and shall not be repaired, modified or disassembled. In case of failure, contact your local distributor. Do not try to open any parts of the SR300 assembly. The SR300 does not have any parts that shall be accessed, changed or adjusted by the user. Modification or dissembling of SR300, especially the infrared laser projector, may result in an emission over class I. The manufacturer of the Intraoral Scanner will notice the manufacturer of the SR300 assembly or return the module before conducting any failure analysis.

Electromagnetic compatibility information

The Intraoral Scanner is suitable for use in professional home healthcare environment and does not execute any essential performance function in the understanding of the applied product standards. Nevertheless, it had been tested for electromagnetic compatibility (EMC) and needs special precautions.

The Intraoral Scanner needs to be installed and put into service according to the Electromagnetic Compatibility Guideline *(see Appendix C - Electromagnetic Compatibility Guideline, p. 68).*

⚠ Caution

The use of accessories other than those specified in *Technical data and label (see Replacement parts, p. 53)* may result in increased emission or decreased immunity of the Intraoral Scanner.

The Intraoral Scanner allows for wireless network connection. However, it may be interfered with radio frequency (RF) sources in the vicinity of the device (e.g, electromagnetic security systems, cellular telephones, radio frequency identification (RFID) or other in-band transmitters), even if that other equipment complies with the International Special Committee on Radio Interference (CISPR) emission requirements.

Portable and mobile RF communications equipment may affect the performance of Intraoral Scanner. Please avoid strong electromagnetic interference, e.g. nearby mobile phones, microwave ovens during the use of the Intraoral Scanner.

The Intraoral Scanner should not be used adjacent to or stacked with other equipment and that if adjacent or stacked use is necessary, the Intraoral Scanner should be observed to verify normal operation in the configuration in which it will be used.

Other portable radio frequency communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the Intraoral Scanner, including cables specified by the manufacturer. Otherwise, degradation of the performance of this device could result.

2.9 Disposal

<u>∧</u> Caution

Reprocess prior to disposal to prevent cross-contamination.

Electrical and electronic devices must be disposed of separately from normal household waste in order to promote reuse, recycling and other forms of recovery, to avoid adverse effects on the environment and human health as a result of the presence of hazardous substances in some of their components and to reduce the quantity of waste to be eliminated with a view to reducing landfill. This includes accessories such as remote controls, power cords, etc.



<u>∧</u> Caution

Do not dispose of this device as unsorted municipal waste. Collect separately.

For details on how to dispose of your device and its packaging, please contact your local distributor.

3. Installation

<u>∧</u> Caution


Keep the original box and packaging material for safe transportation of the device in case of service, repair and disposal. Unpack and pack your scanner following the instructions.

For instructions on how to pack the scanner, see Packing the scanner (p. 45).

3.1 Unpacking the scanner

Store the shipping box in a safe place in case you need to transport or ship the scanner in the future.

Model 001 (cart)



Model 100 (laptop)




Figure 3: Steps for unpacking the scanner – Model 100

3.2 Setting up the scanner

<u>∧</u> Caution

Make sure you have read and understood these instructions for use before installing and operating the scanner. Comply with all precautions and safety information.

Install the scanner in accordance with the manufacturer's instructions. Only use the parts provided or specified.

The scanner is suitable for use in laboratories, dental offices and equivalent environment. However, the protection against dust and water has not been specifically tested. *See Technical data and label, p. 48.*

Make sure the building and its electrical circuitry is grounded and the handpiece is never connected directly to the laptop when using Intraoral Scanner (laptop model).

Do not install the scanner in places with high temperature differences, near heat sources, in direct sunlight or in damp environment (e.g. exposed to splash water). Do not expose this device to rain or moisture. Do not install the scanner in dusty environment.

Do not block any ventilation openings.

Turn off the scanner before plugging or unplugging the power cable.

Always handle the handpiece with great care as it contains fragile components. Dropping the handpiece and its sleeves on the floor may damage it permanently.

Parts identification




Figure 4: Identification of the scanner parts


🛄 Note

The laptop (scanner unit of Model 100) can be replaced by the scanner unit of Model 010 (portable).

1.	Scanner	
2.	Screen	
З.	Handpiece	
	a. optics	
	b. sleeve	
	c. light ring	
4.	Cable hook	
5.	Handle	
6.	Wheel lock	
7.	Integrity-check sleeve	
8.	Power switch	
9.	Power button	
10.	Gesture/voice recognition	
	module	
11.	Power cord/power supply	
12.	Holder	
13.	Handpiece connector power	
	supply	
14.	USB cable	
15.	Handpiece connector	
16.	SharpAdjust	

Connecting the handpiece

- Plug the handpiece cable into the handpiece connector at the back of the screen.
- Plug the handpiece cable into the handpiece connector.
- Plug the handpiece connector via USB into the socket on the laptop.






Figure 5: Plugging the handpiece – Model 001 (cart)

Figure 6: Plugging the handpiece – Model 100 (laptop)

<u>∧</u> Caution

The handpiece must always be connected to the scanner through the handpiece connector.

Make sure the handpiece is plugged as instructed before it comes into contact with patients.

🛄 Note

Perform the integrity check to validate installation or when the handpiece is dropped or subject to a shock, to confirm that it still works as specified.

Disconnecting the handpiece

If the scanner unit is shut down, disconnect the handpiece by unplugging the USB from the socket.

If the scanner unit is still on, disconnect handpiece as follows:

- 1. Select the *Configuration* icon of the main menu, then the *Handpiece Management* context menu.
- 2. Select the *Disconnect* button.
- 3. To release the cord, unplug the USB from the socket.

Assembling the handpiece sleeve

Slide the sleeve on the handpiece (without sleeve) and rotate clockwise to lock into position.

Disassembling the handpiece sleeve

Rotate the sleeve counterclockwise to unlock and slide it off the handpiece.



Figure 7: Assembling and disassembling the handpiece sleeve

<u>∧</u> Caution

The sleeve must be assembled to the handpiece before scanning. If it is not assembled correctly, the device won't scan properly.

Never leave the handpiece elsewhere than in its dedicated handpiece holder.

Plugging the scanner unit

\land Caution

The Intraoral Scanner (cart model) is a protection class I device. To avoid risk of electric shock it shall only be connected to a power outlet with protective earthing.

The Intraoral Scanner (laptop model) is a protection class II device. Always use the power supply provided.

Make sure to comply with all applicable standards and technical regulations for any computer hardware used. Computer hardware is not intended to be used in close proximity (within a distance of 1.5 meters) to the patient. Do not use extension cords or multiple-socket extension cords to plug the hardware.

Protect the power cord from being walked on and from pinching.

Model 001 (cart)

- 1. Plug the power cord in the back of the scanner.
- 2. Plug the power cord into the power outlet.
- 3. Flip the power switch in the *power on* position (see Scanner controls, p. 23)
- 4. Press the power button (see Scanner controls, p. 23)

Model 100 (laptop)

- 1. Plug the power cord in the back of the laptop.
- 2. Plug the power cord into the power outlet.
- 3. Press the power button (see Scanner controls, p. 23)

Refer to instructions for using the device (see Basic principles and user interface, p. 23).

Unplugging the scanner unit

Before any maintenance on the scanner, it should be unplugged as follows:

Model 001 (cart)

- 1. Turn off the device (see Shutting down the scanner, p. 25).
- 2. Flip the power switch in the *power off* position.
- 3. Unplug the power cord from the power outlet.
- 4. Loop the cord around the hook(s).

Model 100 (laptop)

- 1. Turn off the device (see Shutting down the scanner, p. 25).
- 2. Unplug the power cord from the power outlet.

System language

The system language is preset. If this is not the case, contact your local distributor.

4. Basic principles and user interface

Scanner controls


Power switch: used to power on the scanner unit




Position (down)	Scanner status
I	Power on
0	Power off

Model 001 (cart)

Power button: used to start up the scanner unit





Status	Scanner status
Light on	Turned on
Light off	Turned off

Model 001 (cart)

Model	100	(laptop)

Light ring: provides the user with a feedback of handpiece's status



Color	Handpiece status
Blinking, white to blue	Preheating
Blue	Preheated
Green	Scanning & acquiring data
Red	Scanning & not acquiring data

Turning on the scanner

▲ Caution

Turn on the device only with the handpiece fully assembled (with handpiece sleeve).

Model 001 (cart)

- 1. Press the power switch in the *power on* position.
- 2. Press the power button.
- 3. Select your user icon and type in your PIN.

Model 100 (laptop)

- 1. Press the power button.
- 2. Select your user icon and type in your PIN.

First use

When the scanner is activated for the first time, a wizard pops up to assist the user with the basic settings: create administrator account, create users, set credentials to local network and other scanner settings.

User interface

The user interface is mainly icon-driven and provides direct instructions for performing the scans. Here are the main items that you can access from the main menu that is displayed on the left side of the screen. Each menu has its own context menu to access its specific functionalities.



Plan, Scan, Align, Review or Export.



Data Transfer (WiFi)

The Intraoral Scanner can be used offline, as a standalone device. It can also be connected to a wireless network to transfer, share, or back up data.

<u> (</u>Caution

Use protected WiFi setup and access. Keep a distance to other communication devices of 13 cm when you are using 2.4 GHz and 19 cm when you are using 5 GHz.

Model 001 (cart)

- 1. Starting outside of the screen, swipe your finger from the left side to the center of the screen.
- 2. You now have access to the menu of the operating system.
- 3. Follow operating system instructions to connect to wireless networks.

Model 100 (laptop)

Follow operating system instructions to connect to wireless networks.

Shutting down the scanner

<u>∧</u> Caution

Wait until the scanner is completely shut down before unplugging the power cord.

Model 001 (cart)

- 1. Click the top icon in the main menu.
- 2. Select Shutdown.
- 3. Confirm (a warning pops up if a case is uploading to DWOS Connect).
- 4. Flip the power switch in the *power off* position if the power cord needs to be unplugged.

Model 100 (laptop)

Follow shutdown instructions for your operating system.

Gesture control - Model 001 (cart)

The cart model is equipped with a gesture recognition module. It works with an infrared laser sensor [for safety information, refer to *Further information (p. 12)*] which is located above the screen as shown on *Figure 4: Identification of the scanner parts (p. 18)*. Using gesture control instead of the touch-screen prevents cross-contamination during scanning.

Step	Hand gesture	Triggered action
1	Place hand wide open in front of the sensor.	Activates the gesture control feature.
2	Point one finger upwards.	Displays context menu.
3	Move hand wide open from side to side.	To navigate inside the context menu.
4	Point one finger upwards.	Stops the navigation. You can therefore move your hand without affecting the selection.
5	Move hand towards the sensor while the finger is still pointing upwards.	Enters the selection.
6	Move hand wide open in any direction.	Moves the scan in 3D as if it was linked to your hand.

Voice control - Model 001 (cart)

The cart model is equipped with a voice recognition module. It works with a microphone which is located as shown on *Figure 4: Identification of the scanner parts (p. 18)*. Using voice control instead of the touch-screen prevents cross-contamination, and lets you scan without moving your hands away from the patient or handpiece. The commands are listed below with their associated action.

Voice command	Triggered action
io start	Launches the scan for the current element (upper, lower or occlusion).

io pause	Pauses the current scan (re-launch with the start command).
io next	Switches to the next step.
io reset	Erases current data.
io undo	Removes surface data in a chronological fashion.
io redo	Restores surface data in a chronological fashion.
iolock	Protects the current scan against modification. The locked surfaces turn into a reddish color.
io unlock	Releases the locked surfaces so that they can again be modified. They turn back into their original color.
io screenshot	Takes a screenshot of the current view and attaches it to the order.
io take photo	Within the <i>Add Picture</i> feature of the <i>Plan Editor</i> . takes a photo from the screen built-in camera or from the handpiece camera, according to which one is active.
io detect	Detects the scanbody when the handpiece is pointing at it during scanning.

5. Scanning

5.1 Accuracy

This scanner performs highly accurate scans. For scanning accuracy as measured according to ISO 20896-1 and Dental Wings testing standard, see *Integrity check (p. 37)*.

⚠ Caution

To achieve scanning accuracy, comply with all given instructions.

Considering that the final accuracy and fit of the restoration highly depend upon the preciseness of the user's manipulations and judgment, the clinical outcome is easily influenced by the following factors ¹:

- Scanning too fast
- Jaws shifting during occlusion scan
- Inadequate occlusion alignment

5.2 Step-by-step

Setup

<u>∧</u> Caution

Make sure the scanner has been reprocessed as described in *Reprocessing (p. 34)*.

¹ Certain error sources can contribute with deviations in mm range.

- Select a User or a Dentist.
- Enter security PIN.
- Select New case.
- Enter case description, including tooth number, restoration type, material and shade.
- Set *Appointment Date* for the preparation.
- Set *Due Date*, as the target delivery for the restoration at your clinic.
- Specify if a *Pre-Preparation Scan* is desired.
- Select Patient (create or search).
- Add notes and attach files as desired.
- Select Next.

Preheat the handpiece prior to scanning. When the light ring turns red or green, the handpiece is ready for scanning.

<u>∧</u> Caution

Always ensure a biunique relationship between patient, scan and treatment case.

Scan

<u> </u>Caution

Scan the area of interest until sufficient data is collected.

The handpiece heats up over normal body temperature, but it does not pose a hazard to patients due to short term exposure and small contact area. *See Temperature limits, p. 48.*

- Prepare for scanning in the same way as necessary for taking a physical impression. The area to be scanned shall be dry (e.g. no bleeding, no saliva).
- Lift the handpiece from its holder.
- Remove protective packaging from the handpiece, if any.
- Scan (and re-scan as necessary) the area of interest. Scanning is most effective at 5 mm from handpiece sleeve. Avoid contact to breached or otherwise compromised surfaces.
- Scan upper and lower arches as well as bite, following on-screen instructions.
- Select *Next* after each arch.
- Replace the handpiece on its holder.

For further details regarding scanning techniques, visit our website.

Alignment

\land Caution

Verify that upper and lower arches are identified as such.

- Align upper and lower scans onto the bite scan, following on-screen instructions.
- If needed, invert upper and lower arches by selecting *Swap Arches*.
- If needed, switch quadrant by selecting *Switch Sides*.
- Adequate superimposition can be observed when the surface presents an evenly marbled pattern between both colors. Confirm by selecting *Next*.

Review

- Confirm complete occlusion by observing sufficient red areas.
- Identify planned preparations on the scanning data.
- Confirm adequate occlusal spacing on each preparation for restoration material using the gradient slider.
- Confirm absence of undercuts on each preparation by simulating the insertion axis.
- If desired, specify part or complete margin line on each preparation.

Export to DWOS Connect

- In the *Export* screen, verify the case overview.
- Select collaborating laboratory from the drop-down list.
- Review or complete attachments and notes.
- Select Export.

Back to the home screen, a progress bar is displayed until the case is sent. Wait until the progress is completed before turning off the device.

Export to STL

- Click the *Case database* icon from the main menu.
- On the case to export, swipe your fingers towards the right side of the screen.
- Select *Export STL*.
- Select the destination folder.

6. Maintenance

6.1 Handling

Store handpiece, scanner unit, and SharpAdjust according to the specified ambient conditions (see Ambient conditions, p. 51).

Use standard packaging material to protect the handpiece and accessories from contamination between uses *(see Accessories and products used in combination, p. 7).*

Unplug the device during lightning storms or when unused for long periods of time.

When not in use, SharpAdjust shall always have the sleeve cap on, and be kept in the original packaging.

🛄 Note

The product warranty does not apply if damage to the product is the result of a lightning strike.

Never put any load on the scanner.

Make sure the scanner is placed in a safe and stable position.

The surface the scanner is placed on shall be intended to carry the weight.

Always lock the wheels when not moving the cart model. Push down the tabs for locking the wheels.

Before any maintenance action, unplug and reprocess the scanner.

To reduce the risk of electric shock, do not remove the cover. Refer servicing to qualified service personnel.

Protect the handpiece from contamination between uses.

Ensure that the packaging is large enough to contain the handpiece without excessively bending the cable.

Model 001 (cart): Installing additional software may compromise system integrity, slow down performance, and create compatibility issues.

Model 100 (laptop): Install only software validated to be used in combination with the scanner.

Moving the scanner to another room

Model 001 (cart)

- Shut down according to instructions (see Shutting down the scanner, p. 25)
- Flip the power switch in the *power off* position.
- Unplug the cord from the power outlet and loop it around the hook(s).
- Unlock the wheels by lifting up the red tab on each one of them.
- Move the device while holding it by the handle *(see Identification of the scanner parts, p. 18).*
- When arrived at destination lock each wheel by pushing down the tab.

Model 100 (laptop)

- Shut down according to instructions (see Shutting down the scanner, p. 25)
- Unplug the power cord from the outlet.
 You can now move the laptop to a new location.

🛄 Note

If the handpiece is dropped or subject to a shock, run the integrity check to confirm that it still works as specified. *See Integrity check, p. 37.*

Transportation

Always use the original packaging to transport the scanner. Refer to instructions for *Packing the scanner (p. 45)* and *Unpacking the scanner (p. 16)*.

6.2 Reprocessing

Reprocessing means the processing of re-usable medical devices. It includes as applicable:

- preparation at the point of use;
- preparation, cleaning, disinfection;
- drying;
- inspection, maintenance and testing;
- packaging;
- sterilization;
- storage.

<u>∧</u> Warning

Reprocessing shall be done before the first use, then after each use, as well as prior to maintenance, service, repair, and disposal to prevent cross-contamination.

<u>∧</u> Caution

Before any reprocessing, unplug the scanner and SharpAdjust from power outlet.

Always handle the handpiece with great care as it contains fragile components. Dropping the handpiece and its sleeves on the floor may damage it permanently.

Make sure the optics are cleaned after disinfection as follows:

- 1. Disassemble the handpiece sleeve (see Assembling and disassembling the handpiece sleeve, p. 20).
- 2. Clean the optics on the handpiece sleeve and handpiece (without sleeve) *(see Figure 8: Cleaning the optics of handpiece and sleeve, p. 35).*
- 3. Wear gloves. Use a soft, lint-free, non-abrasive, absorbent, disposable cloth moistened with alcohol. Wipe the optics once in one direction. Repeat if needed.





Figure 8: Cleaning the optics of handpiece and sleeve

The Intraoral Scanner consists of semicritical and noncritical areas (see Figure 9: Disinfection levels of handpiece sleeve/scanner, p. 35). Semicritical are the areas contacting mucous membranes. Noncritical are the areas contacting skin, but not mucous membranes. The materials used in the semicritical areas are biocompatible. The semicritical area requires sterilization. The noncritical areas require intermediate or low level disinfection.

Noncritical areas: Intermediate or low level disinfection





Intraoral Scanner (without handpiece sleeve)

Handpiece sleeve

Figure 9: Disinfection levels of handpiece sleeve/scanner

Appendix A - Reprocessing of the handpiece sleeve and Appendix B - Reprocessing of the Intraoral Scanner (without handpiece sleeve) contain further details on how to reprocess..

The instructions given have been validated by Dental Wings as being capable of preparing the Intraoral Scanner for re-use (250 cycles). It remains the responsibility of the user to ensure that the reprocessing as actually performed using equipment, material and personnel in the dental or reprocessing facility achieves the desired result. This requires validation and routine monitoring of the process. Likewise, any deviation by the user from the instructions provided is performed under the sole responsibility of the user and should be properly evaluated for effectiveness and potential adverse consequences.

6.3 Data backup

<u>∧</u> Caution

Backup your data at regular intervals. It is the responsibility of the user to perform and maintain data backups in order to prevent loss of data.

The Intraoral Scanner performs automatic backups of the database, as long as you define a destination for it:

1. Log in as Administrator.

2. Select the Configuration icon from the main menu, then the System Settings context menu.

3. In the field *Backup target* directory, type in a location or click the folder icon to browse to the desired destination for storing the files created by the automatic backup feature.

6.4 Software updates

The user is automatically notified about pending updates. Update the software as indicated to benefit from continuous improvements and to maintain regulatory compliance.

<u>∧</u> Caution

Online updates are used to solve safety and/or information security issues. Such updates are mandatory. Such will be clarified by the information provided with the online update.

6.4 Maintenance tools

The Intraoral Scanner is a self-maintenance device. There are two maintenance tools available:

- The integrity-check sleeve is a non-medical, non-electrical component delivered with all models of the Intraoral Scanner.
- SharpAdjust is a non-medical, electrical component of the Intraoral Scanner, with its own label. SharpAdjust is delivered to users upon request.

Users are prompted by the software to periodically apply the integrity-check sleeve to validate the integrity of the optics. If the integrity cannot be validated, users are guided to run SharpAdjust to fine-tune the optics.

Integrity check

The integrity-check sleeve is a maintenance tool used to validate the integrity of the optics of the scanner.

<u>∧</u> Caution

Make sure the integrity-check sleeve is reprocessed as required before and after use. *See Appendix B* - *Reprocessing of the Intraoral Scanner (without handpiece sleeve), p. 63.*

- 1. Make sure the handpiece (without sleeve) is connected to the scanner unit (*see Disassembling the handpiece sleeve, p. 20*).
- 2. Select the *Configuration* icon of the main menu, then *Handpiece Management* context menu.
- 3. Slide the integrity-check sleeve on the handpiece and rotate clockwise to lock into position (see Figure 10: Using the integrity-check sleeve, p. 37).
- 4. Select the *Start test* button and wait for confirmation and further instructions to display on the screen.
- 5. When the process is over, rotate the integrity-check sleeve counterclockwise to unlock and slide it off the handpiece.




Figure 10: Using the integrity-check sleeve

D Note

The scanner does not require calibration. However, the integrity check should also be performed if the handpiece sleeve is dropped or subject to a shock, to confirm that it still works as specified.

SharpAdjust

SharpAdjust is a maintenance tool used to fine-tune the optical system of the Intraoral Scanner. It is applied in case the integrity check is unsuccessful two times in a row.

<u>∧</u> Caution

Make sure SharpAdjust is reprocessed as required before and after use. *See Appendix B - Reprocessing of the Intraoral Scanner (without handpiece sleeve), p. 63.*

When SharpAdjust is not in use, store it with the sleeve cap on, in the original packaging.

Parts identification




Figure 11: Identification of SharpAdjust parts

- 1. Cap
- 2. SharpAdjust sleeve
- 3. SharpAdjust sleeve positions
- 4. Dial switch
- 5. Switch positions

Connecting SharpAdjust

SharpAdjust is connected and ready to use if the light is on. If SharpAdjust is not working properly, contact your local distributor.

Model 001 (cart)

- 1. Plug the USB cable into SharpAdjust.
- 2. Plug the other end of the cable into a power outlet using a USB power supply.
- 3. Remove the cap from the SharpAdjust sleeve.




Figure 12: Plugging SharpAdjust for Model 001 (cart)

Model 100 (laptop)

- 1. Plug the USB cable into SharpAdjust.
- 2. Plug the other end of the cable into the laptop USB port.
- 3. Remove the cap from the SharpAdjust sleeve.




Figure 13: Plugging SharpAdjust for Model 100 (laptop)

Running SharpAdjust

- 1. Connect SharpAdjust to power outlet *(see Connecting SharpAdjust, p. 38).*
- 2. Make sure the handpiece (without sleeve) is connected to the scanner unit *(see Disassembling the handpiece sleeve, p. 20).*
- 3. Slide the handpiece into the SharpAdjust sleeve and rotate clockwise to lock into position.
- 4. Set SharpAdjust sleeve to position A and turn the dial switch to 1.
- Launch SharpAdjust by selecting the *Configuration* icon from the main menu, then *Handpiece Management* from the context menu. Follow on-screen instructions.
- 6. When the process is complete, rotate the handpiece counterclockwise to unlock and slide it out of the sleeve.




Figure 14: Using SharpAdjust to fine-tune the handpiece optics

Disconnecting SharpAdjust

- 1. Make sure that the handpiece is removed from SharpAdjust.
- 2. Unplug SharpAdjust.
- 3. Reprocess as described in *Reprocessing (p. 34)*.
- 4. Put the cap back on the SharpAdjust sleeve and store the tool in its original packaging.

6.5 Changing fine fuses

Model 001 (cart)

Model 001 (cart) of the Intraoral Scanner contains two fine fuses (*see Technical data and label, p. 48*). They are located at the back of the scanner unit, inside the power switch assembly (*see Figure 4: Identification of the scanner parts, p. 18*). If necessary, exchange them in the following way:

- 1. Using a small flathead screwdriver (approx. width 4 mm), lever out the insert.
- 2. Change the corresponding fine fuses.
- 3. Reassemble.








Figure 15: Instructions for changing fine fuses - Model 001 (cart)

6.6 Service

<u>∧</u> Caution

Reprocess prior to service to prevent cross-contamination.

To reduce the risk of electric shock, do not remove the cover. Refer servicing to authorized service personnel.

Do not replace components of the scanner without the prior authorization from your distributor. Refer servicing to authorized service personnel.

Servicing is required when the device has been damaged in any way, has been exposed to rain or moisture, does not operate normally, or has been dropped. Contact your local distributor.

The Intraoral Scanner does not require regular service. However, in certain countries it might be mandatory according to national law. A risked based approach taking into account the amount of usage is acceptable. It shall be conducted according IEC 62353.

When returning your scanner for service or repair, pack the scanner as described in *Packing the scanner (p. 45)*.

7. Troubleshooting

Find below a list of messages (some are abbreviated) that could appear while using the Intraoral Scanner, and the actions that should be taken to solve the problem.

On-screen message	Required action	
Start up and log in		
Cannot start setup	Destart the device and try again	
Credentials could not be verified.	Restart the device and try again.	
It seems that there is no valid license to run this software on your system	Contact your local distributor to obtain a license.	
The scanner initialization failed	Ensure that the handpiece is properly connected and restart the device.	
Hardware and configuration		
Integrity check unsuccessful		
SharpAdjust unsuccessful	Follow on-screen instructions or contact your local distributor.	
Integrity check failed because of an unforeseen error		
An error has occurred.	Restart the device and try again.	
No configuration file found for the connected handpiece.		
The configuration file does not match the connected handpiece.	 Check internet connection. Click the <i>Configuration</i> icon in the main menu and select the <i>Handpiece</i> <i>Management</i> context menu. In the Overview section, click on the <i>Download</i> icon. 	
The connected handpiece has been qualified for demonstration only, not for clinical use.	The scan accuracy of the connected handpiece is not sufficient for patient cases.	

On-screen message	Required action			
Database				
Case creation failed.				
could not be computed/imported	• Restart the device and try again.			
Unable to save	Restart the device and try again.			
Patient/Dentist update/deletion failed.	•			
Database backup execution failed.	Try to export in a different folder and/or a different file name.			
Implant kit / ipflib file errors.	Select a valid implant kit file.			
Software				
Server errors.				
Error loading 3D graphic resource.	-			
Unable to save the recorded message.	•			
Cannot recognize microphone.	Restart the device and try again.			
An error occurred while recording.				
Failed to convert case to XOrder	•			
Dentist assignment failed.	•			
Wrong client version.	Run the updater again.			
Case export failed.	Try to export in a different folder and/or a different file name.			
Scanning				
Scanned area too small for meshing.	Scan more tooth surface before proceeding.			
Unable to create mesh from scanned points.				
Could not swap arches	Clear scan data and rescan.			
Failed to create a temporary directory	Clean up hard drive.			
This case is not ready to export.	Complete the scan workflow before exporting the case.			
No scan to save.	Complete the scan procedure and retry.			

On-screen message	Required action
The scan has been interrupted to prevent overheating	Wait one minute to let the handpiece cool down and then resume scanning.
DWOS Connect	
No internet connection	Check internet connection.
Server not found/not reachable	Check Internet connection.
Authentication to DWOS Connect failed.	Try again later.
Failed to authenticate on DWOS Connect	Check DWOS Connect credentials.
Bad configuration for remote server.	Check that the DWOS Connect configurations are valid.
No remote server is configured.	Configure your case exchange server connection.
This case is not in queue.	Check if the case is in the upload queue in the workspace screen. If yes, restart the device and try again.
This case is already in queue.	Check if the case is in the upload queue in the workspace screen. If not, restart the device and try again.
Upload was canceled.	If you did not cancel the upload on purpose, restart the device and try again.
Please select a supplier	Select a supplier in the export screen and export the case again.
Failed to upload case Will automatically retry later.	Wait until it automatically retries. If it still fails, restart the device and try again.
Concurrent upload/multiple DWOS connect configurations not supported yet.	Contact your local distributor.
Internal server error (no token).	
A case cannot be exported twice.	Delete the case in DWOS Connect web interface and try again.
DWOS Connect configuration mismatch	Use DWOS Connect web interface to delete the case.

If the recommended action does not solve the problem, or for other on-screen messages, contact your local distributor.

8. Packing the scanner

Model 001 (cart)

Packing the scanner for transportation must be done as follows:

- Shut down the scanner unit.
- Flip the power switch in the *power off* position.
- Unplug the power cord, wind it carefully and place it inside one of the packaging inserts.
- Place the handpiece in its case.
- Pack the scanner in its original box and packaging material as shown below.




Figure 16: Steps for packing the scanner - Model 001 (cart)

Model 100 (laptop)




Figure 17: Steps for packing the scanner - Model 100 (laptop)

9. Distributors and service

Refer servicing to authorized service personnel. In case of questions or repair, please contact your local distributor.

Alternatively, contact support@dental-wings.com (in English, French or German).

10. Technical data and label

Г

Technical data					
Type/Model		DW-IO-001 (cart)	DW-IO-100 (laptop)		
Scanner unit input		100 - 240 V ~, 50 Hz - 60 Hz, 540 VA	100 - 240 V~, 50 - 60 Hz, 72 VA Handpiece Connector input: 5 V , 15 W		
Fuse		5 mm x 20 mm, medium speed, 6 A, ceramic (high breaking capacity), 250 V ~ (2 x M6AH250V)	no fuse		
	Frequency	2.4 GHz or 5 GHz			
	Standards	IEEE 802.11g/n/ac			
WiFi	Security	Protected Access (WPA2) Protected Setup (PIN & PBC)			
	Effective Radiated Power Output	0.085 W for IEEE 802.11 n/ac (dual band) 5 GHz 0.353 W for IEEE 802.11 g/n (single band) 2.4 GHz			
	Temperature limits	40°C - 48°C (104°F - 118°F)	40°C - 48°C (104°F - 118°F)		
	Weight	approx. 130 g (approx. 0.29 lbs)			
Handpiece	Dimensions (Ø at tip x L)	20 mm x 207 mm (0.8" x 8.1")			
Папарієсе	Туре (асс. IEC 60601-1)	Applied part type BF			
	Field of view	19.9 mm x 25.5 mm			
	Depth of field	15 mm			

Technical data		
Type/Model	DW-IO-001 (cart)	DW-IO-100 (laptop)
LED risk group (acc. IEC 62471)	Exempt ¹	
Laser class (acc. IEC 60825)	Class 1	no laser
Protection class (acc: IEC 60529)	IPXO (noncritical area - no pro	otection)
Protection class (electrical)	Class 1 Equipment; Basic insulation, relies on an earth for protection. The earth connection is achieved with a 3-conductor mains cable ending in a 3-prong AC connector which shall be plugged into a corresponding AC outlet.	Class 2 Equipment: Double insulated electrical equipment which does not require a safety connection to earth for protection.
Pollution Degree		ive pollution occurs. Temporary ensation is to be expected.
Mode of operation	Continuous	
Computer / operating system	Integrated high- performance computer storage for 8000 cases Dedicated high-end graphic card Windows 10 operating system Integrated gesture and voice recognition modules	Minimum requirements ² Memory: 16 GB DDR4 at 2400 MHz Processor: Intel® Core [™] i7- 8750H Graphic card: NVIDIA GeForce GTX 1060 with 6 GB GDDR5 Hard disk: 256 GB SSD Webcam
	21" multi-touch screen	17" screen

1 Handpiece with handpiece sleeve measured in worst case scenario (60 mm distance).

² Minimum requirements have been specified to allow for smooth operation of the software. The software might also run on configurations with higher performance. The suitability of such configurations has to be validated by the user.

Technical data		
Type/Model	DW-IO-001 (cart)	DW-IO-100 (laptop)
Dimensions uncrated (L x W x H)	520 mm x 460 mm x 1270 mm (20.5" x 18.1" x 50.0")	422 mm x 321 mm x 31 mm (16.6" x 12.6" x 1.2")
Dimensions crated (L x W x H)	600 mm x 510 mm x 1370 mm (23.6" x 20.1" x 53.9")	613 mm x 409 mm x 216 mm (24.1" x 16.1" x 8.5")
Weight uncrated	38.0 kg (84 lb)	5.9 kg (13.0 lbs)
Weight crated	54 kg (119 lb)	9.2 kg (20.2 lbs)

Technical data		
Type/Model	SharpAdjust	
Input	5V, 150 mA	
LED risk group (acc. IEC 62471)	Exempt	
Protection class (acc: IEC 60529)	IPXO (noncritical area - no protection)	
Protection class (electrical)	Class 3 equipment, designed to be supplied from a separated extra-low voltage (SELV) power source. The voltage from a SELV supply is low enough that under normal conditions a person can safely come into contact with it without risk of electrical shock.	
Pollution Degree	2; Normally only non-conductive pollution occurs. Temporary conductivity caused by condensation is to be expected.	
Mode of operation	Continuous	
Power source	Mandatory requirements: SharpAdjust is powered via USB cable. A separate USB power supply containing the safety markings required by national regulation can be used, as well as a USB port on a laptop or personal computer.	
Dimensions uncrated (L x W x H)	112 mm x 94 mm x 169 mm (4.4" x 3.7" x 6.7")	

Г

Technical data	
Type/Model	SharpAdjust
Dimensions crated (L x W x H)	241 mm x 190 mm x 160 mm
	(9.5" x 7.5" x 6.3")
Weight uncrated	800 g
Weight crated	1230 g

Ambient conditions	
Operating temperature	15°C - 30°C (59°F - 86°F)
Storage conditions	10 °C - 40°C (50°F - 104°F), controlled moisture (no condensation)
Transportation temperature	-40°C - 65°C (-40°F - 149°F)
Relative humidity	20% - 80%
Atmospheric pressure	60 kPa - 106 kPa

Software applications

Intraoral Scanner software application

Scanning parameters	
Patient characteristics	No restrictions apply
Optical technology	LED light High-speed measuring cameras Video cameras
Scanning accuracy/time	According to ISO 20896-1 and Dental Wings testing standard. For further details contact your local distributor <i>(see Distributors and service, p. 47)</i> .
Output format	Open STL or cloud sharing through DWOS Connect

Included

х

Optional

Scanner labels

This is an example of the label on your device:



+D840DWI00010/\$\$+7DWIOC-YY-ZZZZZZ/16DYYYYMMDD3

Figure 18: Device label – Model 001



+D840DWI01000/\$\$+7DWIOK-YY-ZZZZZZ/16DYYYYMMDDB

Figure 19: Device label - Model 100 (bright background)




Figure 20: Device label – Model 100 (dark background)

Replacement parts	Model 001	Model 100	Part No.	Image	Intended use
Handpiece	٠	•	6100-1 (Connection cable length 2.54 m)		Replacement or additional handpiece for the Intraoral Scanner.
Handpiece Storage Case	٠	٠	70-0045		Protects the handpiece and accessories during shipment and storage.
Handpiece Sleeve	٠	•	6190-1		Replacement or additional handpiece sleeve for the Intraoral Scanner.
Handpiece Sleeve Storage Case	٠	٠	70-0095		Protects handpiece sleeves during shipment and storage.
Power Cord	٠		22-0156 (US, CN: Length 2.5 m) 22-0159 (EU) 22-0165 (UK) 22-0167 (BR)		The country- specific power cord connects the Intraoral Scanner to power outlet.
Handpiece Connector Power Supply		•	21-0200 (US, EU, UK, AU)	(B)	The country- specific power supply connects the Intraoral Scanner to the power outlet.

Replacement parts	Model 001	Model 100	Part No.	Image	Intended use
Handpiece Connector Power Supply		٠	21-0228 & 22-0303 (US, ZH) 22-0301 (EU, BR) 22-0300 (UK)		The country- specific power supply connects the Intraoral Scanner to the power outlet.
Integrity-check Sleeve	•	•	6191-1	Ĩ	The integrity-check sleeve is a maintenance tool used to validate the integrity of the optics of the Intraoral Scanner.
SharpAdjust	٠	٠	6195-1		The SharpAdjust is a maintenance tool used to fine-tune the optical system of the Intraoral Scanner.
SharpAdjust Storage Case	٠	٠	6195-2		Protects SharpAdjust during shipment and storage.
Client Shipping Crate	٠		6092-1		This box serves as (replacement) box for return shipment.

Replacement parts	Model 001	Model 100	Part No.	Image	Intended use
Client Shipping Box		•	6392-1		This box serves as (replacement) box for return shipment.

11. Explanation of symbols





Caution: US Federal law restricts this device to sale by or on the order of a licensed health-care practitioner.



Contents of the transport package are fragile. Handle with care.



Indicates the correct upright position of the transport package



Do not dispose of this device as unsorted municipal waste. Collect separately.



Follow Instructions for Use



Device includes a radio frequency transmitter; interference may occur in the vicinity of equipment marked with this symbol.



Type BF Applied Part



Device weight is as indicated in the label, no working load permitted.



Temperature limit for transport



Humidity limitation



Atmospheric pressure limitation

Please note that the above symbols may appear with inverted colors (e.g. on the nameplate).

Appendix A - Reprocessing of the handpiece sleeve

Manufacturer: Dental Wings Inc. Method: Sterilization and intermediate or low level disinfection Device: Intraoral Scanner handpiece sleeve




Figure 21: Intraoral Scanner handpiece sleeve - semicritical surfaces

Precautions

- Always wear gloves.
- Do not use abrasive cleaners.
- Scratches on the optics cause damage to the handpiece sleeve. Be very careful in applying the soft, nylon-bristled brush and specific cloth.
- Protect personnel and surroundings from contamination while brushing.
- Carefully avoid inserting any sharp instrument in the gaps surrounding the optics as this may create concentrate forces on the glass susceptible to facture or delaminate the optics.
- Do not use corrosive cleaning agents or disinfectants with acids, bases, oxidizing agents, and solvents (with exception of alcohol).
- Avoid circular movement when wiping optics.
- Make sure the handpiece sleeve is completely dry before packing.
- Regularly inspect and clean the equipment used.
- When sterilizing multiple handpiece sleeves in one steam sterilizer cycle ensure that the sterilizer's maximum load is not exceeded.

Limitations on reprocessing

Repeated reprocessing has an effect on the handpiece sleeve. End of life is determined by cumulative effects of reprocessing, as well as damage and wear due to use. We strongly recommend that the user regularly inspects the handpiece sleeve. User are solely responsible for issues arising from the use of a damaged or worn handpiece sleeve.

Equipment and material

- Steam sterilizer (safety marking; qualified)
- Enzymatic detergent intended for cleaning of medical devices made of metal, glass and plastics
- Intermediate or low level disinfectant intended for surface disinfection of medical devices made of metal, glass and plastics; proven efficiency (e.g. FDA clearance, VAH/DGHM listing, CE marking)
- Alcohol (suitable for cleaning optics)
- Potable tap water (unless national regulation requires differently)
- Soft, lint-free, non-abrasive, absorbent, disposable cloth
- Soft, nylon-bristled brush
- Syringe
- Gloves (disposable, made of Nitrile or Latex)
- Standard packaging material suitable for steam sterilization (temperature resistance up to at least 137°C (278.6°F), sufficient steam permeability)

Instructions

Point of use

- Immediately remove excess soil within semicritical area using the cloth specified.
- Reprocess within two hours maximum.

Preparation for decontamination

- 1. Disassemble handpiece sleeve from handpiece (see Instructions for Use, Disassembling the handpiece sleeve (p. 20)).
- 2. Transport handpiece sleeve to designated reprocessing area (there are no requirements on containment and transportation).

Cleaning: Automated

Automated cleaning has not been validated since users do not have the equipment to clean automatically. Therefore, automated cleaning must not be applied to the handpiece sleeve.

Instructions

Cleaning: Manual

- 1. Select specific enzymatic detergent so that they comply with the equipment description of this procedure. Follow their instruction for use in terms of storage, concentration, application, temperature and post-rinsing.
- 2. Immerse handpiece sleeve in enzymatic detergent and soak as described in its instructions for use.
- 3. If residues are visible, thoroughly brush handpiece sleeve completely with the specified brush.
- 4. Use a syringe filled with detergent to flush all hard-to-reach areas of the handpiece sleeve.
- 5. Thoroughly rinse handpiece sleeve as described in the enzymatic detergent's instructions for use.
- 6. Dry using a fresh cloth. Use a one-directional wipe for the optics.

Disinfection: Manual

- 1. Select intermediate or low level disinfectant so that they comply with the equipment description of this procedure. Follow their instruction for use in terms of storage, concentration, application, temperature and post-rinsing.
- 2. Immediately after disinfection, clean the optics with a one-directional wipe and alcohol using a fresh, moist cloth (*see Instructions for Use,Reprocessing (p. 34)*).

Drying

Rest handpiece sleeve (optics facing down) on a fresh cloth in a clean place at room temperature until it is completely dry.

Inspection and Testing

- Visually inspect the handpiece sleeve thoroughly inside and outside for defects, damage and wear (e.g. for scratches, discoloration, corrosion, bubbles under the coating, coating deliminiation).
- Check that there is neither clouding nor scratches or chip outs on the optics.
- Replace handpiece sleeve in case of damage and wear.

Maintenance

Discard damaged instruments.

Instructions

Packaging

A standard sterilization packaging material shall be used. Ensure that the pack is large enough to contain the handpiece sleeve without stressing the seals.

Sterilization

- 1. Select steam sterilizer so that they comply with the equipment description of this procedure. Follow their instruction for use in terms of validation, loading weight, operating time and functional testing.
- 2. Select fractionated vacuum method (with sufficient device drying), with a minimum of 3 minutes at 132°C (270°F).
- 3. Do not exceed 134°C (273°F).
- 4. Drying time at least 30 minutes.

Storage

Store the handpiece sleeve in its sterilization packing in a dry and dust-free location.

Additional Information

The Instructions for Use contain more detailed descriptions and figures on assembly/disassembly to handpiece, transportation, handling (including storage), disposal, technical data, ambient temperature etc.

Manufacturer Contact

See the Instructions for Use back cover.

The following material has been used to validate the reprocessing described above:

• Enzymatic detergent: CIDEZYME®/Enzol® (ASP Johnson & Johnson), neodisher® MultiZym (Dr. Weigert)

• Disinfectant (alcohol-based): Bacillol® 30 and Bacillol® 30 Sensitive (both Bode Chemie), Isopropanol 70% (Jedmon Products Ltd.), ProSpray[™] and ProSpray C-60[™] (both Certol® International)

- Alcohol: Isopropanol 70% (Jedmon Products Ltd.)
- Cloth: Clean Wipes (Sticklers), WFw FiberWipes™ (AFL)

Fundamental suitability of the handpiece sleeve (initial state and after repeated reprocessing) and the instructions provided above for effective reprocessing have been validated by an independent government accredited and recognized test laboratory using the equipment and material specified in the test report, CIDEZYME®/Enzol® (both ASP Johnson & Johnson) and ProSpray™ (Certol® International).

Non-detraction of functionality was demonstrated in repeated processing using the equipment and material specified in the test report, CIDEZYME®/Enzol® (both Johnson & Johnson) and ProSpray[™] (Certol® International). In addition, it has been validated by Dental Wings using the aforementioned equipment and material, and neodisher® MultiZym (Dr. Weigert), Bacillol® 30 and Bacillol® 30 Sensitive (Bode Chemie), Isopropanol 70% (Jedmon Products Ltd.), ProSpray C-60[™] (Certol® International).

The instructions given have been validated by Dental Wings as being capable of preparing the Intraoral Scanner handpiece sleeve for re-use (250 cycles). It remains the responsibility of the user to ensure that the reprocessing as actually performed using equipment, material and personnel in the dental or reprocessing facility achieves the desired result. This requires validation and routine monitoring of the process. Likewise, any deviation by the user from the instructions provided is performed under sole responsibility of the user and should be properly evaluated for effectiveness and potential adverse consequences.

Date issued: July 2022

Appendix B - Reprocessing of the Intraoral Scanner (without

handpiece sleeve)

Manufacturer: Dental Wings Inc.

Method: Intermediate or low-level disinfection

Device: Intraoral Scanner (without handpiece sleeve), as per *Instructions for Use, Setting up the scanner (p. 17)*




Figure 22: Intraoral Scanner (without handpiece sleeve) - non-critical surfaces

Precautions

- Always wear gloves.
- Make sure the handpiece sleeve is disassembled from the handpiece and reprocessed separately as specified (*Appendix A Reprocessing of the handpiece sleeve (p. 58)*)
- Do not autoclave.
- Never immerse the handpiece in liquids.
- Do not use abrasive cleaners.
- Scratches on the optics cause damage to the handpiece. Be very careful when applying the specified cloth.
- Do not use corrosive cleaning agents or disinfectants with acids, bases, oxidizing agents, and solvents (with the exception
 of alcohol).
- Avoid circular movement when wiping optics
- Do not rinse the handpiece or the scanner unit.
- Make sure the handpiece is completely dry before assembled to the handpiece sleeve.
- Make sure the scanner unit is completely dry before being plugged into a power outlet.

Limitations on reprocessing

Repeated reprocessing has minimal effect on the Intraoral Scanner (without handpiece sleeve). End of life is normally determined by damage and wear due to use. We strongly recommend that the user regularly inspects the Intraoral Scanner (without handpiece sleeve). Users are solely responsible for issues arising from the use of damaged or worn Intraoral Scanner (without handpiece sleeve).

Equipment and material

- Intermediate or low level disinfectant intended for surface disinfection of medical devices made of metal, glass and plastics; proven efficiency (e.g. FDA clearance, VAH/DGHM listing, CE marking)
- Alcohol (suitable for cleaning optics)
- Potable tap water (unless national regulation requires differently)
- Soft, lint-free, non-abrasive, absorbent, disposable cloth
- Gloves (disposable, made of Nitrile or Latex)

Instructions

Point of use

- Immediately remove excess soil using the cloth specified.
- Reprocess within two hours maximum.

Preparation for decontamination

- 1. Unplug scanner (see Instructions for Use, Unplugging the scanner unit (p. 21)).
- 2. Disconnect the handpiece from the scanner (see Instructions for Use, Disconnecting the handpiece (p. 19)).
- 3. Select specific intermediate or low-level disinfectant according to its instructions for use. Follow it in terms of storage, concentration, application and temperature.

Cleaning: Automated

Automated cleaning must not be applied for the Intraoral Scanner (without handpiece sleeve).

Cleaning: Manual

Clean all surfaces with the cloth specified, moistened with potable tap water. Use a one-directional wipe for the optics.

Disinfection: Automated

Automated disinfection must not be applied for Intraoral Scanner (without handpiece sleeve).

Disinfection:

- 1. Apply intermediate or low-level disinfectant as described in the disinfectant's instructions for use. Use the cloth specified for wiping.
- 2. Immediately after disinfection, clean the optics with a one-directional wipe and alcohol using a fresh, moist cloth (*see Instructions for Use, Reprocessing (p. 34)*).

Drying

Dry at room temperature until completely dry.

Instructions

Inspection and Testing

- Visually inspect for damage and wear (e.g. for scratches, discoloration, corrosion).
- Check that there is neither clouding nor scratches or chip-outs on the optics or the optics of the handpiece.
- In case of doubt, conduct an integrity check (*see Instructions for Use, Maintenance tools (p. 37*).
- Replace handpiece in case of damage and wear (e.g. deep scratches, corrosion, failed maintenance as per *Instructions* for Use, Maintenance tools (p. 37).

Maintenance

Maintenance as per Instructions for Use, Maintenance (p. 32)

Sterilization

Sterilization must not be applied for the Intraoral Scanner (without handpiece sleeve). Sterilization is not required (all surfaces are classified noncritical).

Packaging/Storage

No requirements in particular. If the handpiece is stored on the scanner unit, cover it with an open see-through sterilization pouch. A size of $9 \text{ cm} \times 20 \text{ cm}$ is required. A standard plastic bag may be used to store the complete handpiece when detached from the scanner. A minimum size of $25 \text{ cm} \times 40 \text{ cm}$ is required.

Additional Information

Consult the Instructions for Use for more details.

Manufacturer Contact

See the Instructions for Use back cover.

The following material has been used to validate the reprocessing described above:

- Disinfectant (alcohol-based): Bacillol® 30 and Bacillol® 30 Sensitive (both Bode Chemie), Isopropanol 70% (Jedmon Products Ltd.), ProSpray C-60[™] (Certol® International)
- Alcohol: Isopropanol 70% (Jedmon Products Ltd.)
- Cloth: Clean Wipes (Sticklers), WFw FiberWipes™ (AFL)

Fundamental suitability of the Intraoral Scanner (without handpiece sleeve) (initial state and after repeated reprocessing) and the instructions provided above for effective reprocessing has been validated by Dental Wings. Non-detraction of functionality was demonstrated in repeated reprocessing using the equipment and material specified in the test reports, including Bacillol® 30 and Bacillol® 30 Sensitive (both Bode Chemie), Isopropanol 70% (Jedmon Products Ltd.), ProSpray C-60™ (Certol® International).

The instructions given have been validated by Dental Wings as being capable of preparing the Intraoral Scanner (without handpiece sleeve) for re-use (250 cycles). It remains the responsibility of the user to ensure that the reprocessing as actually performed using equipment, material and personnel in the dental or reprocessing facility achieves the desired result. This requires validation and routine monitoring of the process. Likewise, any deviation by the user from the instructions provided is performed under sole responsibility of the user and should be properly evaluated for effectiveness and potential adverse consequences.

Date issued: July 2022

Appendix C - Electromagnetic Compatibility Guideline

Guidance and manufacturer's declaration - electromagnetic emissions

The Intraoral Scanner is intended for use in the electromagnetic environment specified below. The user of the Intraoral Scanner should assure that it is used in such an environment.

Emission test	Compliance	Electromagnetic environment – guidance	
RF emissions CISPR 11	Group 1	The Intraoral Scanner uses radio frequency (RF) energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.	
RF emissions CISPR 11	Class B	The Intraoral Scanner is suitable for use in all establishments,	
Harmonic emissions IEC 61000-3-2	Class A	including domestic establishments and those directly connected	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	to the public low voltage power supply network that supplies buildings used for domestic purposes.	

Table 1

Guidance and manufacturer's declaration - electromagnetic immunity

The Intraoral Scanner is intended for use in the electromagnetic environment specified below. The user of the Intraoral Scanner should assure that it is used in such an environment.

Immunity Test	IEC 60601 test level	Compliance level	Electromagnetic environment – guidance
Electrostatic	±8 kV contact	±8 kV contact	Floors should be wood, concrete or ceramic
discharge (ESD) IEC 61000-4-2	± 2/4/8/15 kV air	± 2/4/8/15 kV air	tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	±2 kV	±2 kV	Maine neuror quelity chould be that of a typical
	100 kHz repetition frequency	100 kHz repetition frequency	Mains power quality should be that of a typic commercial or hospital environment.
Surge IEC 61000-4-5	± 0.5/1 kV line to line ± 0.5/1/2 kV line to ground	± 0.5/1 kV line to line ± 0.5/1/2 kV line to ground	Mains power quality should be that of a typical commercial or hospital environment.

Voltage dips and interruptions IEC 61000-4-11	0% U _T ; 0.5 cycle at 0°, 45°; 90°, 135°, 180°, 225°, 270°, and 315° 0% U _T ; 1 cycle and 70% U _T ; 25/30 cycles Single phase: at 0° 0% U _T ; 250/300 cycles	0% U _T ; 0.5 cycle at 0°, 45°; 90°, 135°, 180°, 225°, 270°, and 315° 0% U _T ; 1 cycle and 70% U _T ; 25/30 cycles Single phase: at 0° 0% U _T ; 250/300 cycles	Mains power quality should be that of a typical commercial or hospital environment. If the user of the Intraoral Scanner requires continued operation during power shortages, it is recommended that the Intraoral Scanner be powered from an uninterruptible power supply or a battery.
Rated power frequency magnetic field IEC 61000-4-8	30 A/m 60 Hz	30 A/m 60 Hz	Power frequency magnetic fields should be a levels characteristic of a typical location in a typical commercial or hospital environment.

G	uidance and manufacturer's dec	laration – electromagnetic im	munity			
	nded for use in the electromagne it is used in such an environment		v. The user of the Intraoral			
Immunity Test	IEC 60601 test level	Compliance level	Electromagnetic environment – guidance			
Portable and mobile RF communications equipment should be used no closer to any part of the Intraoral Scanner, including cables, than the recommended distance calculated from the equation applicable to the frequency of the transmitter.						
			Equation for distance			
			Equation for distance calculation			
	3 V	3 V	-			
Conducted disturbances	3 V 0.15 MHz - 80 MHz	3 V 0.15 MHz - 80 MHz				
Conducted disturbances inducted by RF fields IEC 61000-4-6						

Radiated RF EM fields IEC 61000-4-3	3 V/m 80 MHz - 2.7 GHz 80% AM at 1 kHz	3 V/m 80 MHz - 2.7 GHz 80% AM at 1 kHz	d = 1.2 √ P 80 MHz to 800 MHz d = 2.3 √ P 800 MHz to 2.7 GHz
Immunity to proximity fields from RF wireless communications equipment IEC 61000-4-3 (IEC61000-4- 39)	IEC 60601-1-2:2014; Table 9	IEC 60601-1-2:2014; Table 9	d = 6/E √ P 380 MHz to 5.8 GHz

P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and E is the Immunity Test Level (V/m), and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,^a should be less than the compliance level in each frequency range.^b Interference may occur in the vicinity of equipment marked with the following symbol:



NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Intraoral Scanner is used exceeds the applicable RF compliance level above, the Intraoral Scanner should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Intraoral Scanner.

b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 10 V/m.

IEC 60601-1-2:2014, Table 9					
Band		Service	P May power	d Distance	E Immunity test level
(MHz)			Max. power (W)	(m)	(V/m)
380-390	TETRA 400		1.8	0.3	27

430-470	GMRS 460; FRS 460	2	0.3	28
704-787	LTE Band 13, 17	0.2	0.3	9
800-960	GSM 800/900; TETRA 800; iDEN 820; CDMA 850; LTE Band 5	2	0.3	28
2400- 2570	Bluetooth; WLAN; 802.11 b/g/n; RFID 2450; LTE Band 7	2	0.3	28
5100- 5800	WLAN 802.11 a/n	0.2	0.3	9

Recommended separation distances between portable and mobile RF communications equipment and the Intraoral Scanner

The Intraoral Scanner is intended to be use in an electromagnetic environment in which radiated RF disturbances are controlled. The user of the Intraoral Scanner can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Intraoral Scanner as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output	Separation distance according to frequency of transmitter (M)			
power of transmitter (W)	150 kHz to 800 MHz d = 1.2 √ P	800 MHz to 2.7 GHz d = 2.3 √P	380 MHz to 5.8 GHz d = 6/E √P	
0.01	0.12	0.23	N.A.	
0.085 ¹	N.A.	N.A.	0.19	
0.1	0.38	0.73	N.A.	
0.352 ¹	N.A.	N.A.	0.13	
1	1.2	2.3	N.A.	
10	3.8	7.3	N.A.	
100	12	23	N.A.	

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

1) Maximum output power of the RF transmitter included in the Intraoral Scanner used in IEEE 802.11ac and 802.11n dual band (5 GHz).

2) Maximum output power of the RF transmitter included in the Intraoral Scanner used in IEEE 802.11g and 802.11n single band (2.4 GHz).

www.dentalwings.com



Dental Wings Inc.

160 Rue St-Viateur E Montreal QC H2T 1A8 CANADA

T +1 514 807-8485 (International) T +1 888 856-6997 (North America) F +1 514 223-3981

Intraoral Scanner Instructions for Use_EN 29-0100 v.4.3

CE

Distributor stamp