



Certificate of Calibration

Gas Flowmeter

Issued by V:KIT Ltd

Certificate No: 2777-GFM

Issue Date: 02-Jul-2020

V:Kit Ltd

16 John Bradshaw Court
Alexandria Way
Congleton,
Cheshire, CW12 1LB, UK
Tel: +44 (0) 1260 274266
www.v-kit.com - info@v-kit.com

Customer: Kirantoni SRL

Address: 132, Miron Costin 18 street

Chisinau

Republic of Moldova

Customer PO No.: ref: 9780

Date of Receipt: 17-Jun-2020

Calibration performed by: C. Anonuevo

Calibration Date: 30-Jun-2020

Expiry Date: 29-Jun-2022

Manufacturer: V:KIT Ltd

Model: VKIT-GFM3

Equipment Condition: Clean

Serial Number: 950178

Calibration Range: 0 - 500 ml/min

Test Media: Nitrogen

Laboratory Temperature: 19.9°C

Laboratory Barometric Pressure: 1013 mBar

Reference Equipment: Traceable to National Standards

Calibration Method

The electronic Flowmeter was tested at the required flow rates and compared with NIST Traceable primary standards.

Test Equipment Used

Barometer PIN:51107

Thermometer: PIN:21317

Flow Reference 1: PIN:86225

Flow Reference 2: PIN:86226

Calibration Results

Actual Flow (ml/min)	Indicated Flow (ml/min)
0.0	0.0
10.0	9.9
50.0	50.2
75.0	75.5
100.0	100.2
200.0	200.0
300.0	300.1
400.0	401.0

Approved By: Tom Gowans

Approval Date: 02-Jul-2020