



Bio-Rad
Certificate of Analysis

Material Description:	Ampicillin 10µG 4x50 disks
Material Number:	66128
Batch Number:	64635173
Manufacture Date:	2024-10-14
Expiration Date:	2026-10-14

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
AMP 10	
Ink colour	Conforms
Green	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
15-22	
Staphylococcus aureus ATCC_25923	Conforms
27-35	

Refer to current CLSI guidelines

Sessinou Sonon

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-10-25 / 09:06:25 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Gentamicin 10µG 4x50 disks
Material Number:	66608
Batch Number:	64635195
Manufacture Date:	2024-11-13
Expiration Date:	2026-11-13

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
GMN 10	
Ink colour	Conforms
Black	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
19-26	
Staphylococcus aureus ATCC_25923	Conforms
19-27	
Pseudomonas aeruginosa_ATCC 27853	Conforms
17-23	

Refer to current CLSI guidelines

Tarik Hamdi

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-22 / 12:21:49 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Azithromycin 15µG 4x50 disks
Material Number:	67008
Batch Number:	64606245
Manufacture Date:	2024-04-12
Expiration Date:	2025-10-09

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	AZM 15
Ink colour	Conforms
	Pink
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Staphylococcus aureus ATCC_25923	Conforms
	21-26

Refer to current CLSI guidelines

Mélanie Fanchon

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-05-21 / 08:03:41 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Meropenem 10µG 4x50 disks
Material Number:	67048
Batch Number:	64635535
Manufacture Date:	2024-11-13
Expiration Date:	2026-11-13

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	MEM 10
Ink colour	Conforms
	Green
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
	28-35
Staphylococcus aureus ATCC_25923	Conforms
	29-37
Pseudomonas aeruginosa_ATCC 27853	Conforms
	27-33

Refer to current CLSI guidelines

Tarik Hamdi

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-22 / 13:29:26 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Colistin 10µG 4x50 disks
Material Number:	67268
Batch Number:	64635178
Manufacture Date:	2024-10-31
Expiration Date:	2026-10-31

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
COL 10	
Ink colour	Conforms
Brown	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
11-17	
Pseudomonas aeruginosa_ATCC 27853	Conforms
11-17	

Refer to current CLSI guidelines

Tarik Hamdi

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-12 / 12:39:02 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Ceftazidime 10µG 4x50 disks
Material Number:	67298
Batch Number:	64635542
Manufacture Date:	2024-10-21
Expiration Date:	2026-10-21

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	CZD 10
Ink colour	Conforms
	Green
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
	23-29
Pseudomonas aeruginosa_ATCC 27853	Conforms
	21-27

Refer to current EUCAST guidelines

Sophie Soignet

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-05 / 09:55:20 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Tigecycline 15µG 4x50 disks
Material Number:	67398
Batch Number:	64614401
Manufacture Date:	2024-07-11
Expiration Date:	2026-07-11

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
TGC 15	
Ink colour	Conforms
Orange	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
20-27	
Staphylococcus aureus ATCC_25923	Conforms
20-25	
Streptococcus pneumoniae ATCC_49619	Conforms
23-29	

Refer to current CLSI guidelines

Sessinou Sonon

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-08-13 / 09:21:38 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Tetracycline 30µG 4x50 disks
Material Number:	67448
Batch Number:	64635185
Manufacture Date:	2024-11-08
Expiration Date:	2026-11-08

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
TET 30	
Ink colour	Conforms
Orange	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
18-25	
Staphylococcus aureus ATCC_25923	Conforms
24-30	

Refer to current CLSI guidelines

Tarik Hamdi

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-22 / 13:59:58 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Cefotaxime 5µG 4x50 disks
Material Number:	67718
Batch Number:	64632830
Manufacture Date:	2024-09-23
Expiration Date:	2026-09-23

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	COX 5
Ink colour	Conforms
	Green
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
	25-31
Haemophilus influenzae ATCC_49766	Conforms
	29-37

Refer to current EUCAST guidelines

Stéphanie Cambier

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-10-07 / 09:58:47 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Nalidixic Acid 30µG 4x50 disks
Material Number:	68618
Batch Number:	64593618
Manufacture Date:	2024-02-23
Expiration Date:	2026-02-22

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	NAL 30
Ink colour	Conforms
	Brown
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
	22-28

Refer to current CLSI guidelines

Hélène Bruxelles

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-03-13 / 08:20:01 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Ciprofloxacin 5µG 4x50 disks
Material Number:	68648
Batch Number:	64635186
Manufacture Date:	2024-11-07
Expiration Date:	2026-11-07

Characteristic	Results
Labeling	Conforms
Specifications see above	
Acronym	Conforms
CIP 5	
Ink colour	Conforms
Brown	
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
29-37	
Staphylococcus aureus ATCC_29213	Conforms
21-27	
Pseudomonas aeruginosa_ATCC 27853	Conforms
25-33	

Refer to current EUCAST guidelines

Sophie Soignet

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-11-18 / 09:24:18 UTC

Date/Time



Bio-Rad
Certificate of Analysis

Material Description:	Trimethoprim 5µG 4x50 disks
Material Number:	68888
Batch Number:	64622516
Manufacture Date:	2024-08-29
Expiration Date:	2026-08-29

Characteristic	Results
Labeling	Conforms
	Specifications see above
Acronym	Conforms
	TMP 5
Ink colour	Conforms
	Brown
.....	
Zone Diameter Interpretive Criteria (mm)	
.....	
Escherichia coli ATCC_25922	Conforms
	21-28
Staphylococcus aureus ATCC_25923	Conforms
	19-26

Refer to current CLSI guidelines

Sophie Soignet

This certificate has been verified and electronically approved by an authorized Quality representative.
This e-signature is performed within a secure system.

2024-09-11 / 09:07:31 UTC

Date/Time