

Specificații tehnice (F2.1)

[Acest tabel va fi completat de către ofertant în coloanele 2, 3, 4, 6, iar de către compania contractantă – în coloanele 1, 5,]

Se indică denumirea bunului(lor)/serviciului(lor) obiect(e) al procedurii de achiziții

Denumirea bunurilor	Modelul articolului	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant
1	2	3	4	5	6
Hioki	RM3545-01	China	HIOKI	Conform specificațiilor caietului de sarcini	Echipament pentru măsurarea rezistenței cu: Resistance measurement: 0.000 00mΩ (10mΩ range) to 1200.0MΩ (1000MΩ range), 12 ranges Low power resistance measurement: 0.00mΩ (1000mΩ range) to 1200.00Ω (1000Ω range), 4 ranges Temperature measurement (thermistor): - 10.0 to 99.9°C Temperature measurement (analog input): - 99.9 to 999.9°C Temperature correction, Temperature conversion, Self-calibration, offset voltage compensation (OVC), comparator (ABS/REF%), BIN, key-lock (OFF, menu lock, all lock), display digit count selection function (7 digits/6 digits/5 digits), automatic power supply frequency settings (AUTO/50Hz/60Hz), scaling, judgment sound setting, auto hold, statistical calculations, clock, self-test, L2105 LED Comparater Attachment output GP-IB interface

Director Lokmera SRL

Nicolai Iasîbaș