



Technical Specification No 17

9 LITERS SAFE VACCINE TRASPORTATION BOX

- In Manufacturer: TAMKAN
- Country of Origin : TURKIE
- Model: **TK1277**

Technical Specification	Yes	No	Remarks
Configuration			
Technical specifications			
9 Liter Safety Vaccine			
Double-sided reinforced and double-lockable carrying handle	X		☐ leaflets , user manual
External dimensions: 40 x 30 x 27 cm.	X		☐ leaflets , user manual
Internal dimensions are 24 x 22 x 18 cm.			
Being able to monitor the indoor temperature without opening the lid, thanks to the digital degree mounted on the front body and extending the probe into the container.	X		
	X		
	X		
	X		
The digital degree battery can be easily replaced without removing the device. The device must have an on-off button.			
Gross Volume 12 liters (net: 9 liters)			
There should be at least 3 pockets where the ice packs will be placed, and direct			

<p>contact of the ice packs and vaccines should be prevented. Ice packs should lower the bag temperature quickly.</p> <p>9 Liter SAFETY Vaccine Transport Bag - TK1277 7 ice packs should be able to be placed in the slots on the lid and the bottom of the container. The internal organization of the container is such that the ice packs do not come into contact with the vaccine.</p> <p>Thanks to the polyurethane insulation and ice batteries, the inside of the container can be kept stable at 2 - 8 degrees for up to 29 hours. All surfaces of the vaccine transport bag are capable of maintaining the temperature.</p> <p>7 free ice packs should be provided with each bag. The inside temperature of the bag drops below 8 degrees in 10 minutes. Keeping below 8 degrees for 29 hours.</p> <p>Accessory:</p> <ul style="list-style-type: none"> • lithium battery - 2 pcs; <p>Battery working time: min 6 hours;</p> <ul style="list-style-type: none"> • Car adopter - 1 piece; • Charger - 1 pc; • Adjustable shoulder strap - 1 piece; • Cover for accessories - 1 piece; 			
Certification			
<input type="checkbox"/> ISO 9001	X		<input type="checkbox"/> certificate attached
<input type="checkbox"/> ISO 13485	X		<input type="checkbox"/> certificate attached
<input type="checkbox"/> CE/CE Declaration of Conformity 93/42/EEC	X		<input type="checkbox"/> certificate attached

<input type="checkbox"/> ISO 10002	X		<input type="checkbox"/> certificate attached
Technical Specification	Yes	No	Remarks
Other services included in offer			
<input type="checkbox"/> transport to the final customer	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> instalation and put in function	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> training for the technical staff for final customer	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> training for the medical staff for final customer	X		<input type="checkbox"/> bidder declaration attached
warranty for 24 months			
GUARANTEE			
<input type="checkbox"/> 24 months guaranteed from signing the receiving minutes	X		<input type="checkbox"/> bidder declaration attached
SERVICE AND MAINTENANCE			
<input type="checkbox"/> the bidder has technical facilities necessary for the service both for ambulances as well as for medical equipment according to the general guarantee conditions and the user guide of the manufacturer	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Maximum time for technical intervention – 48 hours.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Maximum time for remedial measures, in total 72 hours.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The technical service and current repair will be performed without waiting in a queue.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Temporary replacement of the equipment has to be ensured according to the periods mentioned above.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> During the guarantee period, at the reasonable request of the user, repair, adjustments and maintenance of medical equipment ,according to the specifications in the manufacturers' guides, will be done free of charge.			
<input type="checkbox"/> The spare parts and workmanship are for free, except from the consumables for vehicles established by the manufacturer	X		<input type="checkbox"/> bidder declaration attached

Technical Specification	Yes	No	Remarks
AVAILABILITY OF SPARE PARTS			
<input type="checkbox"/> the bidder will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows: <ul style="list-style-type: none"> ○ spare parts free of charge, including the workmanship for the 36 months of guarantee period ○ for the rest of the period – against payment for at least 10 years period is fulfilled. 	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary	X		<input type="checkbox"/> bidder declaration attached
GUIDES			
<input type="checkbox"/> Technical Guide and User Guide.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> All guides will be available in Romanian /Russian.	X		<input type="checkbox"/> bidder declaration attached