

• **Body Thickness Correction Key**⑩

The radiography conditions can be corrected for the patient's body thickness.

- 1** Press the Body Thickness Correction key.

The body thickness correction window is displayed.



- 2** Select the body thickness correction conditions.

The Body Thickness Correction key displays the selected conditions.

Reference [4.3 "Editing Radiography Programs"](#)