

EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.

AB-0583-T

21003760-  
İNG

02-21

REQUIRED TESTS	RESULT	COMMENTS
<b>PHYSICAL PROPERTIES</b>		
Tensile Strength / Dry	P	
Bursting Strength / Dry	P	
Water Permeability	P	
Lint and Other Particles Generation From Nonwoven	P	
<b>MICROBIOLOGICAL TESTS</b>		
Dry-Bacterial Penetration	P	
Microbial Cleanliness (Bioburden)	P	
P: Pass F: Fail R: Refer to retailer technologist. Test results were evaluated according to EN 13795-1:2019 Standard Performance Properties Less Critical limit values		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95 %. The declaration of conformity was given in accordance with the Simple Acceptance Decision Rule. Tests marked (\*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.  
Testing reports without signature and seal are not valid.

ZENMED TIBBİ CİHAZLAR  
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.  
Mahmutbey Mah. 2567.Sk. No:1 İş Yeri No:11,21  
Bağcılar / İSTANBUL / TÜRKİYE  
Tel: 0212 445 45 25 Tic.Sic.No: 294704  
Güneşli V.D.: 997 067 1542