

Certificate of Compliance



No. 0A150331.AMEUU84

Certificate's Holder:

ARI Technology (Group) Co., Ltd.
 Building 15, The Fourth Industrial Zone, Gongming Town, Guangming District, Shenzhen, China.

Certification ECM Mark:



Product:

Syringe Pump, Infusion Pump, Feeding Pump, Insulin Pump

Model(s):

ASP-1800, ASP-1800C, ASP-1800G, ASP-1800D, AIP-1200, AIP-1200Y, AIP-Y1200, AFP-500, AIP-300

Verification to:

Standard:
 EN 60601-1:2006/AC:2010,
 EN 60601-1-2:2007/AC:2010, EN 60601-1-4:1996,
 EN 60601-1-8:2007, EN ISO 8835-2:2009,
 EN ISO 8835-3:2009/A1:2010, EN ISO 8835-4:2009,
 EN ISO 8835-5:2009, EN ISO 5356-1:2004,
 EN ISO 10993-1:2009, EN ISO 14971:2012,
 EN ISO 14155:2011, EN ISO 15223:2012, EN 1041:2008

related to CE Directive(s):
 93/42/EEC (Medical Devices)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Certification Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. Regulation can be found at www.entecerma.it.

Whereas the Manufacturer is responsible of the CE certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible to maintain efficient the internal production control to ensure the product(s) are in compliance with the Certification ECM Mark.

Date of issue 31 March 2015

Expiry date 30 March 2020

Chief Manager

Tim Bianchi



Deputy Manager

Viola Miller



Ente Certificazione Macchine Srl

Via Ca' Bella, 243 - Loc. Castello di Serravalle - 40053 Valsamoggia (BO) - ITALY