

In attention to:

Public Institution

“Office for External Assistance Programs Management” OGP AE

Tender no: ocds-b3wdp1-MD-1673285147976

Manufacturer statement (Lot-1)

We, Sümer Uluslararası Sanayi ve Ticaret A.Ş., herewith confirm that Eyecon Medical SRL is authorized to participate in this tender and state the following:

1. The Medical Oxygen Generator, model Oxyfresh-200, as a part of the whole system Oxygen Generator Station, is certified by CE Certificate number 2195-MED-1230801 valid until 2024-05-26 according to 93/42 EEC Directive, classified as a Medical Device of class IIb according to Declaration of Conformity number TD.OD.ECD.01 from 06.04.2020
2. We confirm that all the components of the Oxygen Generator Station:
 - Compressor (SK25, Kaeser, Germany)
 - Dryer (IDFA-60, SMC, Japan)
 - Filters (MHA-260/G+G01+H+C+DS (Anti-Bacterial), Turkey)
 - Air Tank (AVT500-11 bar/Galvanized Steel, Turkey)
 - Oxygen Generator (Oxyfresh-200 + Filters (MHA-65/D+DS), Turkey)
 - Oxygen Tank (AVT500-11 bar/Stainless Steel 304 L, Turkey)
 - Container (20 Feet + All Accessories, Turkey)



have the mutual compatibility of devices (components) according to the instructions of manufacturers and that this reassembly was performed in accordance with these instructions.

The system was packaged and assembled together with all components and the users will be informed with relevant information including the relevant manufacturers' instructions.

The entire activity is subject of appropriate methods of internal control and inspection.

3. We offer the warranty that in the next 10 years we can assure the supply and production of spare parts of the Oxygen Generator System.

15.02.2023

Osman Çetinel, CEO

A blue ink signature of Osman Çetinel is written over the company name and address.
SÜMER
ULUSLARARASI SAN. ve TIC. A.Ş.
Başkent OSB Başkent Bulvarı No:81 Malıköy
Tel:394 61 71 Fax:394 61 95 Sincan/ANKARA
Sincan Y.D. 06100 8595

