

CERTIFIED REFERENCE MATERIAL

Ref No: SB46124.10MG

Lot No: SAMPLE

Certification Date: xxxxxx

Barcode: 41554388

Expiry date: xxxxxx

Description: 2-Hydroxymethyl-1-methyl-5-nitroimidazole

CAS No: 936-05-0

Empirical formula: C₅H₇N₃O₃

MW: 157.123

Certified Purity / Uncertainty: 0.9990 +/- 0.0010 g/g (99.90 +/- 0.10 %)

Purity=100% - Assay organic impurities

Storage Conditions: Store in a refrigerator at temperatures between 2°C to 8°C

Method of certification:	Concept of Certification and traceability statement:	Intended use:	Instructions for the correct use of this reference material:	Stability and storage:	Level of homogeneity:
CRM's calibration procedure (WQP 5.15.1/22).	The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02. Metrological traceability is established through in-house validated method. The measurement results are traceable to SI.	For Laboratory Use Only. This CRM is intended for: Calibration of TLC, GC/FID, GC/TCD, GC/ECD, GC/MS, GC/MS/MS, LC/UV, LC/MS and LC/MS/MS. Validation of analytical methods. Preparation of "working reference samples". Detection limit and linearity studies. This statement is not intended to restrict the use for other purposes.	This CRM can be used directly or can be diluted in an appropriate solvent. Only a clean glassware should be used. Hazardous situation: The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this chemical are available as safety data sheet.	This CRM is with a guaranteed purity +/- 2% deviation prior to the expiration date. Stability is guaranteed, provided that the material is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions.	The material was tested for homogeneity by analyzing randomly selected samples according to an in-house procedure. The level of homogeneity proved satisfactory for a sample volume of min. 2 mg. The uncertainty incorporates the sample standard deviation combined with the uncertainty calculated from homogeneity and stability studies.

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)

- Accredited according to ISO/IEC 17025

- Accredited according to ISO 17034

This document is designed and the certified value and uncertainty are determined in accordance with ISO 33401, ISO 33405, and Eurachem / CITAC Guides

Names of certifying officers:

Laboratory:



Margarita Dimitrova

Manager:



Krassimira Taralova

Analytical Data:

Column:	Chromolith® Performance RP-18e 4.6x100 mm, 2µm particle size	Isocratic elution (time)	A%	B%
Column temperature:	20°C	0	10	90
Detector:	DAD	7	10	90
Flow rate:	2.0 ml/min			
Injection volume:	10µl			
Mobile phase:	A: Water B: Acetonitrile			



CPAchem Ltd. is ISO 17034 (Cert No AR-1835)
and ISO/IEC 17025 (Cert No AT-1836) accredited by ANAB

CPAchem Ltd
www.cpachem.com
e-mail: info@cpachem.com; tel.: +359 42952901
for France, Belgium and Switzerland:
e-mail: acsd2@wanadoo.fr; tel.: 01 30 57 57 32 / fax: 01 30 57 57 33

Peak Integration Report

Sample Name:	2-Hydroxymethyl-1-methyl-5-nitroimidazole_41554388	Inj. Vol.:	10.00
Instrument Method:	Chromolith isocratic_10_90_2 ml flow		
Inj. Date / Time:	21-Mar-2024 / 10:34	Run Time:	7.00

No.	Time min	Peak Name	Peak Type	Area mAU*min	Height mAU	Rel.Area %
1	0.82	2-Hydroxymethyl-1-methyl-5	BMB*	49.444	1157.407	99.93
TOTAL:				49.44	1157.41	99.93

