

Certificate of Analysis

ThermoFisher
SCIENTIFIC


Anode Buffer Container 3500 Series


Product No. 4393927
Lot No. 2202755
Date of Manufacture 29MAR2022
Expiration Date 20NOV2022

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.924
CONDUCTIVITY TEST		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1760 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Luciana Torcasio (Mar 30, 2022 10:08 EDT)
Quality Assurance
Issued 30MAR2022