



## TECHNICAL FILE – DECLARATION OF CONFORMITY

DESCRIPTION	Intellect Mobile 2
-------------	--------------------

CLASSIFICATION	Ila
----------------	-----

Revision	Effective Date	Originator	Description
A	31 oct 2019	Dombovári, B.	Initial Release
B	18 nov 2019	Allard, T.	Add Ultrasound applicators codes Add ISO13485 to the standards table
C	See Agile	Pouy, S.	GMDN code update Addition of <ul style="list-style-type: none"> <li>- GMDN 61186 for COMBO device, which is a multi-modality physical therapy system</li> <li>- GMDN 11248 for Ultrasound device</li> <li>- GMDN 63505 for ultrasound system handpiece</li> </ul>
D	See Agile	Dombovári, B.	Remove UMDNS code 18823



<b>DECLARATION OF CONFORMITY</b>	
<b>MANUFACTURER</b>	<b>DJO FRANCE SAS</b> 3 rue de Bethar Centre Européen de Frêt 64990 Mouguerre FRANCE
<b>EU AUTHORIZED REPRESENTATIVE (MDD)</b>	N/A
<b>PRODUCT</b>	Intellect Mobile 2
<b>PART NUMBER LIST</b>	TF-FRA-014-3
<b>MDD CLASSIFICATION</b>	Class IIa
<b>RED CLASSIFICATION</b>	Class 1
<b>CONFORMITY ASSESSMENT ROUTE</b>	Annex VII (MDD) Annex II (RED)
<b>GMDN CODE</b>	46573, 63505, 61186, 11248
<b>UMDNS CODE</b>	17908
<p>WE, THE MANUFACTURER, <b>DJO FRANCE SAS</b>, DECLARE UNDER SOLE RESPONSIBILITY THAT THE ITEM TO WHICH THIS DECLARATION IS RELATED IS IN CONFORMITY WITH:</p> <ul style="list-style-type: none"> <li>ALL RELEVANT PROVISIONS OUTLINED IN THE OFFICIAL JOURNAL OF THE EUROPEAN COMMUNITY COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES. THE ITEM COMPLIES WITH ALL RELEVANT PROVISIONS OF THE ANNEX I ESSENTIAL REQUIREMENTS, AS AMENDED UP TO AND INCLUSIVE OF COUNCIL DIRECTIVE 2007/47/EC.</li> <li>DIRECTIVE 2011/65/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL OF 8 JUNE 2011 ON THE RESTRICTION OF THE USE OF CERTAIN HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT (ROHS-2)</li> <li>DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL OF 16 APRIL 2014 ON THE HARMONISATION OF THE LAWS OF THE MEMBER STATES RELATING TO THE MAKING AVAILABLE ON THE MARKET OF RADIO EQUIPMENT AND REPEALING DIRECTIVE 1999/5/EC</li> </ul>	
<b>STANDARDS APPLIED</b>	EN ISO 13485:2016/AC:2016ISO      Medical devices - Quality management systems - Requirements for regulatory purposes
	EN 60601-1:2006/A1:2013 IEC 60601-1:2005/A1:2012      Safety Requirements for Medical Electrical Systems. Ed. 3.1
	IEC 60601-1-2:2014 EN 60601-1-2:2015      Electromagnetic Compatibility – Requirements and Tests. Ed. 3
	IEC 60601-2-5:2009 EN 60601-2-5:2015      Medical electrical equipment – Particular requirements for the safety of ultrasonic physiotherapy equipment. Ed. 3
	IEC 60601-2-10:2012/AMD1:2016 EN 60601-2-10:2016/A1:2016      Medical electrical equipment – Particular requirements for the basic safety and essential performance of nerve and muscle stimulators. Ed. 2
	IEC 60601-1-11:2015 EN 60601-1-11:2015      Medical electrical equipment —General requirements for basic safety and essential performance — Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
	EN 60601-1-6:2010/A1:2013 IEC 60601-1-6:2010/AMD1:2013      Medical electrical equipment - General requirements for basic safety and essential performance - Collateral standard: Usability
	IEC 62366-1:2015 EN 62366-1:2016 IEC 62366 :2014      Medical devices – Application of Usability
	IEC 62133-1: 2017 (Nickel Systems) IEC 62133: 2012(Lithium Systems)      Secondary cells and batteries containing alkaline or other non-acid electrolytes - Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications
	IEC 62304:2006/AMD1:2015 EN 62304/A1:2015      Medical device software – Software life-cycle processes
EN ISO 14971:2012      Medical Devices – Application of Risk Management to Medical Devices	

	ISO 10993-1:2018	Biological Evaluation of medical devices – Part 1: General requirements for basic safety and essential performance
	ASTM D4169-16	Standard practice for performing testing of shipping containers and systems
	MEDDEV 2.7/1 rev.4	Clinical evaluation: Guide for manufacturers and notified bodies
	ETSI EN 301 489-1 V1.8.1 (2008-04)	Electromagnetic compatibility and Radio Spectrum Matters (ERM); Electromagnetic Compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements
	ETSI EN 301 489-3 V1.4.1 (2002-08)	Electromagnetic compatibility and Radio spectrum Matters (ERM); Electromagnetic Compatibility (EMC) standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD) operating on frequencies between 9 kHz and 246 GHz
	2014/53/EU	Radio Equipment Directive (RED)
	(EU) 207/2012	Electronic instructions for use of medical devices
<b>NOTIFIED BODY</b>	BSI Group The Netherlands B.V. Say Building, John M. Keynesplein 9, 1066 EP Amsterdam Netherlands 2797	
<b>EC CERTIFICATE(S)</b>	CE 681250	
<b>PLACE OF ISSUE</b>	64990 Mouguerre, France	
<b>SIGNATURE</b>	<p>SIGNED FOR AND ON BEHALF OF <b>DJO FRANCE SAS</b></p> <p><i>Dombovári</i></p> <p>Vu pour certification matérielle de la Signature de M. <i>Britta DOMBOVÁRI</i> BAYONNE, le <i>28/08/2020</i></p> <p>POUR LE PRÉSIDENT DE LA CHAMBRE DE COMMERCE ET D'INDUSTRIE DE BAYONNE PAYS BASQUE</p> <p><b>For le Président de la Chambre de Commerce et d'Industrie de Bayonne Pays Basque</b></p> <p><b>F. HARRIAGUE</b></p>	





APOSTILLE  
(Convention de La Haye du 5 octobre 1961)

- 1.- République française  
Le présent acte public
- 2.- a été signé par F. HARRIAGUE
- 3.- agissant en qualité Agent
- 4.- est revêtu du sceau de Chambre de Commerce  
et d'Industrie de Bayonne

- 5.- à Pau Attesté le 29/9/20
- 7.- Par le Procureur Général près la Cour d'Appel
- 8.- sous n° 2012062
- 9.- Sceau 10.- Signature



Le Procureur Général  
Michel BEAULIER

“L’apostille confirme seulement l’authenticité de la signature, du sceau ou du timbre sur le document. Elle ne signifie pas que le contenu du document est correct ou que la République Française approuve son contenu.”

