


# Valleylab™ FT10 Energy Platform Specification Guide



### Valleylab™ FT10 energy platform —

our next generation advanced energy platform with improved LigaSure™ vessel sealing technology and expanded electrosurgical features.

Be the first to see how over **2 million devices** sold annually will see performance enhancements.



spread to surrounding tissue<sup>3</sup>

- Intuitive use4

# Valleylab™ FT10 Energy Platform

## **Technical Specifications**

#### General

Output configuration	Isolated output			
Cooling	Natural convection and fan			
Display	7 in. LCD touchscreen			
Connector ports	LED illuminated Smart™ connector readerse on the LigaSure™/Bipolar receptacle			
Enclosure	Magnesium			
Mounting	• Valleylab™ Universal Generator Car (VLFTCRT)			
	Operating-room boom systems			
	• Any stable, flat surface such as a table or cart top			

### **Dimensions and Weight**

Height	6.7 in. (17.0 cm)		
Width	14.5 in. (35.8 cm)		
Length	18.2 in. (46.2 cm)		
Weight	22.3 lb. (10.1 kg)		

#### **Operating Parameters**

Ambient temperature range	50 to 104 °F (10 to 40 °C)
Relative humidity	30% to 75% non-condensing
Atmospheric pressure	700 to 1060 millibars

#### Transport and Storage

Ambient temperature range	14 to 149 °F (-10 to +65 °C)
Relative humidity	25% to 85% non-condensing
Atmospheric pressure	500 to 1060 millibars

#### **Duty Cycle**

The Valleylab™ FT10 energy platform is capable of operating a duty cycle of 25%, defined as 10 seconds active and 30 seconds inactive, in any mode for a period of 4 hours.

#### Internal Memory

Real-time clock	Battery type - Lithium CR1620;
battery	Battery capacity - 75 mAh
Storage capacity	4 GB

#### **Audio Tones**

Activation Tones	Tone	Duration	Volume
CUT	660 Hz ± 5%	Entire Activation Duration	User adjustable from 45 dBA to 65 dBA (-0/+6 dBA @ 1 m)
COAG	940 Hz ± 5%	Entire Activation Duration	User adjustable from 45 dBA to 65 dBA (-0/+6 dBA @ 1 m)
VALLEYLAB	800 Hz ± 5%	Entire Activation Duration	User adjustable from 45 dBA to 65 dBA (-0/+6 dBA @ 1 m)
BIPOLAR	940 Hz ± 5%	Entire Activation Duration	User adjustable from 45 dBA to 65 dBA (-0/+6 dBA @ 1 m)
LIGASURE	440 Hz ± 5%	Entire Activation Duration	User adjustable from 45 dBA to 65 dBA (-0/+6 dBA @ 1 m)

#### Radio Frequency Identification (RFID)

Frequency Range	13.56 MHz		
RF Output Power	68.17 dBuV/m @ 3 meters		
Type of Antenna	Integral Loop Antenna		
Modulation	Amplitude-shift Keying (ASK)		
Mode of Operation (Simplex/Duplex)	Duplex		
Contains Transmitter Module FCC ID	2AAVI-JDK1901		
Contains IC ID	11355A-JDK1901		

#### Wireless Fidelity (WiFi)

Transmit/Receive Frequency Range	2.4000 ~ 2.4835 GHz (Industrial Scientific Medical Band)			
Standards	IEEE 802.11b, 802.11g, 802.11n			
RF Output Power	11b: 17 ± 1.5 dBm			
	11g: 15 ± 1.5 dBm			
	11n: 14 ± 1.5 dBm			
Data Rate	11b: 1/2/5.5/11 Mbps			
	11g: 6/9/12/24/36/48/54 Mbps			
	11n: (20 MHz): MCSO-7 (Up to 72 Mbps)			
	11n: (40 MHz): MCSO-7 (Up to 150 Mbps)			
Securities	WEP 64/128, WPA, WPA2, and IEEE 802.1x			
Type of Antenna	Internal Antenna (1T1R)			
Contains Module FCC ID	NDD9578111008			
Contains IC ID	4701A-78111306			

#### Leakage

## Leakage Currents and Patient Auxiliary Currents (IEC 60601-1:2012)

Touch Current	< 100 μΑ ΝC, < 500 μΑ SFC			
Earth Leakage Current	< 500 μA NC, < 1000 μA SFC			
Patient Auxiliary Current (< 1kHz)	< 10 μA NC, < 50 μA SFC			
Patient Auxiliary Current (> 1kHz)	Scaled with frequency per IEC 60601-1:2012, but does not exceed 10mA NC/SFC			
Patient Leakage Current	< 10 μA NC, < 50 μA SFC			
Total Patient Leakage Current	< 50 μA NC, < 100 μA SFC			

NC – Normal Condition

SFC – Single Fault Condition (as defined in IEC 60601-1:2012)

Total Patient Leakage Current – Measurement of patient leakage current with all patient outputs connected together

#### **High Frequency Leakage (IEC 60601-2-2)**

Bipolar	< 68.9 mARMS		
Monopolar measured directly at the ESU terminals	< 100 mARMS		
LigaSure™/BPR measured directly at the ESU terminals	< 100 mARMS		

#### REM Contact Quality Monitoring System

Interrogation frequency	68–75 kHz		
Interrogation current	< 100 µA RMS		
Interrogation Voltage	< 12 VRMS		
Impedance Sense Range	5 Ω to 135 Ω		
Impedance Accuracy (RF not activated)	±7Ω		
Impedance Accuracy (RF Activated)	Greater of $\pm$ 14 $\Omega$ or 20%		

#### **Backup Power**

The Valleylab™ FT10 energy platform retains all user programmed features, calibration, and statistical data when switched off and unplugged. The Valleylab™ FT10 energy platform operates within specification when switched over to a supplied-line power by hospital backup systems.

#### **Output Characteristics**

Mode	Rated Load $(\Omega)$	Rated Output Power (W)	Peak Voltage	Current RMS Max	Crest Factor*	Duty Cycle
Monopolar CUT	(12)	1 0 1 1 1 1 1	reak voltage	TVIGA	Crest ractor	Duty Cycle
PURE	300	300	910	1.25	1.42	100%
BLEND	300	200	1100	1	2.5	50%
VALLEYLAB	300	200	1549	1	3.8	25%
Monopolar COAG						
FULGURATE	500	120	3135	1	5.7	6.25%
SPRAY	500	120	3575	1	6.5	4.76%
SOFT	100	120	240	1.55	1.42	100%
Bipolar						
LOW (1-15 W)	100	15	88	1	1.42	100%
MEDIUM (16-40 W)	100	40	143	2	1.42	100%
HIGH (45-95 W)	100	95	310	2	1.42	100%
LigaSure™						
LIGASURE	20	350	163	5.5	1.42	N/A
<b>Bipolar Resection</b>						
CUT	500	375	495	2.4	1.42	100%
COAG	100	175	212	3.2	1.42	100%

#### **Output Waveforms**

TissueFect<sup>™</sup> Tissue Sensing Technology, an automatic adjustment, controls all modes. As tissue resistance increases from zero, the energy platform outputs constant current followed by constant power followed by constant voltage. The maximum output voltage is controlled to reduce capacitive coupling and video interference and to minimize sparking.

- 1. Based on Covidien memo: "LigaSure Data Sources for VLFT10 White Papers." September 2015 RE00025819 Rev A
- 2. Data from product validation testing. Covidien report, "R0064457, LigaSure™ Renal Bench Burst Pressure Evaluation of the Valleylab™ FT10." May 29, 2015
- 3. Based on Covidien In-vivo GLP Acute report: "Verification Report GLP Acute Animal Lab LigaSure™ Preclinical Evaluation of Valleylab™ FT10." May 19, 2015: RE00005503; Report page 4, Attachment pg. 33-39
- 4. Based on product validation testing. Covidien report, "Product Validation of Valleylab™ FT10 Surgeon & Nurse Evaluation in Simulated Use." May 26, 2015; RE00005401 Rev A; pg. 11

