

Detalii specifice și standarde tehnice

LOT „Procurare de Termociclor”:

№	Name of product	Minimal technical specifications	Offered specifications; model, manufacturer.
1	Termociclor	<p>The thermal cycler for PCR with 96 positions for 0.2 ml tubes. Power configurations: VAC 220-240V, 50 Hz.</p> <ul style="list-style-type: none"> - Memory for minimum 5 PCR programs, that can be edited manually. - Temperature range: 04-100°C; - Real-time graphical display of actual and set block temperatures, - Real-time digital display of all sensor values; - Temperature accuracy: +/-0.5°C; - Heating rate: up to 6°C/s; - Cooling rate: Up to 3°C/s; - Calibrated to traceable international standards. 	<p>The thermal cycler for PCR with 96 positions for 0.2 ml tubes. Power configurations: VAC 220-240V, 50 Hz.</p> <ul style="list-style-type: none"> - Memory for 1000 PCR programs, that can be edited manually. - Temperature range: 0-100°C; - Real-time graphical display of actual and set block temperatures, with touchscreen, - Real-time digital display of all sensor values; - Temperature accuracy: +/-0.25°C; - Heating rate: up to 6°C/s; - Cooling rate: Up to 6°C/s; - Calibrated to traceable international standards. <p>Model: VeritiPro Dx Thermal Cycler, Manufacturer: ThermoFisher Scientific, Singapore</p>
	Mandatory requirements	<ul style="list-style-type: none"> - Original manufacturer packing for all items; - Quality Certificates (CE and/or DC) or any other equivalent; - ISO 9001-2015 	<ul style="list-style-type: none"> - Original manufacturer packing for all items; - Quality Certificates (CE and/or DC) - ISO 9001-2015

- **Toate bunurile vor fi testate de beneficiar. Acceptarea bunurilor și semnarea actelor de recepție va avea loc după testarea pozitivă efectuată de beneficiar.**

NUMELE FURNIZORULUI GBG MLD SRL

Semnătura și ștampila autorizate _____ Tudor Ceaicovschi, director general

Locul: Chisinau

Data: 18.08.2023