

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Laboratory Equipment (Microscope)
Name and address of the applicant	<b>Leica Microsystems CMS GmbH</b> Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems Ltd. Shanghai Floor 1, 2, 3A, 4A, and 6, Building T20-1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade Zone, 201206 Shanghai, PEOPLE'S REPUBLIC OF CHINA  Zhang Jia Gang Optical Instrument Co; Ltd Tangqiao Town, Luyuan Nanlin Development Zone, 215616 Zhangjiagang City, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated voltage: 12 VDC Rated power: 18 W Protection class: III
Trade mark (if any)	Leica
Model/type Ref.	DM1000 LED
Additional information (if necessary)	Certificate DE 3 – 30741 issued 2018-06-26 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 61010-2-101:2018
as shown in the Test Report Ref. No. which forms part of this certificate	713229324

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0277 Rev. 00

Date, 2022-01-28



( Erich Thurner )