



Bio-Rad  
Certificate of Analysis

---

Material Description:	CEFOTAXIME-CLAV.ACID 30-10 G 4X50 DISKS
Material Number:	12012125
Batch Number:	64572379
Manufacture Date:	2023-10-31
Expiration Date:	2025-11-29

---

**Characteristic**

**Results**

Labeling	Specifications see above	Conforms
----------	--------------------------	----------

Acronym	CCT 40	Conforms
---------	--------	----------

Ink colour	Green	Conforms
------------	-------	----------

.....  
Acceptable increase in zone  
Diameter for ATB tested alone  
versus its zone when tested in  
combination with clavulanic acid  
.....

Escherichia coli ATCC_25922	2 mm increase	Conforms
-----------------------------	---------------	----------

Klebsiella pneumoniae ATCC_700603	3 mm increase	Conforms
-----------------------------------	---------------	----------

Refer to current CLSI guidelines



Bio-Rad  
Certificate of Analysis

---

Material Description:	CEFOTAXIME-CLAV.ACID 30-10 G 4X50 DISKS
Material Number:	12012125
Batch Number:	64572379
Manufacture Date:	2023-10-31
Expiration Date:	2025-11-29

---

**Characteristic**

**Results**

Herve Gamba

---

This certificate has been verified and electronically approved by an authorized Quality representative. This e-signature is performed within a secure system.

2023-11-13 / 13:56:15 UTC

---

Date/Time