

Product Specifications

CL03.0216

HYDREX

Buffer pH 7.00

Tampão pH 7.00

Tampon pH 7.00

Buffer pH 7.00

Puffer pH 7.00

Tampón pH 7.00

Tampone pH 7.00

**For laboratory use, pH measurements, traceable to SRM from NIST
pH 7.00 ($\pm 0.02/20^{\circ}\text{C}$) Phosphate****Traceability: Reference standards acc. NIST SRM 189c, 188, 185i,
186lg, 186llg, 187e**

Mol weight:

Density: 1,01 g/ml

HS Nr: 38229000

CAS Nr:

UN:

ADR:

Specifications:

pH	= 7.13 @ 0°C
pH	= 7.05 @ 10°C
pH	= 7.00 @ 20°C
pH	= 6.99 @ 25°C
pH	= 6.98 @ 30°C
pH	= 6.97 @ 40°C
pH	= 6.96 @ 50°C
pH	= 6.96 @ 60°C
pH	= 6.97 @ 70°C
pH	= 6.98 @ 80°C
pH	= 7.00 @ 90°C