ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: HBsAb Quantitative EIA Test Kit

PURCHASE NO.: 37740066B

BATCH NO.: 2003706

EXPIRATION DATE: 2021/03/04

QUANTITY: 33kits

SPECIFICATION		STANDARD	CONCLUSION
Appearance		No obvious sediments or floccules in liquid component and color meets requirement	Complies
Sensitivity Standard	10mIU/mL	0.105<OD<0.400	Complies
	100mIU/mL	0.500<OD<1.600	Complies
	200mIU/mL	OD>1.200 and OD>100.0mIU/mL OD	Complies
Compliance of Negative Standard		OD < Calibrator 2# OD	Complies
Calibrator 1#		OD < 0.105	Complies
Calibrator 2#		OD > 0.105	Complies
Calibrator 3#		OD > 0.250	Complies
Calibrator 4#		OD > 0.500	Complies
Calibrator 5#		OD > 1.000	Complies
R value		R > 0.98	Complies
Precision		CV < 15%	Complies
The above results based on evaluation on the whole kit with all components manufactured by ACON			

Quality Department: Antony. Hu

Date: Apr. 21, 2020

Digitally signed by Monastîrschii Viorica
Date: 2021.03.02 15:14:38 EET
Reason: MoldSign Signature
Location: Moldova

REPUBLICA MOLDOVA, mun. CHISINAU
S.R.L.
"SANMEDICO"
IDNO 1006...
SOCIETATEA CU RĂSPUNDERE LIMITATĂ