

Technical Specification No

Manual tourniquet system

- Manufacturer: Arimedical
- Country of Origin : China
- Model:
- Certification ISO , CE ,
-

Technical Specification	Yes	No	Remarks
Configuration <ul style="list-style-type: none"> ➤ manual tourniquet system -1 piece , portable , light wight, reusable cuffs 1 pc for adult and 1 pc for children, with transport bag 	X		<ul style="list-style-type: none"> ➤ manual tourniquet system -1 piece , portable , light wight, reusable cuffs 1 pc for adult and 1 pc for children, with transport bag
Certification:			
<ul style="list-style-type: none"> ➤ ISO 9001 ➤ ISO 13485 ➤ CE ➤ CE Declaration of Conformity 93/42/EEC 	X X X X		<ul style="list-style-type: none"> ➤ certificate attached ➤ certificate attached ➤ certificate attached ➤ certificate attached
Technical Specification	Yes	No	Remarks
Other services included in offer			
<ul style="list-style-type: none"> ➤ transport to the final customer ➤ instalation and put in function ➤ training for the technical staff for final customer ➤ training for the medical staff for final customer ➤ warranty for 24 months 	X X X X X		<ul style="list-style-type: none"> ➤ bidder declaration attached ➤ bidder declaration attached ➤ bidder declaration attached ➤ bidder declaration attached ➤ bidder declaration attached
GUARANTEE			
<ul style="list-style-type: none"> ➤ 24 months guaranteed from signing the receiving minutes 	X		<ul style="list-style-type: none"> ➤ bidder declaration attached
SERVICE AND MAINTENANCE			
<ul style="list-style-type: none"> ➤ the bidder has technical facilities necessary for the service both for ambulances as well as for medical equipment according to the general guarantee 	X		<ul style="list-style-type: none"> ➤ bidder declaration attached

<p>conditions and the user guide of the manufacturer</p> <p>➤ Maximum time for technical intervention – 48 hours.</p> <p>➤ Maximum time for remedial measures, in total 72 hours.</p> <p>➤ The technical service and current repair will be performed without waiting in a queue.</p> <p>➤ Temporary replacement of the equipment has to be ensured according to the periods mentioned above.</p> <p>➤ During the guarantee period, at the reasonable request of the user, repair, adjustments and maintenance of medical equipment ,according to the specifications in the manufacturers' guides, will be done free of charge.</p> <p>➤ The spare parts and workmanship are for free, except from the consumables for vehicles established by the manufacturer</p>	<p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p>		<p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p>
Technical Specification	Yes	No	Remarks
AVAILABILITY OF SPARE PARTS			
<p>➤ the bidder will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows:</p> <ul style="list-style-type: none"> ○ spare parts free of charge, including the workmanship for the 36 months of guarantee period ○ for the rest of the period – against payment for at least 10 years period is fulfilled. <p>➤ The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary</p>	<p>X</p> <p>X</p> <p>X</p> <p>X</p>		<p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p>
GUIDES			
<p>➤ Technical Guide and User Guide.</p> <p>➤ All guides will be available in Romanian /Russian.</p>	<p>X</p> <p>X</p>		<p>➤ bidder declaration attached</p> <p>➤ bidder declaration attached</p>