

LIMITED LIABILITY COMPANY
“SCIENCE-TECHNICAL CENTRE “ENERGODIAGNOSTIKA”

Legal address: room 8, office XXI, floor 2, Borisovskaya st. 1, MOSCOW, RUSSIA, 105187

Actual address: office II, Stroiteley st. 7, REUTOV, RUSSIA, 143966

Tel/Fax: +7-495 791 78 83, +7-498 661-61-35

e-mail: mail@energodiagnostika.ru <http://www.energodiagnostika.ru>

Checking of measuring instruments Certificate No. RA.RU 310546

CALIBRATION CERTIFICATE

No. C-ВЯП/20-12-2024/397209975

Valid until:
19th of December 2025

Measurement Item: Magneto metrical device for determining the concentration of stresses «TSC»

Federal Register Fund Number: 20675 – 11

Serial Number: 5M-32-052

in the set of scanning devices with: Type 1-8M (№ 1-8M-710), Type 6-12 (№6-12-111), Type 11-6K (№057),
Type 4 (w/n)

Verified for magnetic field strength measurements in the range \pm (10 to 1999) A/m

According to verification procedure TSC.000.000.000 OM, Appendix 2nd

Using standard gauge registration No. 28808–05, magnetic induction measure YKM-1, ser. No. 01

Measurement class 2nd Class, in range 0,012...2,5 mTl ME \pm 1,5 %

Influent factor values Atmospheric pressure – 098 kPa, Air temperature – 19°C, Relative humidity – 43%

Based on the results of periodical verification the above mentioned measurement items recognized as valid

Verification mark



Data Entering Number 397209975

to the FRF Ensuring Unity of Measurements <https://fgis.gost.ru/fundmetrology/cm/results/1-397209975>

General Director

Dubov Alexander.

Verification engineer

Dubov Andrey.

Verification date: 20th of December 2024