

## Applicability:

Test	Applicable:
1. Safety, Finishing, Start-up, Service and Remote Support Tests	<input type="checkbox"/>
2. OTC (Vision C/Air) Tests	<input type="checkbox"/>
3. FMTS (V) Tests	<input type="checkbox"/>
4.1 Bucky Tests	<input type="checkbox"/>
4.2 Wall Stand Tests	<input type="checkbox"/>
4.3. Bucky Table Tests	<input type="checkbox"/>
5. Portable/Mobile Positioner	<input type="checkbox"/>
6. Integrated Position Tests (V, C and Air)	<input type="checkbox"/>
7. U-Arm Tests	<input type="checkbox"/>
8. Straight Arm Tests	<input type="checkbox"/>
9. Imaging and X-ray Tests	<input type="checkbox"/>
10. Advanced Function Tests	<input type="checkbox"/>
11. Accessories Tests	<input type="checkbox"/>
12. Specific Project Tests (AED retrofit system)	<input type="checkbox"/>

## 1. Safety, Finishing, Start-up, Service and Remote Support Tests

### Test SU. Safety, Finishing and Start-up Test Results

Tests	Result
<b>Test SU.1</b> Earthing measured impedance is <100 mΩ (SAT only)	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test SU.2</b> Surfaces are smooth, clean and cold with no scratches or discolouration. Lights work comfortably bright. Cover gaps small and uniform. The buttons react to pressing.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test SU.3</b> The cable sleeves are intact, the cables are fastened, they do not cross sharp surfaces and cannot snag device motion or trip someone.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test SU.4</b> Device, safety and company/product labels present, intact and clearly visible. Correct device S/N visible in the console software. Check specific device requirements in VT-V10 Vision product labelling.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test SU.5</b> All components power up, applications are accessible, the tablet console and LEDs are on, user interfaces responsive are, the System state is on "Stand By". The buttons on the stand respond. The language and metric system conform to the device location.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test SU.6</b> All components power down, no LEDs are active, the console shuts down. Only the backup power on the EP remains present (LEDs on the VXS). The stand cannot be moved easily.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

<b>Test SU.7</b> Operator access is available. Avanse and Stand App auto-start within 30s. The system is unavailable to the Operator. The system can be accessed remotely with standard Admin access. The Service Shell loads automatically.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test SU.8</b> Activating the Emergency stop button on the mini console stops all movement of the device.	F <input type="checkbox"/> P <input type="checkbox"/>	N/A <input type="checkbox"/>

**Safety, Finishing and Start-up Notes:**

### 3. FMTS (V) Test results

Tests	Result
<b>Test FM.1</b> The brakes release on all release handle contact, quietly. The motion is smooth and quiet over the full range. The cables bend without resistance. The brake holds when the handle is released. No rotation backlash.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test FM.2</b> The rotation brake release quietly, the motion is balanced and smooth. The brake holds position when the button is released. Mechanical indexing at the horizontal and vertical beam allows easy sliding in and out.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test FM.3</b> On the touchscreen Hand PA selection the stand shows Free position. When turned to WS bucky the console identifies Stand position (voice and on UI).	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test FM.4</b> The touchscreen changes the exposure technique in < 2s. The battery never goes below 50% and the tablet does not loose connection.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test FM.5</b> By pressing the brake release button for the third axis, the tube goes slightly upwards or holds current position.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

FMTS Notes:

### 4. Detector Stand Test Results

#### 4.1 Bucky Test results

Repeat the tests for every bucky on the system.

Tests	Result
-------	--------

<b>Test BC.1</b> An Orange LED turns on when a grid is inserted and turns off when it's removed. Avanse shows a grid code consistent with the grid frame.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test BC.2</b> The bucky trolley moves smoothly and locks into place in the fully out and fully inserted positions.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test BC.3</b> The detector slides easily into the bucky trolley and comes out without resistance.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test BC.4</b> The detector rotates smoothly and locks into the portrait and landscape positions. Blue and green LEDs and the Avanse bucky indication show Portrait/Landscape indicators correctly.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>

## 4.2 Wall Stand Test results

### Test WS: WS Positioning and Safety

Tests	Result
<b>Test WS.1</b> The elevation brake releases quietly on the handle grip. Motion is smooth, balanced and quiet over full range. The cables bend without resistance. The brake engages when the handle is released and holds fast.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test WS.2</b> The tilt brake releases quietly, motion is smooth, balanced and quiet over full range. The cables bend without resistance. Moving into and out of 90° indexed positions is smooth. The brake holds the bucky with no backlash.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test WS.3</b> The system auto-positions smoothly and accurately, without stopping or errors, the status shows READY→MOVING→IDLE. When Move is released the system stops.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test WS.4</b> Motorised WS elevation and bucky tilt are smooth and quiet on press and stop when button is released. The bucky stops automatically in horizontal and vertical positions.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test WS.5</b> On emergency stop the WS stops immediately. It cannot be moved again until the EM stop is released.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

## 4.3 Bucky Table Test Results

Tests	Result
<b>Test BT.1</b> The table-top brake releases quietly, floating is smooth and quiet in full range. The brake holds and great force is needed to move the table top in any direction.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
<b>Test BT.2</b> The bucky moves with no resistance, smoothly and quietly over the entire range and stays in position when released.	P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

<b>Test BT.3</b> The table auto-positions smoothly and accurately, without stops or errors. When Move is released the table stops.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test BT.4</b> The table elevates on single/double click smoothly and quietly from top to bottom. It stops as soon as the footswitch is released.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test BT.5</b> On emergency stop, the table stops immediately. It cannot be moved again until the EM stop is released.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>

**Bucky, Wall Stand and Bucky Table Notes:**

## 6. Integrated Position Test Results (V, C and Air)

### Test IP.1 Auto-positioning and Accuracy WS

Tests	Result		
<b>Test IP.1.1</b> On remote press, the stand auto-positions onto WS smoothly and accurately centred, without stopping or errors. When the remote is released the system stops. It carries on to final position when pressed again. In the final position the stand status is green (IDLE).	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.2</b> On SID increase, the system remains centred onto WS within 10mm.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.3</b> The WS SID value on the collimator tape should be around 4cm less than the value shown on the tube-side console display .	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.4</b> On manual tube height adjustment the stand automatically adjusts the bucky height to match the beam height and stops when centred.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.5</b> On manual WS height change, the OTC adjusts the tube height to match the beam height and stops when centred to WS bucky	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.6</b> The stand adjust the bucky height to match the manually inclined tube beam and stops when centred.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.7</b> When pointing at an empty WS bucky, the stand status is RED. Exposure is not possible.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.8</b> When the WS and beam are not in sync, the stand status is RED. Exposure is not possible.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.1.9</b> The system positions the tube smoothly above the horizontal WS bucky programmed position.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

## Test IP.2 Auto-positioning and Accuracy Bucky Table

Tests	Result		
<b>Test IP.2.1</b> The system auto-positions into Table smoothly and accurately centred, without stopping or errors. In the final position the stand status is green (IDLE).	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.2</b> While moving the tube left/right, the beam remains centred onto table within 10mm.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.3</b> Table SID value on the collimator tape should be around 4cm less than the value shown on the tube-side console display .	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.4</b> . On manual tube translation left/right the table automatically adjusts the bucky position to sync with the inclined beam.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.5</b> . On manual tube angle adjustment the table automatically adjusts the bucky position to sync with the beam.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.6</b> When the table is elevated under the tube, the tube moves simultaneously with the table up and down.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.7</b> When pointing at an empty table bucky, the stand status is RED. Exposure is not possible.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IP.2.8</b> When the table bucky and beam are not in sync the stand status is RED. Exposure is not possible.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b><u>Integrated Positioning Notes:</u></b>			

## 9. Imaging and X-ray Test Results

### IM.1 Console, Exam Entry, APR, System Status, Exposure

Tests	Result		
<b>Test IM.1.1</b> The entered patient data appears on top of the screen, requested procedures are added. The system state on entry is on Stand by.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.1.2</b> The status box exists for each detector. Portable detectors go offline when the battery is removed and return automatically when re-inserted.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.1.3</b> Selecting a projection or patient size updates the exposure technique from APR within 2s. Smaller and paediatric size patients reduce the exposure settings.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.1.4 (SAT Only)</b> The worklist search shows the test exam and it loads correctly.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.1.5</b> kV allows setting of 40 to 125/150, mA 10 to generator power x 1.2, e.g. 630 for 50kW. The generator allows preparation at max power of 100kV, the state goes into PREPARATION (Yellow).	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

**Imaging and X-ray Notes:**

### IM.2 Collimation

Tests	Result		
<b>Test IM.2.1</b> When the light is turned on it turns off automatically after ~30s. The blades move smoothly with no resistance. When fully opened at 1m SID, the whole detector active area is covered. The cross-hairs are in the centre of the light field.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.2.2</b> The system goes through PREPARATION into EXPOSURE (red) state. The acquired image appears on the screen. The lead markers are in the corners of the exposed area shown in the image.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.2.3</b> The light field is slightly smaller than 20x20cm (95%) for both SID values.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.2.4</b> The filters change manually or automatically.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.2.5</b> The collimator rotates smoothly with a mechanical index when facing the front.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

## IM.3 Exposure Accuracy, AEC, DAP, Detector Tests results

### IM 3.1 kV Accuracy (SAT Only)

Test configuration:

- Free technique, the collimator pointing at the floor at 1m height, no detector, no filtration
- The X-ray instrument positioned in the center of the light field with collimator opening set to 10 x 10 cm

Focus	mA	mAs	kV <sub>target</sub>	kV <sub>actual</sub>	Tolerance [%]
SF	100	20	50		
SF	100	20	70		
SF	100	20	100		
SF	100	20	150		
LF	250	20	50		
LF	250	20	70		
LF	250	20	100		
LF	250	20	150		
SF	10	20	80		
SF	160	20	80		
LF	200	20	80		
LF	500	20	80		
<b>Test 1:</b> All exposures are without error. The values from the x-ray instrument are within $\pm 10\%$ of the set values compared to the values returned by the generator					P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

### kV, Exposure Time and mAs Accuracy and Dose Level (FAT Only)

The set doses are not the same for each generator. Pay attention to the focal spot (SF-small focus, LF – large focus). If the focal point doesn't match for the given values (kV<sub>target</sub>, mA<sub>target</sub>, mAs<sub>target</sub>), adjust them and write them in the table below. To obtain LF, it is best to adjust mA.

	SF	LF	
kV <sub>target</sub>	60	80	
mA <sub>target</sub>	1.6	4	
mAs <sub>target</sub>	160	+/- 400	
mS <sub>target</sub>	10	-/+ 10	
kV <sub>actual</sub>			

Tolerance [%]			
All exposures without error. Values from the x-ray instrument within $\pm 10\%$ of set values compared to the values returned by the generator.			P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
mAs <sub>actual</sub>			
Time <sub>actual</sub> [ms]			
Both exposures without error. Time and mAs of each exposure in $\pm 10\%$ of target.			P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>
Dose [ $\mu\text{Gy}$ ]			
Tolerance [%]			
Draw a ROI box over the image. Compare the dose reading from the measurement instrument to the dose value indicated in the ROI box. The dose measurements ( $\mu\text{Gy}$ ) should be within 10% of each other.			P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>

**Note (Vision M only):** Exposure doses: 1) 60kVp; 50mA; 5mAs; 10ms (SF/LF N/A)

### IM 3.2 Dose Reproducibility (SAT Only)

Test configuration:

- Free technique, no detector, collimator into floor 1m SID to instrument
- X-ray instrument in the center of the light field, collimator opening 10 x 10 cm
- 1.5 mm CU additional filtration into the collimator
- Exposure technique: 77kV 200mA 10mAs

	Actual value	Tolerance [%]
Exposure 1 / Dose [ $\mu\text{Gy}$ ]		
Exposure 2 / Dose [ $\mu\text{Gy}$ ]		
Exposure 3 / Dose [ $\mu\text{Gy}$ ]		
Average Dose [ $\mu\text{Gy}$ ]		
Maximum tolerance + 10%		
Minimum tolerance - 10%		
All exposures are without error. The Dose value of each exposure is within $\pm 10\%$ of the average measurement.		
P <input type="checkbox"/> F <input type="checkbox"/> N/A <input type="checkbox"/>		

### IM 3.3 AEC (SAT Only)

Test configuration:

- Put a 25 mm Al filter under the collimator
- Place the X-ray instrument next to AEC field, no grid
- For the second test add another 1 mm Cu filter under the collimator

	kV	mA	Dose [μGy]	1. field [μGy]	2. field [μGy]	3. field [μGy]	4. field [μGy]	5. field [μGy]				
1	70	160										
<i>The same test with added 1 mm Cu under collimator.</i>												
2	70	160										
The tolerance range is ≤10% between fields, within ±10% of the calibrated AEC value and the same values with the added Cu filter.							P	<input type="checkbox"/>	F	<input type="checkbox"/>	N/A	<input type="checkbox"/>

### IM 3.4 DAP and Detector

Test configuration:


- Position the beam to point at the detector directly, set exposure values to 70kV, 5mAs, 1mm Cu filter, SID 100cm, no grid, the collimator opened to full size of the detector, the X-ray measuring instrument is positioned in the light field.

#### DAP

	kV	mA	mAs	1.measurement [mGycm <sup>2</sup> ]	2.measurement [mGycm <sup>2</sup> ]	3.measurement [mGycm <sup>2</sup> ]			
2	70	200	20						
Test 1: The difference between 3 measurements is ≤20%					Test Evaluation	P	<input type="checkbox"/>	F	<input type="checkbox"/>

#### DETECTOR

- For this test, ensure that the tube and detector are perpendicular to each other, use a spirit level or digital protractor.
- If the PVI value in any ROI ≥50,000 you have overexposed the detector and have to repeat the test: check the filter, adjust exposure parameters to achieve lower dose.

Flatness - Draw 9 ROI boxes on the image	Max	Min	Mean	$\frac{\text{Max-Mean}}{\text{Mean}}$	$\frac{\text{Mean-Min}}{\text{Mean}}$				
									
Test 2: $(\text{Max-Mean})/\text{Mean} < 0.3$ , $(\text{Mean-Min})/\text{Mean} < 0.3$ <b>Note (Vision M Only):</b> Expose at: 70kVp; 5mAs.				P	<input type="checkbox"/>	F	<input type="checkbox"/>	N/A	<input type="checkbox"/>

Exposure Notes:

Test IM.4 Image Quality, Orientation, Processing, Annotations and Export Test Results

Tests	Result	
<b>Test IM.4.1</b> On Hand PA exposure the system goes through preparation (yellow) then exposure (red), emitting a beep. The QA phantom image is displayed shortly after, processed using Extremities process group.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.2</b> The highest detector resolution is clearly visible on the QA phantom image.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.3</b> All 6 contrast phantom steps are visible on the QA phantom image.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.4</b> The annotations list offers only "L" and "R".	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.5</b> The top left of the QA phantom is shown on the top left of the displayed image.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.6</b> Exam burning to CD completes without errors. The CD automatically opens on another PC and displays the exam correctly.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.7</b> On close exam, images are sent to PACS without error and are accessible from the local archive.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test IM.4.8</b> (SAT Only) The image prints accurately with correct annotations.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>

Imaging Notes:

10. Advanced Imaging Functions Tests Results

AD Long Anatomy Imaging

Tests	Result	
<b>Test AD.1</b> Auto-stitching completes without error, the system positions and exposes in the correct order. Final long image is produced automatically and stitched correctly with continuous ruler.	P <input type="checkbox"/> F <input type="checkbox"/>	N/A <input type="checkbox"/>

<b>Test AD.2</b> The stitching tool stitches the manually positioned long anatomy images automatically and the ruler is continuous..	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
--	----------------------------	----------------------------	------------------------------

Advanced Imaging Function Notes:

## 11. Accessories Test Results

### Test AC.1 Mobile Table

Tests	Result		
<b>Test AC.1.1</b> The table top floats smoothly in all directions when unlocked	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test AC.1.2</b> The table top floats smoothly with a load of >130kg	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

### Test AC.2 Long Anatomy (Stitching) Platform Test Results

Tests	Result		
<b>Test AC.2.1</b> The platform moves smoothly in all directions when unlocked. When locked the platform doesn't move.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test AC.2.2</b> The platform doesn't deform (or deform slightly) when force is applied on the plastic back. The adjustment of the ruler is easy.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test AC.2.3</b> The footrest works properly in both up and down positions.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test AC.2.4</b> The ruler markings are seen clearly in the image when exposed.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

## 12. Specific Project and Imaging Test Results

### Test SP.1 AED Retro-fit System Tests Results

Perform tests IM.1.1 – IM1.3, IM4.1 – IM4.5, IM5.1 – IM5.2

Tests	Result		
-------	--------	--	--

<b>Test SP.1.1</b> There are no scratches or discolouration on any surface. All surfaces are smooth and clean, with no sharp edges or other snag points. No surfaces are hot.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test SP.1.2</b> The system starts up normally, the detector connects and is ready for exposure.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test SP.1.3</b> The system goes into sleep and recovers without an error, the detector re-connects and can acquire an image.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>
<b>Test SP.1.4</b> The detector acquires the image.	P <input type="checkbox"/>	F <input type="checkbox"/>	N/A <input type="checkbox"/>

<b>Product Acceptance Test type:</b>	<b>FAT / SAT</b>
<b>Test passed:</b>	<b>YES / NO</b>
<b>Tested By (Name/Signature):</b>	
<b>Date:</b>	