

26-30, NISHIAZABU 2-CHOME, MINATO-KU, TOKYO 106-8620 JAPAN



ΕN

## **EU Declaration of Conformity**

12-CEM-01-L

Manufacturer:

**FUJIFILM** Corporation

single registration number (SRN)

**TBD** 

Address:

26-30, Nishiazabu 2-chome, Minato-ku,

Tokyo 106-8620, JAPAN

Authorized Representative:

FUJIFILM Europe GmbH

single registration number (SRN)

TBD

Address:

Heesenstrasse 31

40549 Duesseldorf, GERMANY

Basic UDI-DI:

454741010100000000000353K

Trade name:

FUJI MEDICAL DRY LASER IMAGER DRYPIX Smart

Model Number:

DRYPIX6000

Applicable Products:

xxx51978 or later

Classification (MDR, Annex VIII):

Class I (Rule 13)

We, FUJIFILM Corporation, herewith declare in our sole responsibility that the product(s) identified in this declaration conforms to the provisions of the following Regulation(s) and Directive(s).

Regulation:

Medical Device Regulation: REGULATION (EU) 2017/745 and their Annexes.

RoHS Directive: 2011/65/EU,(EU)2015/863

Common specifications ('CS'): No references to any CS

Conformity Assessment Procedure for Regulation (EU) 2017/745:

Annex IV

Place and Date of issue

Signature:

Naotake Mitsumori

Kanagawa, JAPAN

Name: Function: General Manager,

2021-05-21

Qulity Assurance and Regulatory Affairs Division,

Medical Systems Business Division

**FUJIFILM Corporation**