

Declaration of Conformity to Council Directive 93/42/EEC concerning Medical Devices

Manufacturer: Beijing Choice Electronic Technology Co., Ltd.
Room 4104, No.A12 Yuquan Road, Haidian District,
100143 Beijing, P.R.CHINA

European Representative: Shanghai International Holding Corp. GmbH (Europe)
Eiffestraße 80 20537 Hamburg GERMANY

Product: Fingertip Pulse Oximeter

| | |
|---------------------|------------------------|
| Customers type no.: | Manufactures type no.: |
| Microlife OXY300 | MD300C2G/MD300C29 |

UMDNS Code: 17148

Classification: Class IIa, rule 10 to Annex IX of the MDD

Conformity assessment Route: Annex II excluding (4)

**We, the manufacturer, herewith declare that the stated medical devices
meet the transposition into national law, the provisions of Council Directive
93/42/EEC concerning medical devices.**

All supporting documentation is retained at the premises of the manufacturer.

Standards applied:

EN ISO 13485:2012 Medical devices- Quality management systems- Requirements for
regulatory purposes

EN ISO14971:2012 Medical devices –Application of risk management to medical
devices

EN ISO10993-1: 2009/AC:2010 Biological evaluation of medical devices - Part 1:
Evaluation and testing within a risk management system

EN ISO10993-5:2009 Biological evaluation of medical devices - Part 5: Tests for in vitro
cytotoxicity

EN ISO10993-10:2010 Biological Evaluation of Medical Devices-Part 10: Tests for irritation and delayed-type hypersensitivity

EN 60601-1:2006/A1:2013 Medical electrical equipment-Part1: General requirements for safety and essential performance

EN 60601-1-2:2007/AC:2010 Medical electrical equipment-Part1-2: General requirements for safety and essential performance Collateral Standard: Electromagnetic compatibility – Requirements and tests

ISO 80601-2-61:2011 Medical electrical equipment-Particular requirements for basic safety and essential performance of pulse oximeter equipment for medical use

EN1041:2008 Information supplied by the manufacture of medical device

EN 980:2008 Symbols for use in the labelling of medical devices

EN 62304: 2006 Medical device software-Software life-cycle processes

EN60601-1-6: 2010 Medical electrical equipment-Part 1-6: General requirements for basic safety and essential performance-Collateral Standard: Usability

Notified Body:

TÜV SÜD Product service GmbH
Ridlerstr 65, D-80339 München, Germany

Identification Number:

CE 0123

(EC) Certificate(s):

No. G1 16 02 78179 025

Start of CE-marking:

2014-02-14

Place, Date of Declaration:

Beijing, 2017-01-13

Signature:

Lei Chen

Name: Lei Chen

Position: Quality Director