



# CERVICAL BRUSH standard sterile

TD-35

att.no.3 pt.2 subpt.2.3

revision 3

date: 2020-03-02

amendment of: 2023-02-01

**MANUFACTURER:** ZARYS International Group sp. z o.o. sp.k.


**ADDRESS:** ul. Pod Borem 18, 41-808 Zabrze, Poland, phone +48 32 376 07 00-02,06 fax. +48 32 370 38 08

## INDICATION:

- Device designed for obtaining cell smear from the cervix (cervical canal and ectocervix) and from the transformation zone

## FEATURES:

- straight type
- soft bristle tips to prevent cell samples from getting damaged
- made of polypropylene (handle), stainless steel (core) and nylon (bristle)
- optimal outreach thanks to the long handle
- PVC-free
- latex-free
- phthalate-free
- EO sterilized
- packaging: 1 pc./paper-foil



## SIZE AND REFERENCE

CATALOGUE NUMBER	TYPE	SALES PACKAGING	TRANSPORT PACKAGING
SC-001	standard	100 pcs	10 x 100 pcs