



Directive 93/42/EEC on Medical Devices, Annex II Section 4

No. CE 632827

Issued To: Abbott Vascular

3200 Lakeside Drive

Santa Clara California 95054 USA

In respect of:

XIENCE PRO Everolimus-eluting Coronary Stent Systems

BSI has performed a design examination on the above devices in accordance with the Council Directive 93/42/EEC, Annex II Section 4. The design conforms to the requirements of this directive. For marketing of these products an additional Annex II excluding Section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 2797):

Gary E Slack, Senior Vice President Medical Devices

Gay C Stade

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 1 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.





Supplementary Information to CE 632827

Issued To: Abbott Vascular 3200 Lakeside Drive

Santa Clara California 95054 USA

Device Name: XIENCE PRO 48 Everolimus Eluting Coronary Stent System

Intended purpose per IFU:

Indications:

The XIENCE PRO 48 Everolimus Eluting Coronary Stent System is indicated for improving coronary luminal diameter in the following:

- Patients with symptomatic ischemic heart disease due to discrete de novo native coronary artery lesions.
- For restoring coronary flow in patients experiencing acute myocardial infarction who present within 12 hours of symptom onset.
- For the treatment of patients with concomitant diabetes, acute coronary syndrome, dual vessel lesions (two lesions in two different epicardial vessels), lesions residing within small coronary vessels; lesions where treatment results in the jailing of side branches (lesions with a side branch < 2 mm in diameter or an ostial stenosis < 50%); for the treatment of elderly patients (age ≥ 65), and for treatment of both men and women.
- For the treatment of patients presenting with in-stent restenosis in coronary artery lesions; chronic total occluded coronary artery lesions (defined as coronary artery lesions with TIMI flow 0 and lasting longer than 3 months); and coronary artery bifurcation lesions.

In all cases, the treated lesion length should be less than the nominal stent length (48 mm) with a reference vessel diameter of \geq 2.50 mm and \leq 3.75 mm.

Catalog Numbers:					
Stent Diameter	Stent Length [mm]				
[mm]	48				
2.50	1017250-48				
2.75	1017275-48				
3.00	1017300-48	Caralina - Lillian			
3.50	1017350-48	600			

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 2 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.





Supplementary Information to CE 632827

Issued To: Abbott Vascular

3200 Lakeside Drive Santa Clara California 95054 USA

Device Name: XIENCE PRO^X Everolimus Eluting Coronary Stent System

Intended purpose per IFU:

Classification: Class III Implant

1076300-08

1076325-08

1076350-08

1076400-08

Indications:

3.00

3.25

3.50

4.00

The XIENCE PROX Everolimus Eluting Coronary Stent System is indicated for improving coronary luminal diameter in the following:

• Patients with symptomatic ischemic heart disease due to discrete de novo native coronary artery lesions

1076300-15

1076325-15

1076350-15

1076400-15

- For restoring coronary flow in patients experiencing acute myocardial infarction who present within 12 hours of symptom onset
- For the treatment of patients with concomitant diabetes, acute coronary syndrome, dual vessel lesions (two lesions in two different epicardial vessels), lesions residing within small coronary vessels; lesions where treatment results in the jailing of side branches (lesions with a side branch < 2 mm in diameter or an ostial stenosis < 50%); for the treatment of elderly patients (age ≥ 65), and for treatment of both men and women.
- For the treatment of patients presenting with in-stent restenosis in coronary artery lesions; chronic total occluded coronary artery lesions (defined as coronary artery lesions with TIMI flow 0 and lasting longer than 3 months); and coronary artery bifurcation lesions.

In all cases, the treated lesion length should be less than the nominal stent length (8 mm, 12 mm, 15 mm, 18 mm, 23 mm, 28 mm, 33mm or 38mm) with a reference vessel diameter of \geq 2.00 mm and \leq 4.25 mm.

	Catalog Numbers:							
Stent Diameter		Stent Length [mm]						
[mm]	8	12	15	18	23	28	33	38
2.00	1076200-08	1076200-12	1076200-15	1076200-18	1076200-23	1076200-28		
2.25	1076225-08	1076225-12	1076225-15	1076225-18	1076225-23	1076225-28		
2.50	1076250-08	1076250-12	1076250-15	1076250-18	1076250-23	1076250-28	1076250-33	1076250-38
2.75	1076275-08	1076275-12	1076275-15	1076275-18	1076275-23	1076275-28	1076275-33	1076275-38

1076300-18

1076325-18

1076350-18

1076400-18

1076300-23

1076325-23

1076350-23

1076400-23

1076300-28

1076325-28

1076350-28

1076400-28

First Issued: 2015-04-13	Date: 2021-01-12	Expiry Date: 2024-05-26

...making excellence a habit."

1076300-33

1076325-33

1076350-33

1076400-33

Page 3 of 7

1076300-38

1076325-38

1076350-38

1076400-38

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

1076300-12

1076325-12

1076350-12

1076400-12





Supplementary Information to CE 632827

Issued To:

Abbott Vascular 3200 Lakeside Drive Santa Clara California 95054 USA

Device Name: XIENCE PRO^A Everolimus Eluting Coronary Stent System

Intended purpose per IFU:

Indications:

The XIENCE PRO^A Everolimus Eluting Coronary Stent System is indicated for improving coronary luminal diameter in the following:

- Patients with symptomatic ischemic heart disease due to discrete de novo native coronary artery lesions.
- For restoring coronary flow in patients experiencing acute myocardial infarction who present within 12 hours of symptom onset.
- For the treatment of patients with concomitant diabetes, acute coronary syndrome, dual vessel lesions (two lesions in two different epicardial vessels), lesions residing within small coronary vessels; lesions where treatment results in the jailing of side branches (lesions with a side branch < 2 mm in diameter or an ostial stenosis < 50%); for the treatment of elderly patients (age ≥ 65), and for treatment of both men and women.
- For the treatment of patients presenting with in-stent restenosis in coronary artery lesions; chronic total occluded coronary artery lesions (defined as coronary artery lesions with TIMI flow 0 and lasting longer than 3 months); and coronary artery bifurcation lesions.

In all cases, the treated lesion length should be less than the nominal stent length (8 mm, 12 mm, 15 mm, 18 mm, 23 mm, 28 mm, 33mm or 38mm) with a reference vessel diameter of \geq 2.00 mm and \leq 4.25 mm.

Classification: Class III Implant									
	Catalog Numbers:								
Stent Diameter		Stent Length [mm]							
[mm]	8	12	15	18	23	28	33	38	
2.00	1128200-08	1128200-12	1128200-15	1128200-18	1128200-23	1128200-28			
2.25	1128225-08	1128225-12	1128225-15	1128225-18	1128225-23	1128225-28			
2.50	1128250-08	1128250-12	1128250-15	1128250-18	1128250-23	1128250-28	1128250-33	1128250-38	
2.75	1128275-08	1128275-12	1128275-15	1128275-18	1128275-23	1128275-28	1128275-33	1128275-38	
3.00	1128300-08	1128300-12	1128300-15	1128300-18	1128300-23	1128300-28	1128300-33	1128300-38	
3.25	1128325-08	1128325-12	1128325-15	1128325-18	1128325-23	1128325-28	1128325-33	1128325-38	
3.50	1128350-08	1128350-12	1128350-15	1128350-18	1128350-23	1128350-28	1128350-33	1128350-38	
4.00	1128400-08	1128400-12	1128400-15	1128400-18	1128400-23	1128400-28	1128400-33	1128400-38	

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 4 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.





Supplementary Information to CE 632827

Issued To: Abbott Vascular

3200 Lakeside Drive Santa Clara California 95054 USA

Device Name: XIENCE PROS Everolimus Eluting Coronary Stent System

Intended purpose per IFU:

Indications:

The XIENCE PROS Everolimus Eluting Coronary Stent System is indicated for improving coronary luminal diameter in the following:

- Patients with symptomatic ischemic heart disease due to discrete de novo native coronary artery lesions.
- For restoring coronary flow in patients experiencing acute myocardial infarction who present within 12 hours of symptom onset.
- For the treatment of patients with concomitant diabetes, acute coronary syndrome, dual vessel lesions (two lesions in two different epicardial vessels), lesions residing within small coronary vessels; lesions where treatment results in the jailing of side branches (lesions with a side branch < 2 mm in diameter or an ostial stenosis < 50%); for the treatment of elderly patients (age ≥ 65), and for treatment of both men and women.
- For the treatment of patients presenting with in-stent restenosis in coronary artery lesions; chronic total occluded coronary artery lesions (defined as coronary artery lesions with TIMI flow 0 and lasting longer than 3 months); and coronary artery bifurcation lesions.

In all cases, the treated lesion length should be less than the nominal stent length (8 mm, 12 mm, 15 mm, 18 mm, 23 mm, 28 mm, 33 mm, or 38 mm) with a reference vessel diameter of \geq 2.00 mm and \leq 4.25 mm.

Classification: Class III Implant

	Catalog Numbers:							
Stent Diameter		Stent Length [mm]						
[mm]	8	12	15	18	23	28	33	38
2.00	1508200-08	1508200-12	1508200-15	1508200-18	1508200-23	1508200-28	1508200-33	1508200-38
2.25	1508225-08	1508225-12	1508225-15	1508225-18	1508225-23	1508225-28	1508225-33	1508225-38
2.50	1508250-08	1508250-12	1508250-15	1508250-18	1508250-23	1508250-28	1508250-33	1508250-38
2.75	1508275-08	1508275-12	1508275-15	1508275-18	1508275-23	1508275-28	1508275-33	1508275-38
3.00	1508300-08	1508300-12	1508300-15	1508300-18	1508300-23	1508300-28	1508300-33	1508300-38
3.25	1508325-08	1508325-12	1508325-15	1508325-18	1508325-23	1508325-28	1508325-33	1508325-38
3.50	1508350-08	1508350-12	1508350-15	1508350-18	1508350-23	1508350-28	1508350-33	1508350-38
4.00	1508400-08	1508400-12	1508400-15	1508400-18	1508400-23	1508400-28	1508400-33	1508400-38

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 5 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.





Supplementary Information to CE 632827

Issued To: Abbott Vascular 3200 Lakeside Drive

Santa Clara California 95054 USA

Certificate History

Date	Reference Number	Action		
13 April 2015	10154362	New Issue. Transfer from another Notified Body.		
12 January 2016	10159718	DuPont Tyvek Medical Transition Project update.		
24 November 2016	10166114	Certificate Renewal.		
10 August 2017	8695169	Various IFU updates including revised risk and clinical use information and alignment of structure and general consistency. Update symbols on labels for consistency across project families.		
22 December 2017	8868966	Add Synergy Health in Offaly, Ireland as new ETO sterilization site.		
05 March 2018	8888512	Addition of product XIENCE PRO ^A as re-branding of the XIENCE Alpine with no design changes.		
27 February 2019	7780598	Traceable to NB 0086.		
14 October 2019	9749795	Addition of a new drug manufacturing site including minor adaptions to manufacturing process and update to testing monograph.		
20 November 2019	3092491	Change of UPLC column used in the analytical testing for lot release.		

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 6 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.





Supplementary Information to CE 632827

Issued To:

Abbott Vascular 3200 Lakeside Drive Santa Clara California 95054 USA

Certificate History

Date	Reference Number	Action			
Current	3079678	Certificate Renewal. Removal of product codes: - 1017225-08/-12/-15/-18/-23/-28 - 1017250-08/-12/-15/-18/-23/-28/-33/-38 - 1017275-08/-12/-15/-18/-23/-28/-33/-38 - 1017300-08/-12/-15/-18/-23/-28/-33/-38 - 1017350-08/-12/-15/-18/-23/-28/-33/-38 - 1017400-08/-12/-15/-18/-23/-28/-33/-38 Addition of product XIENCE PROs as re-branding of the XIENCE Sierra with no design changes. Update of the supplementary information page to include intended purpose per IFU and device classification as per current BSI template. Reformatting of device models tables. Words "and peripheral" removed from certificate scope.			

First Issued: **2015-04-13** Date: **2021-01-12** Expiry Date: **2024-05-26**

...making excellence a habit."

Page 7 of 7

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.