



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф +7495 231-2272 +7499 502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОНЫ Анти-А, Анти-В и Анти-AB)
Регистрационное удостоверение № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-В во флаконах по 10 мл с синими крышками

Серия: 095810

Единица: 100 мл

Изготовлен: 21.10.2019

Количество единиц 40

Годен до: 21.10.2021

Объем серии: 10000 мл

Паспорт: B095810 от 21.10.2019

Наименование показателя	Норма по ТУ	Результаты испытаний
1. Внешний вид		
1.1 Цоликлон анти-А	Прозрачная жидкость красного цвета.	Соответствует
1.2 Цоликлон анти-В	Прозрачная жидкость синего цвета.	
1.3 Цоликлон анти-AB	Прозрачная бесцветная жидкость.	
2. Серологические свойства		
2.1 Специфичность	Цоликлон анти-А не должен давать агглютинации с эритроцитами групп В(III) и O(I) Цоликлон анти-В не должен давать агглютинации с эритроцитами групп А(II) и O(I) Цоликлон анти-AB не должен давать агглютинации с эритроцитами группы O(I)	Соответствует Соответствует Соответствует
2.1.1 Геммагглютинирующая способность	Агглютинация на плоскости эритроцитов А1 и В с соответствующими Цоликлонами должна появиться не позднее 10 сек. после смешивания	Соответствует 10 секунд
2.3 Титр	Титр Цоликлона анти-А в реакции агглютинации на плоскости с эритроцитами группы А(II) 1:32 - 1:64 Титр Цоликлона анти-В в реакции агглютинации на плоскости с эритроцитами группы В(III)	Соответствует 1:32 - 1:64
	Титр Цоликлона анти-AB в реакции агглютинации на плоскости с эритроцитами групп А(II) 1:32 - 1:64 и В(III) 1:64	Соответствует 1:64 Соответствует 1:32 - 1:64

Цоликлон соответствует требованиям ТУ - 9398-101-51203590-2009

Заведующая ОТК ООО «Медиклон»

М.С. Орлова



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф +7495 231-2272 +7499 502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОНЫ Анти-А, Анти-В и Анти-AB)
Регистрационное удостоверение № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-А во флаконах по 10 мл с красными крышками

Серия: 096111

Единица: 100 мл

Изготовлен: 05.11.2019

Количество единиц 40

Годен до: 05.11.2021

Объем серии: 10000 мл

Паспорт: A096111 от 05.11.2019

Наименование показателя	Норма по ТУ	Результаты испытаний
1. Внешний вид		
1.1 Цоликлон анти-А	Прозрачная жидкость красного цвета.	Соответствует
1.2 Цоликлон анти-В	Прозрачная жидкость синего цвета.	
1.3 Цоликлон анти-AB	Прозрачная бесцветная жидкость.	
2. Серологические свойства		
2.1 Специфичность	Цоликлон анти-А не должен давать агглютинации с эритроцитами групп В(III) и O(I) Цоликлон анти-В не должен давать агглютинации с эритроцитами групп А(II) и O(I) Цоликлон анти-AB не должен давать агглютинации с эритроцитами группы O(I)	Соответствует Соответствует Соответствует
2.1.1 Геммагглютинирующая способность	Агглютинация на плоскости эритроцитов А1 и В с соответствующими Цоликлонами должна появиться не позднее 10 сек. после смешивания	Соответствует 10 секунд
2.3 Титр	Титр Цоликлона анти-А в реакции агглютинации на плоскости с эритроцитами группы А(II) 1:32 - 1:64 Титр Цоликлона анти-В в реакции агглютинации на плоскости с эритроцитами группы В(III)	Соответствует 1:32 - 1:64
	Титр Цоликлона анти-AB в реакции агглютинации на плоскости с эритроцитами групп А(II) 1:32 - 1:64 и В(III) 1:64	Соответствует 1:64 Соответствует 1:32 - 1:64

Цоликлон соответствует требованиям ТУ - 9398-101-51203590-2009

Заведующая ОТК ООО «Медиклон»

М.С. Орлова



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф (495) 231-2272 (499) 502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОН Анти-D Супер)
Регистрационное удостоверение: № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-D Супер во флаконах по 10 мл с зелеными крышками

Серия: 292711 Единица: 100 мл
Изготовлен: 05.11.2019 Количество единиц 40
Годен до: 05.11.2021 Объем серии: 10000 мл.
Паспорт: Дс292711 от 05.11.2019

Наименование показателя	Характеристика нормы по ТУ	Результаты испытаний
1. Внешний вид	Прозрачная жидкость светло-бежевого цвета	Соответствует
2. Серологические свойства		
2.1 Специфичность	Цоликлон Анти-D Супер не должен агглютинировать D(-) эритроциты.	Соответствует
2.2 Гемагглютинирующая способность	Четкая реакция агглютинации должна наступать в течение 30 сек. после смешивания реагента с D(+) эритроцитами	Соответствует 30 сек.
2.3 Титр	Титр Цоликлона Анти-D Супер в реакции агглютинации на плоскости с D(-) эритроцитами 1:32 Титр Цоликлона Анти-D Супер в реакции прямой агглютинации с D(+) эритроцитами в микропласте не ниже 1:256	1:32 1:256

Цоликлон соответствует требованиям ТУ-9398-101-51203590-2009

Заведующая ОТК ООО «Медиклон»

М.С. Орлова



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф +7495-231-2272 +7499-502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОНЫ Анти-A, Анти-B и Анти-AB)
Регистрационное удостоверение: № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-AB

Серия: 098611 Единица: 100 мл
Изготовлен: 05.11.2019 Количество единиц 10
Годен до: 05.11.2021 Объем серии: 10000 мл.
Паспорт: АВ098611 от 05.11.2019

Наименование показателя	Норма по ТУ	Результаты испытаний
1. Внешний вид		
1.1 Цоликлон анти-A	Прозрачная жидкость красного цвета.	Соответствует
1.2 Цоликлон анти-B	Прозрачная жидкость синего цвета.	
1.3 Цоликлон анти-AB	Прозрачная бесцветная жидкость.	
2. Серологические свойства		
2.1 Специфичность	Цоликлон анти-A не должен давать агглютинации с эритроцитами групп B(III) и O(I) Цоликлон анти-B не должен давать агглютинации с эритроцитами групп A(II) и O(I) Цоликлон анти-AB не должен давать агглютинации с эритроцитами группы O(II)	Соответствует Соответствует Соответствует
2.1 Гемагглютинирующая способность	Агглютинация на плоскости эритроцитов A и B с соответствующими Цоликлонами должна появиться не позднее 10 сек. после смешивания	Соответствует 10 секунд
2.3 Титр	Титр Цоликлона анти-A в реакции агглютинации на плоскости с эритроцитами группы A(II) 1:32 - 1:64 Титр Цоликлона анти-B в реакции агглютинации на плоскости с эритроцитами группы B(III) 1:64 Титр Цоликлона анти-AB в реакции агглютинации на плоскости с эритроцитами групп A(II) 1:32 - 1:64 и B(III) 1:64	Соответствует 1:32 - 1:64 Соответствует 1:64 Соответствует 1:32 - 1:64

Цоликлон соответствует требованиям ТУ-9398-101-51203590-2009

Заведующая ОТК ООО «Медиклон»

М.С. Орлова



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф (495) 231-2272 (499) 502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОН Анти-Kell Супер)
Регистрационное удостоверение № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-Kell Супер
Серия: 196410 Единица: 100 мл
Изготовлен: 21.10.2019 Количество единиц 10
Годен до: 21.10.2021 Объем серии: 10000 мл.
Паспорт: K196410 от 21.10.2019

Наименование показателя	Характеристика нормы по ТУ	Результаты испытаний
1. Внешний вид	Прозрачная желтоватая или розоватая жидкость	Соответствует
2. Серологические свойства		
2.1 Специфичность	Цоликлон Анти-Kell супер не должен агглютинировать эритроциты K(-)	Соответствует
2.2 Гематтотинирующая способность	Четкая реакция агглютинации на плоскости должна наступать в течение 30 сек. после смешивания	Соответствует
2.3 Активность	Титр Цоликлона Анти-Kell Супер в реакции прямой агглютинации в микровладе не ниже 1:16	Соответствует 1:16

Цоликлон соответствует требованиям ТУ - 9398-101-51203590-2009
Заведующая ОТК ООО «Медиклон»

М.С.Орлова



ООО "Медиклон"

127276 Москва, Ботаническая ул., 35, т/ф (495) 231-2272 (499) 502-1214

ПАСПОРТ-СЕРТИФИКАТ ПРОИЗВОДИТЕЛЯ
на «Набор реагентов для определения групп крови человека систем
ABO, Резус и Kell» по ТУ-9398-101-51203590-2009
(ЦОЛИКЛОН Анти-D (IgG))
Регистрационное удостоверение № ФСР 2009/06043 от 05 ноября 2009 г

Наименование: Цоликлон Анти-D(IgG)
Серия: 292110 Единица: 100 мл
Изготовлен: 28.10.2019 Количество единиц 10
Годен до: 28.10.2021 Объем серии: 10000 мл.
Паспорт: ДЖ292110 от 28.10.2019

Наименование показателя	Характеристика нормы	Результаты испытаний
1. Внешний вид	Прозрачная жидкость светло-бежевого цвета	Соответствует
2. Серологические свойства		
2.1 Специфичность	Цоликлон Анти-D не должен агглютинировать D(-) эритроциты	Соответствует
2.2 Гематтотинирующая способность	Агглютинация эритроцитов D+ с Цоликлоном в пробирочном тесте с желатином должна появляться не позднее 15 мин.	Соответствует
2.3 Титр	Титр Цоликлона анти-D в пробирочном тесте с желатином 1:128; Титр Цоликлона в непрямом антиглобулиновом тесте: 1:512	Соответствует

Цоликлон соответствует требованиям ТУ - 9398-101-51203590-2009

Заведующая ОТК ООО «Медиклон»

М.С.Орлова

Date: 21-01-20

To Whom It May Concern

Letter of Authorisation

This is to confirm that GBG-MLD SRL of str. Tighina 65, of 607, MN-2001, Chisinau, Republic of Moldova is an authorised distributor for Lorne Laboratories Limited in Moldova.

GBG-MLD SRL is authorised to present proposals, offer quotations, accept orders and participate in tender number 18/0003 for the National Blood Transfusion Center for products on behalf of Lorne Laboratories Limited. This authorisation is valid until 31.12.2021.

The undersigned herewith states that the above is true and correct.



Ian John
Managing Director

LORNE LABORATORIES LIMITED
Unit 1 Cutbush Park Industrial Estate
Danehill
Lower Earley
Berkshire RG6 4UT
United Kingdom

And duly authorised to sign this Authorisation on behalf of Lorne Laboratories Limited





CERTIFICATE OF REGISTRATION

Lorne Laboratories Ltd

Unit 1 Cutbush Park Industrial Estate
Danehill
Lower Earley
Berkshire RG6 4UT UNITED KINGDOM

UL LLC®(UL) issues this certificate to the Firm named above, after assessing the Firm's quality system and finding it in compliance with:

ISO 13485:2016

EN ISO 13485:2016

The manufacture of in vitro diagnostic blood grouping reagents. The purchase for resale of in vitro diagnostic serology test kit.

Authorized by

Michael J. Windler, P.E.

Manager of Global Regulatory Service
Distinguished Member of the Technical Staff
UL Life and Health Sciences
UL LLC



Check Certificate
Status: [here](#)

File Number A12241
Certificate 1458.180626
Initial Issue June 26, 2018

Cycle Start Date June 26, 2018
Effective Date June 26, 2018
Expiry Date May 22, 2020

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL LLC.



00-MB-S0043 Issue 15.0



UL LLC
333 Pfingsten Road
Northbrook, IL 60062-2096
USA

UL and the UL logo are trademarks of Underwriters Laboratories Inc. © 2011.



CERTIFICATO n.
CERTIFICATE No.

42654MC

SI CERTIFICA CHE IL SISTEMA DI GESTIONE PER LA QUALITÀ DI
WE HEREBY CERTIFY THAT THE QUALITY MANAGEMENT SYSTEM OPERATED BY

KIMA S.R.L.

UNITA OPERATIVE / OPERATIVE UNITS

Via Leonardo Da Vinci, 22 - Zona Industriale Tognana - 35028 Piove di Sacco (PD)
Italia

È CONFORME ALLA NORMA / IS IN COMPLIANCE WITH THE STANDARD

UNI CEI EN ISO 13485:2016

Sistema di Gestione per la Qualità / Quality Management System

PER LE SEGUENTI ATTIVITÀ / FOR THE FOLLOWING ACTIVITIES

EA: 29

Commercializzazione di prodotti del Gruppo: kit diagnostici,
terreni di coltura per microbiologia, articoli in plastica per laboratorio analisi,
provette con vuoto predeterminato e aghi sterili.

Trading of the products of the Group: diagnostic kits, culture media for microbiology,
plastic disposable labware, test tubes with predetermined vacuum and sterile needles.

Relativo alla documentazione del Sistema di Gestione per la Qualità, aziendale per l'implementazione del modello della norma di riferimento.
Relativo to the documentation of the Quality Management System for achieving compliance with the reference standard requirements.
For information purposes, a copy of the certificate is available on request at the following address: cert@kimasrl.it

For timely and updated information about any changes in the certificate status referred to in this certificate,
please contact the number +39 02 773341 or email address info@kimasrl.it

Data emissione First issue	Emissione corrente Current issue	Data di scadenza Expiry date
18/01/2007	18/01/2019	17/01/2022

ICIM S.p.A.

Via Po, 10 - 35028 Piove di Sacco (PD)
www.icim.it



ISO 9001:2015

Member of the group of companies: KIMA S.p.A. and KIMA S.p.A. S.p.A.
Signatory of the ISO 9001:2015 and ISO 13485:2016 Agreements

CISO is a member of
-IONet-

www.ionet-certification.com
www.ionet-certification.com

IONet, the association of the world's first class
certification bodies, is a leading international
organization in the field of management
systems certification. It is the only
organization to have been awarded the
award of ISO 9001:2015 and ISO 13485:2016
over 100 subsidiaries all over the globe.



www.ciso.it

CISO is the Italian Federation of Organisms of
Certification and Control of the Italian Association
of Certification Bodies.



THE INTERNATIONAL CERTIFICATION NETWORK

CERTIFICATE

CISQ/ICIM SPA has issued an IQNet recognized certificate that the organization:

KIMA S.r.l.

Via Leonardo Da Vinci, 22 - Zona Industriale Tognana - I-35028 Piove di Sacco (PD)

has implemented and maintains a

Quality Management System

for the following scope:

Trading of the products of the Group: diagnostic kits, culture media for microbiology, plastic disposable labware, test tubes with predetermined vacuum and sterile needles.

which fulfils the requirements of the following standard:

ISO 13485:2016

Issued on: **2019-01-18**

First issued on: **2007-01-18**

Expires on: **2022-01-17**

This attestation is directly linked to the IQNet Partner's original certificate and shall not be used as a stand-alone document.

Registration Number: IT-70247



Alex Stoichitoiu
President of IQNET



Ing. Claudio Provetti
President of CISQ

IQNet Partners*:

AENOR Spain AFNOR Certification France APCER Portugal CCC Cyprus CISQ Italy
CQC China CQM China CQS Czech Republic Cro Cert Croatia DQS Holding GmbH Germany FCAV Brazil
FONDONORMA Venezuela ICONTEC Colombia Inspecta Sertifiointi Oy Finland INTECO Costa Rica
IRAM Argentina JQA Japan KFQ Korea MIRTEC Greece MSZT Hungary Nemko AS Norway NSAI Ireland
NYCE-SIGE México PCBC Poland Quality Austria Austria RR Russia SII Israel SIQ Slovenia
SIRIM QAS International Malaysia SQS Switzerland SRAC Romania TEST St Petersburg Russia TSE Turkey YUQS Serbia
IQNet is represented in the USA by: AFNOR Certification, CISQ, DQS Holding GmbH and NSAI Inc.



Anticoagulanti - Blood collection tubes

SODIO CITRATO - SODIUM CITRATE



18610
NON STERILE

0,25 ml di sodio citrato in provetta in polipropilene.

0,25 ml of sodium citrate in polypropylene test tube.

Ø 12x56 mm.



18613
NON STERILE

0,25 ml di sodio citrato in provetta polipropilene.

0,25 ml of sodium citrate in polypropylene test tube.

Ø 16x58 mm.



18614
NON STERILE

0,25 ml di sodio citrato in provetta polipropilene.

0,25 ml of sodium citrate in polypropylene test tube.

Ø 12x86 mm.



18608
NON STERILE

0,3 ml di sodio citrato in provetta polipropilene. Graduata a 3 ml.

0,3 ml of sodium citrate in polypropylene test tube. Graduated to 3 ml.

Ø 13x75 mm.

SODIO CITRATO - SODIUM CITRATE



18688
NON STERILE

0,3 ml di sodio citrato in provetta in polipropilene. Graduata a 3 ml.

0,3 ml of sodium citrate in polypropylene test tube. Graduated to 3 ml.

Ø 13x75 mm.



18561
NON STERILE

0,5 ml di sodio citrato in provetta in polipropilene.

0,5 ml of sodium citrate in polypropylene test tube.

Ø 16x58 mm.



18562
NON STERILE

0,5 ml di sodio citrato in provetta in polipropilene.

0,5 ml of sodium citrate in polypropylene test tube.

Ø 12x86 mm.



18564
NON STERILE

0,5 ml di sodio citrato in provetta polipropilene.

0,5 ml of sodium citrate in polypropylene test tube.

Ø 16x100 mm.



Provette monouso

Disposable test tubes



18016
NON STERILE

Provetta cilindrica in polipropilene.
Opaca.
Graduata a 2,5-5 ml.

Polypropylene cylindrical test tube.
Translucid.
Graduated to 2,5-5 ml.

Vol. 5 ml.
Ø 12x86 mm.



18018
NON STERILE

Provetta cilindrica in polipropilene.
Con etichetta. Opaca.
Graduata a 2,5-5 ml.

Polypropylene cylindrical test tube.
With label. Translucid.
Graduated to 2,5-5 ml.

Vol. 5 ml.
Ø 12x86 mm.



18020
NON STERILE

Provetta cilindrica in polistirolo.
Trasparente.
Graduata a 2,5-5 ml.

Polystyrene cylindrical test tube.
Transparent.
Graduated to 2,5-5 ml.

Vol. 5 ml.
Ø 12x86 mm.



18021
NON STERILE

Provetta cilindrica in polistirolo.
Con etichetta. Trasparente.
Graduata a 2,5-5 ml.

Polystyrene cylindrical test tube.
With label.
Transparent. Graduated to 2,5-5 ml.

Vol. 5 ml.
Ø 12x86 mm.



18012
NON STERILE

Provetta conica in polipropilene.
Opaca.
Graduata a 0,5-1-2,5-5-10 ml.

Polypropylene conical test tube.
Translucid.
Graduated to 0,5-1-2,5-5-10 ml.

Vol. 10 ml.
Ø 16x100 mm.



18014
NON STERILE

Provetta conica in polipropilene.
Con etichetta. Opaca.
Graduata a 0,5-1-2,5-5-10 ml.

Polypropylene conical test tube.
With label. Translucid.
Graduated to 0,5-1-2,5-5-10 ml.

Vol. 10 ml.
Ø 16x100 mm.



18304
NON STERILE

Provetta conica in polistirolo.
Trasparente.
Graduata a 0,5-1-2,5-5-10 ml.

Polystyrene conical test tube.
Transparent.
Graduated to 0,5-1-2,5-5-10 ml.

Vol. 10 ml.
Ø 16x100 mm.




18306
NON STERILE

Provetta conica in polistirolo.
Con etichetta. Trasparente.
Graduata a 0,5-1-2,5-5-10 ml.

Polystyrene conical test tube.
With label. Transparent.
Graduated to 0,5-1-2,5-5-10 ml.

Vol. 10 ml.
Ø 16x100 mm.



EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
ASO Latex kit	031100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.



Eddy Velthuis
Technical Director



File No A12241;
ISO 13485:2003; ISO 9001:2008

Lorne Laboratories Limited
Unit 1 Cutbush Park Industrial Estate
Danehill, Lower Earley
Berkshire RG6 4UT United Kingdom

Tel: +44 (0) 118 921 2264
Fax: +44 (0) 118 986 4518
Email: info@lornelabs.com
www.lornelabs.com

Registered office as above, Registered in England No. 04540797, VAT No. 800 3655 66



EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
RF Latex kit	830100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.



Eddy Velthuis
Technical Director



File No A12241;
ISO 13485:2003; ISO 9001:2008

Lorne Laboratories Limited
Unit 1 Cutbush Park Industrial Estate
Danehill, Lower Earley
Berkshire RG6 4UT United Kingdom

Tel: +44 (0) 118 921 2264
Fax: +44 (0) 118 986 4518
Email: info@lornelabs.com
www.lornelabs.com

Registered office as above. Registered in England No. 04540797. VAT No. 800 3655 66





EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
CRP Latex kit	850100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.

Eddy Velthuis
Technical Director



File No A12241;
ISO 13485:2003; ISO 9001:2008

Lorne Laboratories Limited
Unit 1 Cutbush Park Industrial Estate
Danehill, Lower Earley
Berkshire RG6 4UT United Kingdom
Tel: +44 (0) 118 921 2264
Fax: +44 (0) 118 986 4518
Email: info@lornelabs.com
www.lornelabs.com

Registered office as above. Registered in England No. 04540797, VAT No. 800 3655 66

Declaration of Conformity

helena
Biosciences Europe

HL-7-0163DC DOI 2015/08 (9)

In Application of the Council Directive 98/79/EC on the approximation of the laws of the Member States relating to *In Vitro* Diagnostic Medical Devices & CE marking.

Declaration of conformance to applicable sections of Annex I - Essential Requirements and Annex III (EC Declaration of Conformity) imposed by sections 2 to 5. The below listed products are not classified under Annex II Lists A or B. Access to the appropriate technical files will be made available to the appropriate body in the event this is required.

Product Code	Description	GMDN Classification Code
5265	Thromboplastin LI	55983
5265H	Thromboplastin LI	55983

I, the undersigned declare that the devices registered against the above GMDN Classification Code conforms to the said Directives.

Full Name: M.J. Stephenson

Title: Managing Director

Signed:



Date: 06 Aug 2015

Tel +44 (0)191 482 8440
Fax +44 (0)191 482 8442
info@helena-biosciences.com
www.helena-biosciences.com

Helena Biosciences Europe
Queensway South, Team Valley Trading Estate,
Gateshead, Tyne and Wear, NE11 0SD,
United Kingdom





120 ML URINE CONTAINERS FOR VACUUM TUBES CONTENITORI URINE 120 ML PER PROVETTE SOTTOVUOTO

Urine containers graduated up to 100 ml, yellow screw cap with device for suction with vacuum tubes. Label on the cap for sampling system protection. With writing surface.

Vol. 120 ml.
Dim. Ø 59 x 68 mm.
Material: polypropylene.

Contentori per urine graduati fino a 100 ml, tappo a vite giallo con dispositivo per aspirazione con provette sottovuoto. Con etichetta applicata sul tappo per la protezione del sistema di prelievo. Con superficie di scrittura.

Vol. 120 ml.
Dim. Ø 59 x 68 mm.
Materiale: polipropilene.

COD.	SUPPLIED / CONFEZIONE
5120	Bulk / Sfuso
5120/TS	Sterile - Bulk / Sfuso
5120/SG	Sterile - Individually wrapped / Confezione singola



60 ML URINE CONTAINERS FOR VACUUM TUBES CONTENITORI URINA 60 ML PER PROVETTE SOTTOVUOTO

Urine containers graduated up to 45 ml, yellow screw cap with device for suction with vacuum tubes. Label on the cap for sampling system protection.

Vol. 60 ml.
Dim. Ø 35 x 70 mm.
Material: polypropylene.

Contentori per urine graduati fino a 45 ml, tappo a vite giallo con dispositivo per aspirazione con provette sottovuoto. Con etichetta applicata sul tappo per la protezione del sistema di prelievo.

Vol. 60 ml.
Dim. Ø 35 x 70 mm.
Materiale: polipropilene.

COD.	SUPPLIED / CONFEZIONE
5060	Bulk / Sfuso
5060/TS	Sterile - Bulk / Sfuso
5060/SG	Sterile - Individually wrapped / Confezione singola



150 ML URINE CONTAINERS CONTENITORI URINE DA 150 ML

Urine containers graduated up to 100 ml, with screw cap and writing surface.

Vol. 150 ml.
Dim. Ø 58 x 72 mm.
Material: polypropylene.

Contentori per urine graduati fino a 100 ml, con tappo a vite e superficie di scrittura.

Vol. 150 ml.
Dim. Ø 58 x 72 mm.
Materiale: polipropilene.

COD.	SUPPLIED / CONFEZIONE	CAP COLOUR / COLORE TAPPO
2120	Bulk, with separated cap / Sfuso, con tappo a parte	Light blue / Azzurro
2120/T	Bulk, with inserted cap / Sfuso, con tappo inserito	Light blue / Azzurro
2120/TS	Sterile, with inserted cap / Sterile, con tappo inserito	Red / Rosso
2120/CS	Aseptic / Asettico - Individually wrapped / Confezione singola	Red / Rosso
2120/SG	Sterile - Individually wrapped / Confezione singola	Red / Rosso



STRAWS WITH URINE COLLECTION DEVICE SONDE CON DISPOSITIVO DI PRELIEVO DELLE URINE

Straws with urine collection device for vacuum test tubes. Suitable for containers without collection device. Available with straw of two sizes.

Sonde per il prelievo delle urine con provette sottovuoto da contenitori non dotati di dispositivo. Disponibili con tubicino di due lunghezze.

COD.	STRAW LENGTH / LUNGHEZZA SONDA	SUPPLIED / CONFEZIONE
31400	90 mm	Non sterile
31400/SG	90 mm	Sterile - Bags of 100 pcs / Sacchetti da 100 pz.
31400/SG/CS	90 mm	Sterile - Individually wrapped / Confezione singola
31500	170 mm	Non sterile





Certification
Médical-Santé



CERTIFICAT
CERTIFICATE OF REGISTRATION
N° 10462 rev. 6

Le LNE certifie que le système de management de la qualité développé par
LNE certifies that the quality management system developed by

ELITECH CLINICAL SYSTEMS SAS
Zone Industrielle
61500 SEES FRANCE

pour les activités
for the activities

Conception, production, contrôle et commercialisation de réactifs de chimie clinique
destinés aux laboratoires d'analyses de biologie.
Validation de la combinaison réactifs et automates.

Design, production, control and sales of clinical chemistry reagents
intended to be used by clinical laboratories.
Validation of the combination reagents and instruments.

réalisées sur le(s) site(s) de
performed on the location(s) of

ELITech Clinical Systems SAS
Zone industrielle - 61500 - SEES - FRA

est conforme aux exigences des normes internationales
complies with the requirements of the international standards

NF EN ISO 13485 : 2016

Début de validité / Effective date : July 28th, 2017 (included)

Valable jusqu'au / Expiry date : July 27th, 2020 (included)

Etabli le / Issued on : July 27th, 2017

cofrac



**CERTIFICATION
DE SYSTEMES
DE MANAGEMENT**
Accréditation n°4-0038
Liste des sites accrédités
et portée disponible sur
www.cofrac.fr

LNE N° 10462-6

Ce certificat est délivré selon les règles de certification G-MED / This certificate is issued according to the rules of G-MED certification

Renouvelle le certificat 10462-5

On behalf of the Certification Director
Cécile VAUGELADE
G-MED Certification Division Manager



Laboratoire national de métrologie et d'essais • Établissement public à caractère industriel et commercial
LNE/G-MED • Organisme notifié par l'autorité compétente française en matière de dispositifs médicaux n° 0459
1, rue Gaston Boissier - 75724 Paris Cedex 15 • Tél. : 01 40 43 37 00 • Fax : 01 40 43 37 37 • www.lne.fr • www.gmed.fr



DECLARATION DE CONFORMITE CE

Nous, ELITech Clinical Systems SAS, zone Industrielle 61500 SEES France, déclarons sous notre seule responsabilité que les réactifs appartenant au groupe 1 «METABOLITES DIVERS », référencés dans la liste ci-jointe, sont conformes aux exigences essentielles des annexes I et III de la Directive Européenne 98/79/CE relative aux dispositifs médicaux de diagnostic *in vitro* et au code de la santé publique.

Cette déclaration est basée sur le contenu de chaque dossier technique DOS-CE-XXXX et s'appuie sur la certification de notre système qualité selon la norme NF EN ISO 13485 : 2015 (Certification valable jusqu'au 27 juillet 2020).

(Voir liste ci-jointe).

DECLARATION OF EC CONFORMITY

We, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, hereby certify, under our own responsibility, that the reagents belonging to Group 1 "MISCELLANEOUS METABOLITES", such as listed hereto, conform to the essential requirements of appendices I and III of European Directive 98/79/EC, relating to *in vitro* diagnostic medical devices and to the public health code.

This declaration is based on the contents of each DOS-CE-XXXX technical file and is supported by the certification of our quality system according to the standard NF EN ISO 13485 : 2015 (Certification valid until July 27th, 2020).

(See attached list).

DECLARACIÓN CE DE CONFORMIDAD

Nosotros, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, declaramos bajo nuestra única responsabilidad que los reactivos pertenecientes al grupo 1 "METABOLITOS VARIOS", referenciados en la lista adjunta, son conformes con los requisitos esenciales de los anexos I y III de la Directiva Europea 98/79/CE sobre dispositivos médicos para diagnóstico *in vitro* y el código de salud pública.

Esta declaración se basa en el contenido de cada DOS-CE-XXXX técnica y está respaldado por la certificación de nuestro sistema de calidad según la norma NF EN ISO 13485 : 2015 (Certificación válida hasta el 27 de Julio 2020).

(Ver lista adjunta)

Sées, le 18 Mai 2018

Valérie GOURDON,

Responsable des Affaires Réglementaires
Regulatory Affairs Manager
Responsable de los Asuntos Reglamentarios

ELITech Clinical Systems SAS

Zone Industrielle
61500 SEES - France
Tél. : +33(0)2 33 81 21 00 - Fax : +33(0)2 22 28 77 51
SIRET 318 365 228 00056

Cécile GOUBAULT,
Directeur Général Délégué
Managing Director
Directora General

Tél. : +33(0)2 33 81 21 00 - Fax : +33(0)2 22 28 77 51
SIRET 318 365 228 00056



DESIGNATION DU REACTIF/ REAGENT DESIGNATION/ DESIGNACIÓN DE REACTIVO	REFERENCES/ REFERENCIAS	NOM DU DOSSIER CE/ EC FILE NAME/ NOMBRE DEL ARCHIVO CE	Code GMDN/ GMDN Code/ Codigo GMDN
ALBUMIN	ALBU-0600/0700/0250	DOS-CE-ALBU_R&Sd	53397
ALBUMIN ENVOY	ALBU-0850		
BILIRUBIN DIRECT 4+1	BIDI-0600/0250		53233
BILIRUBIN TOTAL 4+1	BITO-0600/0250	DOS-CE-BILI 4+1	53229
CREATININE ENVOY	CRSL-0850	DOS-CE-CRSL	53229/53233
CREATININE JAFTE	CRCO-0600/0700	DOS-CE-CRSL	53251
CREATININE PAP SL	CRSL-0630/0250	DOS-CE-CRSL	53250
DIRECT BILIRUBIN ENVOY	BIDV-0850	DOS-CE-BIDV	53233
GLUCOSE ENVOY	GFSL-0850	DOS-CE-GFSL_R&Sd	
GLUCOSE HK SL	GHSI-0600/0250	DOS-CE-GHSI	53301
GLUCOSE PAP SL	GFSL-0495/0500/0700/ 0507/0707/0250/0455/0497	DOS-CE-GFSL_R&Sd	
HEMOGLOBIN	HEMO-0400	DOS-CE-HEMO	32430
IRON TIBC	FECA-0050	DOS-CE-TIBC	53904
LACTATE	LACT-0100	DOS-CE-LACT	53342
MICROPROTEIN PLUS	PRTU-0600/0250	DOS-CE-PRTU_R&Sd	53481
PHOSPHORUS	PHOS-0600/0230	DOS-CE-PHOS_R&Sd	59123
PHOSPHORUS ENVOY	PHOS-0850		
TOTAL BILIRUBIN ENVOY	BITV-0850	DOS-CE-BITV	53229
TOTAL PROTEIN	PRTB-0600	DOS-CE-PRTB	
TOTAL PROTEIN ENVOY	PROB-0850		
TOTAL PROTEIN PLUS	PROB-0600/0700/0250	DOS-CE-PROB_R&Sd	53985
UREA ENVOY	URSL-0850	DOS-CE-URSL_R&Sd	
UREA UV	URUV-0400	DOS-CE-URUV	53587
UREA UV SL	URSL-0400/0420/0500/ 0407/0427/0507/0250/0455	DOS-CE-URSL_R&Sd	
URIC ACID	ACUR-0200/0400	DOS-CE-ACUR	
URIC ACID ENVOY	AUVD-0850	DOS-CE-AUVD	
URIC ACID MONO SL	AUML-0420/0500/0700/ 0427/0497/0507/0707/0250	DOS-CE-AUML_R&Sd	53583
URIC ACID SL	AUSL-0250	DOS-CE-AUSL	

ELITech Clinical Systems SAS

Zone Industrielle
61500 SEES - France
Tél. : +33(0)2 33 81 21 00 - Fax : +33(0)2 22 28 77 51
SIRET 318 365 228 00056

DECLARATION DE CONFORMITE CE

Nous, ELITech Clinical Systems SAS, zone industrielle 61500 SEES France, déclarons sous notre seule responsabilité que les réactifs appartenant au groupe 4 "LIPIDES", référencés dans la liste ci-jointe, sont conformes aux exigences essentielles des annexes I et III de la Directive Européenne 98/79/CE relative aux dispositifs médicaux de diagnostic *in vitro* et au code de la santé publique.

Cette déclaration est basée sur le contenu de chaque dossier technique DOS-CE-XXXX et s'appuie sur la certification de notre système qualité selon la norme NF EN ISO 13485 : 2016 (Certification valable jusqu'au 27 juillet 2020).

(Voir liste ci-jointe).

DECLARATION OF EC CONFORMITY

We, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, hereby certify, under our own responsibility, that the reagents belonging to Group 4 "LIPIDS", such as listed hereto, conform to the essential requirements of appendices I and III of European Directive 98/79/EC, relating to *in vitro* diagnostic medical devices and to the public health code.

This declaration is based on the contents of each DOS-CE-XXXX technical file and is supported by the certification of our quality system according to the standard NF EN ISO 13485 : 2016 (Certification valid until July 27th, 2020).

(See attached list).

DECLARACIÓN CE DE CONFORMIDAD

Nosotros, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, declaramos bajo nuestra única responsabilidad que los reactivos pertenecientes al grupo 4 "LIPIDOS", referenciados en la lista adjunta, son conformes con los requisitos esenciales de los anexos I y III de la Directiva Europea 98/79/CE sobre dispositivos médicos para diagnóstico *in vitro* y el código de salud pública.

Esta declaración se basa en el contenido de cada DOS-CE-XXXX técnica y está respaldado por la certificación de nuestro sistema de calidad según la norma NF EN ISO 13485 : 2016 (Certificación válida hasta el 27 de Julio 2020).

(Ver lista adjunta)

Sées, le 28 Juillet 2017

Valérie GOURDON
Responsable des Affaires Réglementaires
Regulatory Affairs Manager
Responsable de los Asuntos Reglamentarios

Cécile GOUBAULT,
Directeur Général Délégué
Managing Director
Directora General

ELITech Clinical Systems SAS
Zone Industrielle
61500 SEES - France
Tel : +33(0)2 33 81 21 00 - Fax : +33(0)2 22 28 77 51
N°SIRET 315 092 228 0006

ELITech Clinical Systems SAS
Zone Industrielle
61500 SEES - France
Tel : +33(0)2 33 81 21 00 - Fax : +33(0)2 22 28 77 51
N°SIRET 315 092 228 0006

DECLARATION DE CONFORMITE CE

Nous, ELITech Clinical Systems SAS, zone Industrielle 61500 SEES France, déclarons sous notre seule responsabilité que les réactifs appartenant au groupe 5 «CONTROLES/ CALIBRATORS/ STANDARDS » référencés dans la liste ci-jointe, sont conformes aux exigences essentielles des annexes I et III de la Directive Européenne 98/79/CE relative aux dispositifs médicaux de diagnostic in vitro et au code de la santé publique.

Cette déclaration est basée sur le contenu de chaque dossier technique DOS-CE-XXXX et s'appuie sur la certification de notre système qualité selon la norme NF EN ISO 13485 : 2016 (Certification valable jusqu'au 27 juillet 2020).

(Voir liste ci-jointe).

DECLARATION OF EC CONFORMITY

We, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, hereby certify, under our own responsibility, that the reagents belonging to Group 5, "CONTROLS/ CALIBRATORS/ STANDARDS", such as listed hereto, conform to the essential requirements of appendices I and III of European Directive 98/79/EC, relating to in vitro diagnostic medical devices and to the public health code.

This declaration is based on the contents of each DOS-CE-XXXX technical file and is supported by the certification of our quality system according to the standard NF EN ISO 13485 : 2016 (Certification valid until July 27th, 2020).

(See attached list).

DECLARACIÓN CE DE CONFORMIDAD

Nosotros, ELITech Clinical Systems SAS, Zone Industrielle 61500 SEES France, declaramos bajo nuestra única responsabilidad que los reactivos pertenecientes al grupo 5 "CONTROLES/ CALIBRADORES/ ESTÁNDARES", referenciados en la lista adjunta, son conformes con los requisitos esenciales de los anexos I y III de la Directiva Europea 98/79/CE sobre dispositivos médicos para diagnóstico in vitro y el código de salud pública.

Esta declaración se basa en el contenido de cada DOS-CE-XXXX técnica y está respaldado por la certificación de nuestro sistema de calidad según la norma NF EN ISO 13485 : 2016 (Certificación válida hasta el 27 de Julio 2020).

(Ver lista adjunta)

Sées, le 28 Juillet 2017

Valérie GOURDON,
Responsable des Affaires Réglementaires
Regulatory Affairs Manager
Responsable de los Asuntos Reglamentarios

Cécile GOUBAULT,
Directeur Général Délégué
Managing Director
Directora General

ELITech Clinical Systems SAS
Zone Industrielle
61500 SEES - France
Tel : +33(0)2 33 81 21 00 Fax : +33(0)2 22 28 77 51

ELITech Clinical Systems SAS
Zone Industrielle
61500 SEES - France
Tel : +33(0)2 33 81 21 00 Fax : +33(0)2 22 28 77 51

GRUPE 5 – CONTROLES/CALIBRANTS/STANDARDS GROUP 5 – CONTROLS/CALIBRATORS/STANDARDS GRUPO 5 – CONTROLES/CALIBRADORES/ESTÁNDARES

DESIGNATION DU REACTIF/ REAGENT DESIGNATION/ DESIGNACIÓN DE REACTIVO	REFERENCES/ REFERENCIAS	NOM DU DOSSIER CE/ EC FILE NAME/ NOMBRE DEL ARCHIVO CE	Code GMDN/ GMDN Code/ Codigo GMDN
CHOLESTEROL HDL 2G CALIBRATOR	HDL-0011/0041	DOS-CE-HDLL-CAL	44696
CHOLESTEROL LDL 2G CALIBRATOR	LDL-0011/0041	DOS-CE-LDLL-CAL	41728
CHOLESTEROL Standard 200 mg/dL	CHOL-0055	DOS-CE-CHOL200	44698
CK-MB CONTROL	CKMB-0900	DOS-CE-CKMB-CT	44693
CREATININE Standard 2 mg/dL	CREN-0055	DOS-CE-CREN2	44700
ELICAL 2	CALI-0550	DOS-CE-CALI2	47868
ELITROL I	CONT-0060	DOS-CE-ELIT I	47869
ELITROL II	CONT-0160	DOS-CE-ELIT II	41818
GLUCOSE Standard 100 mg/dL	GLUP-0055	DOS-CE-GLUP100	47869
ISE CONTROL I	ISCT-0046	DOS-CE-ISCT	53482
ISE CONTROL II	ISCT-0047	DOS-CE-TRU200	44702
MICROPROTEIN PLUS Standard 100 mg/dL	PRTU-0022	DOS-CE-TRU200	53588
TRIGLYCERIDES Standard 200 mg/dL	TRIG-0055	DOS-CE-URUV50	44704
UREA Standard 50 mg/dL	URUV-0055	DOS-CE-ACUR6	
URIC ACID Standard 6 mg/dL	ACUR-0055		

DECLARATION OF CONFORMITY

1) Manufacturer (Name, department): **Monobind Inc.**
 Address: 100 North Pointe, LAKE FOREST, CA 92630. UNITED STATES

2) European authorized representative: **CEpartner4U BV**,
 Address: **ESDOORNLAAN 13, 3951DB MAARN, THE NETHERLANDS**;
 (on product labels printed as:
 CEpartner4U, ESDOORNLAAN 13, 3951DB MAARN, THE NETHERLANDS Tel: +31 (0)6 516 536 26;
 or as: CEpartner4U, 3951DB; 13, NL tel: +31 (0)6 - 516.536.26)

3) Product(s) (name, type or model/batch number, etc.):

Immunoassay products;

ELISA,

CLIA,

Control,

Instruments

(see appendix)

4) The product(s) described above is in conformity with:

Document No.	Title	Edition / Date of issue
L 331; 98/79/EC	In-Vitro-Diagnostic Directive	1998-10-27

5) Additional information (conformity procedure, Notified Body, CE certificate, etc.):

Conformity assessment procedure for CE marking: IVD Directive, Annex III

Lake Forest, USA; 2011-09-27

A Shatola

Tony Shatola; QA Director, Monobind Inc.

(Place & date of issue (yyyy-mm-dd)) (name, function and signature of manufacturer)

Maarn, NL; 2011-09-27

Olga Terlinck; Consultant, CEpartner4U BV

(Place & date of issue (yyyy-mm-dd)) (name, function and signature of authorized representative)

Appendix

Date: 2011-09-26

Device types	Item# ELISA	Item# CLIA	Item# Control	Item# Instrument	EDMS code	Risk Class	Certificate #	First date of CE marking
Thyroid								
T3 - Triiodothyronine	125-300	175-300			12.04.01.05.00	Low		2005-11-11
FT3 - Free Triiodothyronine	1325-300	1375-300			12.04.01.01.00	Low		2005-11-11
T4 - Thyroxine	225-300	275-300			12.04.01.07.00	Low		2005-11-11
FT4 - Free Thyroxine	1225-300	1275-300			12.04.01.02.00	Low		2005-11-11
TSH - Thyrotropin	325-300	375-300			12.04.01.11.00	Low		2005-11-11
Rapid TSH - Rapid Thyrotropin	6025-300	6075-300			12.04.01.11.00	Low		2010-06-29
T3U - Triiodothyronine Urinate	525-300	575-300			12.04.01.06.00	Low		2005-11-11
TBG - Thyroxine-Binding Globulin	3525-300	3575-300			12.04.01.09.00	Low		2005-11-11
Tg - Thyroglobulin	2225-300	2275-300			12.04.01.08.00	Low		2005-11-11
T3, T4 & TSH - Triiodothyronine, Thyroxine & Thyrotropin Combo (VAST)	8025-300	8075-300			12.04.01.01.00	Low		2005-11-11
T3 - Triiodothyronine (SBS)	8125-300	8175-300			12.04.01.01.00	Low		2010-06-29
T4 - Thyroxine (SBS)	8225-300	8275-300			12.04.01.01.00	Low		2010-06-29
T3, FT4 & TSH - Free Triiodothyronine, Free Thyroxine & Thyrotropin Combo (VAST)	7025-300	7075-300			12.04.01.01.00	Low		2010-06-29
Neonatal Thyroid & Genetics								
NTSH - Neonatal Thyrotropin	3425-300	3475-300			12.04.01.90.00	Low		2005-11-11
NT4 - Neonatal Thyroxine	2625-300	2675-300			12.04.01.12.00	Low		2005-11-11
N17OHP - Neonatal 17 OH Progesterone	5525-300				12.05.01.07	Low		2008-02-01
Biotinidase	8825-300				12.07.02.90.00	Low		2011-09-26
Autoimmune Thyroid								
Anti-Tg - Anti-Thyroglobulin Antigen	1025-300	1075-300			12.10.03.04.00	Low		2005-11-11
Anti-TPO - Anti-Thyroperoxidase Antigen	1125-300	1175-300			12.10.03.01.00	Low		2005-11-11
Fertility & Prenatal								
LH - Luteotropin	625-300	675-300			12.05.01.05.00	Low		2005-11-11
FSH - Folitropin	425-300	475-300			12.05.01.04.00	Low		2005-11-11
PRL - Prolactin	725-300	775-300			12.05.01.08.00	Low		2005-11-11
PRL - Prolactin Sequential	6025-300	6075-300			12.05.01.06.00	Low		2005-11-11
hCG - Human Chorionic Gonadotropin	825-300	875-300			12.05.02.05.00	Low		2005-11-11
Rapid hCG - Rapid Human Chorionic Gonadotropin	3325-300				12.05.02.05.00	Low		2005-11-11
FSH, LH, hCG, sPRL Combo (VAST)	8325-300	8375-300			12.05.01.90.00	Low		2006-08-24
AFP, hCG, uES Combo (VAST)	8525-300	8575-300			12.05.01.90.00	Low		2010-06-29
Steroid								
Cortisol	3625-300	3675-300			12.06.02.04.00	Low		2005-11-11
DHEA-S - Dehydroepiandrosterone sulfate	5125-300	5175-300			12.05.01.02.00	Low		2010-06-29
DHEA - Dehydroepiandrosterone	7425-300	7475-300			12.05.01.02.00	Low		2011-09-26

Device types	Item# ELISA	Item# CLIA	Item# Control	Item# Instrument	EDMS code	Risk Class	Certificate # CE-marking	First date of CE-marking
E2 - Estradiol	4825-300	4875-300			12.05.01.03.00	Low		2010-06-29
uE3 - Estradiol, Unconjugated	5025-300	5075-300			12.05.02.02.00	Low		2010-06-29
Progesterone	4825-300	4875-300			12.05.01.06.00	Low		2010-06-29
Testosterone	3725-300	3775-300			12.05.01.10.00	Low		2007-11-01
Free Testosterone	5325-300	5375-300			12.05.01.10.00	Low		2010-06-29
17OH-P - 17-Hydroxyprogesterone	5225-300	5275-300			12.05.01.07.00	Low		2010-06-29
17OHP - 17-Hydroxyprogesterone Exl. Range	9825-300	9875-300			12.05.01.07.00	Low		2010-10-18
Vitamin D3 - 25-Hydroxyvitamin D3	7725-300	7775-300			12.06.03.10.00	Low		2011-09-26
Growth & Bone Metabolism								
hGH - Human Growth Hormone	1725-300	1775-300			12.06.04.02.00	Low		2005-11-11
PTH - Parathyroid Hormone	7825-300	7875-300			12.06.03.13.00	Low		2011-09-26
Diabetes								
Insulin	2425-300	2475-300			12.06.01.03.00	Low		2005-11-11
Insulin Rapid	5825-300				12.06.01.03.00	Low		2010-06-29
C-peptide	2725-300	2775-300			12.06.01.01.00	Low		2005-11-11
Insulin & C-peptide Combo (VAST)	7325-300	7375-300			12.06.01.03.00	Low		2005-11-11
Cardiac Markers								
CKMB - Circulating Creatine Kinase (MB)	2925-300	2975-300			12.13.01.02.00	Low		2005-11-11
CTnI - Troponin I	3825-300	3875-300			12.13.01.07.00	Low		2005-11-11
DIG - Digoxin	925-300	975-300			12.08.01.01.00	Low		2005-11-11
HS-CRP - High Sensitivity C-Reactive Protein	3125-300	3175-300			12.13.01.90.00	Low		2005-11-11
Myoglobin	3225-300	3275-300			12.13.01.05.00	Low		2005-11-11
Infectious Diseases								
IgG - Anti/H1 Pylori	1425-300	1475-300			15.01.04.03.00	Low		2005-11-11
IgM - Anti/H1 Pylori	1525-300	1575-300			15.01.04.03.00	Low		2005-11-11
IgA - Anti/H1 Pylori	1625-300	1675-300			15.01.04.03.00	Low		2005-11-11
Cancer Markers								
AFP - Alpha-Fetoprotein	1925-300	1975-300			12.03.90.01.00	Low		2005-11-11
CA 125 Ovarian Cancer Antigen	3025-300	3075-300			12.03.01.06.00	Low		2005-11-11
CA 15-3 Breast Cancer Antigen	5625-300	5675-300			12.03.01.02.00	Low		2010-06-29
CA 19-9 - Pancreatic Cancer Antigen	3925-300	3975-300			12.03.01.03.00	Low		2005-11-11
CEA - Carcinoembryonic Antigen	1825-300	1875-300			12.03.01.31.00	Low		2005-11-11
CEA - Carcinoembryonic Antigen Next Generation	4825-300	4875-300			12.03.01.31.00	Low		2010-06-29
hHCG - Free Beta Human Chorionic Gonadotropin	2025-300	2075-300			12.03.01.90.00	Low		2005-11-11
Allergy & Anemia								
Ferritin	2825-300	2875-300			12.07.01.02.00	Low		2005-11-11
Folate	7925-300	7975-300			12.07.01.03.00	Low		2010-06-29
IgE - Immunoglobulin E	2925-300	2975-300			12.02.01.02.00	Low		2005-11-11
hTfR - Transferrin Soluble Receptor	8625-300	8675-300			12.07.01.06.00	Low		2010-06-29
Vitamin B12	7825-300	7875-300			12.07.02.04.00	Low		2011-09-26

Miscellaneous Controls								
Anti-Tg & Anti-TPO - Positive & Negative - Anti-Thyroglobulin, Anti-Thyroperoxidase				AIT-101		12.50.01.16.00	Low	2010-06-29
High Level Fertility Control - Single Level - Progesterone, Estradiol, Human Chorionic Gonadotropin				FC-300		12.50.01.16.00	Low	2010-06-29
Maternal Control - Tri Level - Human Chorionic Gonadotropin, Free Beta Human Chorionic Gonadotropin Subunit, Alpha Feta Protein, Estradiol				MC-300		12.50.01.16.00	Low	2010-06-29
Thyroglobulin Control - Tri Level H, Pylori IgG Control - Positive & Negative				TG-300 HPY- IG-300		12.50.01.16.00 12.50.01.16.00	Low Low	2010-06-29 2010-06-29
Miscellaneous Instruments								
IC hardware + dedicated accessories + software - Autoplex ELISA Analyzer & CLIA Processor					IN006	21.02.10.01	Low	2010-06-29
IC hardware + dedicated accessories + software - Lumax Chemiluminescence Strip Reader					IN001	21.02.10.01	Low	2006-06-24
IC hardware + dedicated accessories + software - Neo-Lumax Chemiluminescence Strip Reader					IN010	21.02.10.01	Low	2011-09-26
IC hardware + dedicated accessories + software - Impulse 2 Chemiluminescence Strip Reader					IN005	21.02.10.01	Low	2006-08-24
IC hardware + dedicated accessories + software - Impulse 3 Chemiluminescence Strip Reader					IN007	21.02.10.01	Low	2010-06-29
IC hardware + dedicated accessories + software - Lumax 3.0 ELISA Reader					IN004	21.02.10.01	Low	2007-03-01
IC hardware + dedicated accessories + software - Lumax 3.0 ELISA Reader					IN008	21.02.10.01	Low	2011-09-26
IC hardware + dedicated accessories + software - Elix 3.0 ELISA Strip Reader					IN003	21.02.10.01	Low	2007-08-10
IC hardware + dedicated accessories + software - Neo-Elix ELISA Strip Reader					IN009	21.02.10.01	Low	2011-09-26
IC hardware + dedicated accessories + software - Microplate Washer					IN002	21.02.10.01	Low	2010-06-29





Thyrotropin (TSH) Test System

Product Code: 325-300

1.0 INTRODUCTION

Intended Use: The Quantitative Determination of Thyrotropin Concentration in Human Serum by a Microplate Immunoassay

2.0 SUMMARY AND EXPLANATION OF THE TEST

Measurement of the serum concentration of thyrotropin (TSH), a glycoprotein with a molecular weight of 28,000 Daltons and secreted from the anterior pituitary, is generally regarded as the most sensitive indicator available for the diagnosis of primary and secondary (pituitary) hypothyroidism (1, 2). The structure of human TSH is similar to that of the pituitary and placental gonadotropins, consisting of an 89-amino acid subunit which is similar or identical between these hormones and a 115-amino acid β -subunit, which apparently confers hormonal specificity. The production of the β -subunit is separately regulated with apparent excess production of the β -subunit. The TSH molecule has a linear structure consisting of the protein core with carbohydrate side chains; the latter accounts for 16% of the molecular weight.

TSH measurements are equally useful in differentiating secondary and tertiary (hypothalamic) hypothyroidism from the primary thyroid disease. TSH release from the pituitary is regulated by thyrotropin releasing factor (TRH), which is secreted by the hypothalamus, and by direct action of T4 and triiodothyronine (T3). The thyroid hormones, at the pituitary, increase levels of T3 and T4, reduces the response of the pituitary to the stimulatory effects of TRH. In secondary and tertiary hypothyroidism, concentrations of T4 are usually low and TSH levels are generally low or normal. Either pituitary TSH deficiency (secondary hypothyroidism) or insufficiency of stimulation of the pituitary by TRH (tertiary hypothyroidism) causes this. The TRH stimulation test differentiates these conditions. In secondary hypothyroidism, TSH response to TRH is blunted while the normal or delayed response is obtained in tertiary hypothyroidism.

Further, the advent of immunoassay methods has provided this laboratory with sufficient sensitivity to enable the differentiation of hypothyroidism from euthyroid population and extending the usefulness of TSH measurements. This method is a second-generation assay, which provides the means for discrimination in the hypothyroid-euthyroid range. The functional sensitivity (<20% between assay CV) of the one-hour procedure is 0.195 μ U/ml while the two-hour procedure has a functional sensitivity of 0.005 μ U/ml (3).


In this method, TSH calibrator, patient specimen or control is first added to a streptavidin coated well. Biotinylated monoclonal and enzyme labeled antibodies are added and the reaction mixture. Reaction between the various TSH antibodies and native TSH forms a sandwich complex that binds with the streptavidin coated to the well.

After the completion of the required incubation period, the antibody bound enzyme-thyrotropin conjugate is separated from

the unbound enzyme-thyrotropin conjugate by aspiration or decantation. The activity of the enzyme present on the surface of the well is quantitated by reaction with a suitable substrate to produce color. The employment of several serum references of known thyrotropin levels permits construction of a dose response curve of activity and concentration. From comparison to the dose response curve, an unknown specimen's activity can be correlated with thyrotropin concentration.

3.0 PRINCIPLE

Immunoassay (TYPE 3): The essential reagents required for an immunoassay include high affinity and specificity antibodies (enzyme conjugated and immobilized), with different and distinct epitopes recognized in excess, and native antigen. In this procedure, the immobilization takes place during the assay at the surface of a microplate well through the interaction of streptavidin coated on the well and exogenously added biotinylated monoclonal anti-TSH antibody. Upon mixing monoclonal biotinylated antibody, the enzyme-labeled antibody and a serum containing the native antigen, reaction results between the native antigen and the antibodies, without competition or steric hindrance, to form a soluble sandwich complex. The interaction is illustrated by the following equation:



$Enz-Ab_{(m)}$ = Biotinylated Monoclonal Antibody (Excess Quantity)

$Ag_{(s)}$ = Native Antigen (Variable Quantity)


$Enz-Ab_{(m)}$ = Enzyme-Polyclonal Antibody (Excess Quantity)

$Enz-Ab_{(m)}-Ag_{(s)}-Ab_{(m)}$ = Antigen-Antibodies Sandwich Complex

k_a = Rate Constant of Association

k_d = Rate Constant of Dissociation

Simultaneously, the complex is deposited to the well through the high affinity reaction of streptavidin and biotinylated antibody. The interaction is illustrated below:



Immunized complex = sandwich complex bound to the well surface

After equilibrium is attained, the antibody-bound fraction is separated from unbound antigen by decantation or aspiration. The enzyme activity in the antibody-bound fraction is directly proportional to the native antigen concentration. By utilizing several different concentrations of known antigen values, a dose response curve can be generated from which the antigen concentration of an unknown can be ascertained.

4.0 REAGENTS

Materials Provided:

A. Thyrotropin Calibrators - 7mU/ml - Icons A-G

Seven (7) vials of references for TSH Antigen at levels of 0(A), 0.5(B), 2.5(C), 5.0(D), 10(E), 20(F), and 40(G) μ U/ml. Store at 2-8°C. A preservative has been added.

Note: The calibrators, human serum based, were calibrated using a reference preparation, which was assayed against the WHO 2nd IRP 60/58.

B. TSH Enzyme Reagent - 13mU/ml - Icon $\odot$

One (1) vial containing enzyme labeled affinity purified polyclonal goat antibody, biotinylated monoclonal mouse IgG in buffer, dye, and preservative. Store at 2-8°C.

C. Streptavidin Coated Plate - 96 wells - Icon $\odot$

One 96-well microplate coated with streptavidin and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

D. Wash Solution Concentrate - 20 ml - Icon $\odot$

One (1) vial containing a surfactant in buffered saline. A preservative has been added. Store at 2-8°C.

E. Substrate A - 7mU/ml - Icon $\odot$

One (1) bottle containing tetramethylbenzidine (TMB) in buffer. Store at 2-8°C.

F. Substrate B - 7mU/ml - Icon $\odot$

One (1) bottle containing hydrogen peroxide (H_2O_2) in buffer. Store at 2-8°C.

G. Stop Solution - 8mU/ml - Icon $\odot$

One (1) bottle containing a strong acid (1N HCl). Store at 2-8°C.

H. Product Instructions.

Note 1: Do not use reagents beyond the kit expiration date.

Note 2: Avoid extended exposure to heat and light. Opened reagents are stable for sixty (60) days when stored at 2-8°C. Kit and component stability are identified on the label.

Note 3: Above reagents are for a single 96-well microplate.

4.1 Required But Not Provided:

1. Pipette(s) capable of delivering 50 μ l and 100 μ l volumes with a precision of better than 1.5%.

2. Dispenser(s) for repetitive deliveries of 0.100ml and 0.350ml volumes with a precision of better than 1.5% (optional).

3. Microplate washer or a sequence bottle (optional).

4. Microplate Reader with 450nm and 620nm wavelength absorbance capability.

5. Absorbent Paper for blotting the microplate wells.

6. Plastic wrap or microplate cover for incubation steps.

7. Vacuum aspirator (optional) for wash steps.

8. Timer.

9. Storage container for storage of wash buffer.

10. Distilled or deionized water.

11. Quality Control Materials.

5.0 PRECAUTIONS

For In Vitro Diagnostic Use

Not for Internal or External Use in Humans or Animals

All products that contain human serum have been found to be cross-reactive for Hepatitis B Surface antigen, HIV 1/2 and HCV antibodies by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories", 2nd Edition, 1988, HHS.

Safe disposal of kit components must be according to local regulatory and statutory requirements.

6.0 SPECIMEN COLLECTION AND PREPARATION

The specimens shall be blood, serum in type, and the usual precautions in the collection of venipuncture samples should be observed. For accurate comparison to established normal values, a fasting morning serum sample should be obtained. The blood should be collected in a plain red-top venipuncture tube without additives or gel barrier. Allow the blood to clot. Centrifuge the specimen to separate the serum from the cells.

Samples may be refrigerated at 2-8°C for a maximum period of five (5) days. If the specimen(s) cannot be analyzed within this time, the sample(s) may be stored at temperatures of -20°C for up to 30 days. Avoid use of contaminated devices. Avoid repetitive freezing and thawing. When assayed in duplicate, 0.100 ml of the specimen is required.

7.0 QUALITY CONTROL

Each laboratory should assay controls at levels in the low, normal, and elevated range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Permanent statistical methods should be employed to ascertain trends. The individual laboratory should set acceptable assay performance limits. Other parameters that should be monitored include the 80, 50 and 20% intercepts of the

dose response curve for run-to-run reproducibility. In addition, maximum absorbance should be consistent with past experience. Significant deviation from established performance can indicate uninvolved change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations.

8.0 REAGENT PREPARATION:

1. Wash Buffer

Dilute contents of wash concentrate to 1000ml with distilled or deionized water in a suitable storage container. Store at 2-30°C for up to 60 days.

2. Working Substrate Solution

Pour the contents of the amber vial labeled Solution 'A' into the clear vial labeled Solution 'B'. Place the yellow cap on the clear vial for easy identification. Mix and label accordingly. Store at 2-8°C.

Note: Do not use the working substrate if it looks blue.

Note 2: Do not use reagents that are contaminated or have bacteria growth.

9.0 TEST PROCEDURE:

Before proceeding with the assay, bring all reagents, serum references and controls to room temperature (20-27°C).

Test Procedure should be performed by a skilled individual or trained professional

1. Format the microplate's wells for each serum reference, control and patient specimen to be assayed in duplicate.

Replace any unused microplate strips back into the aluminum bag, seal and store at 2-8°C.

2. Pipette 0.050 ml (50 μ l) of the appropriate serum reference, control or specimen into the assigned well.

3. Add 0.100 ml (100 μ l) of the TSH Enzyme Reagent to each well. It is very important to dispense all reagents close to the bottom of the coated well.

4. Swirl the microplate gently for 20-30 seconds to mix and cover.

5. Incubate 60 minutes at room temperature.

6. Discard the contents of the microplate by decantation or aspiration. If decanting tap and blot the plate dry with absorbent paper.

7. Add 350 μ l of wash buffer (see Reagent Preparation Section) decant (tap and blot) or aspirate. Repeat two (2) additional times for a total of three (3) washes. An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Discard the wash and repeat two (2) additional times.

8. Add 0.100 ml (100 μ l) of working substrate solution to all wells (see Reagent Preparation Section). Always add reagents in the same order to minimize reaction time differences between wells.

DO NOT SHAKE THE PLATE AFTER SUBSTRATE ADDITION

9. Incubate at room temperature for fifteen (15) minutes.

10. Add 0.050ml (50 μ l) of stop solution to each well and mix gently for 15-20 seconds. Always add reagents in the same order to minimize reaction time differences between wells.

11. Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm to minimize well imperfections) in a microplate reader. The results should be read within thirty (30) minutes of adding the stop solution.

** For better low-end sensitivity (< 0.5 μ U/ml), incubate 120 minutes at room temperature. The 40 μ U/ml calibrator should be excluded from absorbance over 3.0 units will be expanded. Follow the remaining steps.

Note: Discard samples reading over 40 μ U/ml by 1:5 and 1:10 with TSH $\odot$ Calibrator. Multiply the results by the dilution factor to obtain accurate results.

10.0 CALCULATION OF RESULTS


A dose response curve is used to ascertain the concentration of thyrotropin in unknown specimens.

- Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
- Plot the absorbance for each duplicate serum reference versus the corresponding TSH concentration in μ U/ml on linear graph paper (do not average the duplicates of the serum references before plotting).
- Draw the best-fit curve through the plotted points.
- To determine the concentration of TSH for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in μ U/ml) from the horizontal axis of the graph (this duplicates of the unknown may be averaged as shown in the following example, the average absorbance (0.775) intersects the dose response curve at (7.66 μ U/ml) TSH concentration (See Figure 1).

Note: Computer data reduction software designed for ELISA assay may also be used for the data reduction. If such software is utilized, the calibration of the software should be accurate.

EXAMPLE 1				
Sample ID	Well Number	Abs	Mean Abs	Value
Cal A	A1	0.018	0.019	0
	B1	0.021		
Cal B	B1	0.076	0.079	0.5
	D1	0.082		
Cal C	E1	0.302	0.298	2.5
	F1	0.293		
Cal D	G1	0.556	0.567	5.0
	H1	0.577		
Cal E	A2	0.926	0.921	10
	B2	0.916		
Cal F	C2	1.632	1.619	20
	D2	1.629		
Cal G	E2	2.694	2.671	40
	F2	2.647		
Control	G2	0.900	0.775	7.66
	H2	0.793		
Patient	A3	1.391	1.383	16.65
	B3	1.375		

*The data presented in Example 1 and Figure 1 are for illustration only and should not be used in lieu of a dose response curve prepared with each assay.



11.0 Q.C. PARAMETERS

In order for the assay results to be considered valid the following criteria should be met:

- The absorbance of calibrator 'G' (40 μ U/ml) should be ≥ 1.3 .
- Four out of six quality control points should be within the established ranges.

"NOT INTENDED FOR NEWBORN SCREENING"

12.0 RISK ANALYSIS

The MSDS and Risk Analysis Form for this product is available on request from Monobind Inc.

12.1 Assay Performance

- It is important that the time of reaction in each well is held constant to achieve reproducible results.
- Pipetting of samples should not extend beyond ten (10) minutes to avoid assay drift.
- Highly lipemic, hemolyzed or grossly contaminated specimens should not be used.
- If more than one (1) plate is used, it is recommended to repeat the dose response curve.
- The addition of substrate solution initiates a kinetic reaction, which is terminated by the addition of the stop solution. Therefore, the substrate and stop solution should be added in the same sequence to eliminate any time deviation during reaction.
- Plate readers measure vertically. Do not touch the bottom of the wells.
- Failure to remove adhering solution adequately in the aspiration or decantation wash steps may result in poor replication and spurious results.
- Use components from the same lot. No intermixing of reagents from different batches, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from Monobind's IFU may yield inaccurate results.
- Patients specimens with TSH concentrations over 40 μ U/ml may be diluted (1:5 or 1:10) with the 'G' calibrator and re-assayed. The sample's concentration is obtained by multiplying the result by the dilution factor.
- All applicable national standards, regulations and laws, including, but not limited to, good laboratory procedures, must be strictly followed to ensure compliance and proper device maintenance.
- It is important to calibrate all the equipment e.g. Pipettes, Readers, Washers and/or the automated instruments used with this device, and to perform routine preventative maintenance.
- Risk Analysis: as required by CE Mark IVD Directive 98/79/EC for this and other devices, by Monobind, can be requested via email from Monobind@monobind.com

12.2 Interpretation

- Measurement and Interpretation of results must be performed by a skilled individual or trained professional.
- Laboratory results alone are only one aspect for determining patient care and should not be the sole basis for therapy. Particularity if the results conflict with other determinants.
- For valid test results, adequate controls and other parameters must be within the listed ranges and assay requirements.
- Test results are altered, such as by mixing parts of different kits, which could produce false test results, or if results are incorrectly interpreted. Monobind shall have no liability.
- If computer control data reduction is used to interpret the results of the test, it is imperative that the predicted values for the calibrators fall within 10% of the assigned concentrations.
- Serum TSH concentration is dependent upon a multiplicity of factors, hypothalamic gland function, thyroid gland function, and the responsiveness of pituitary to TRH. Thus, thyrotropin concentration alone is not sufficient to assess clinical status.
- Serum TSH values may be elevated by pharmacological intervention. Dopaminergics, amiodarone, iodine, phenobarbital, and phenytoin have been reported to increase TSH levels.
- A decrease in thyrotropin values has been reported with the administration of propranolol, methimazole, dopamine and d-thyroxine (4).
- Genetic variations or degradation of intact TSH into subunits may affect the binding characteristics of the antibodies and influence the final result. Such samples normally exhibit different results among various assay systems due to the reactivity of the antibodies involved.

"NOT INTENDED FOR NEWBORN SCREENING"

13.0 EXPECTED RANGES OF VALUES

A study of euthyroid adult population was undertaken to determine expected values for the TSH AccuBind™ ELISA Test

System. The number and determined range are given in Table 1. A nonparametric method (95% Percentile Estimator) was used.

TABLE 1				
Expected Values for the TSH ELISA Test System (in μ U/ml)				
Number	5%	2.5%	Percentile	75%
Low Normal	0.59	Low Range	0.28 - 0.53	
High Normal	6.16	High Range	5.60 - 6.82	

It is important to keep in mind that establishment of a range of values is not to be expected to be found by a given method for a population of "normal" persons is dependent upon a multiplicity of factors: the specificity of the method, the population tested and the precision and the correlation coefficient of the assay. For these reasons each laboratory should depend upon the range of expected values established by the manufacturer only until an in-house range can be determined by the analysis using the method with the population indigenous to the area in which the laboratory is located.

14.0 PERFORMANCE CHARACTERISTICS

14.1 Precision

The within and between assay precisions of the TSH AccuBind™ test system were determined by analyses on three different levels of pool control sera. The number (N), mean (X), value standard deviation (s) and coefficient of variation (C.V.) for each of these control sera are presented in Table 2 and Table 3.

Within Assay Precision (Values in μ U/ml)				
Sample	N	X	s	C.V.
Pool 1	24	0.37	0.03	8.1%
Pool 2	24	6.75	0.43	6.4%
Pool 3	24	25.33	1.94	6.9%

Between Assay Precision (Values in μ U/ml)				
Sample	N	X	s	C.V.
Pool 1	10	0.43	0.04	9.3%
Pool 2	10	6.80	0.54	7.9%
Pool 3	10	28.40	1.07	5.9%

*As measured in ten experiments in duplicate over seven days.

14.2 Sensitivity

The sensitivity (detection limit) was ascertained by determining the variability of the 0 μ U/ml serum calibrator and using the 2 σ (95% certainty) statistic to calculate the minimum dose. For 1 hr incubation - 0.075 μ U/ml. For 2 hr incubation - 0.027 μ U/ml.

14.3 Accuracy

The TSH AccuBind™ ELISA test system was compared with a reference immunochemiluminometric assay. Biological specimens from hypothyroid, euthyroid and hyperthyroid populations were used. The values ranged from 0.01 μ U/ml - 67 μ U/ml. The total number of such specimens was 241. The least square regression equation and the correlation coefficient were computed for the TSH AccuBind™ ELISA method in comparison with the reference method. The data obtained is depicted in Table 4.

TABLE 4		
Method	Mean (X)	Least Square Regression Analysis
This Method	4.54	$y = 0.47 + 0.968(x)$
Reference	4.21	Correlation Coefficient

Only slight amounts of bias between the TSH AccuBind™ ELISA method and the reference method are indicated by the closeness of the mean values. The least square regression equation and correlation coefficient indicates excellent method agreement.

14.4 Specificity

The cross-reactivity of the TSH AccuBind™ ELISA test system to selected substances was evaluated by adding the interfering substance to a serum matrix at various concentrations. The cross-reactivity was calculated by dividing a ratio between dose of interfering substance to dose of thyrotropin needed to produce the same absorbance.

Substance	Cross Reactivity	Concentration
Thyrotropin (hTSH)	1.0000	
Follicleotropin (hFSH)	< 0.0001	1000ng/ml
Lutealotropin Hormone (hLH)	< 0.0001	1000ng/ml
Chorionic	< 0.0001	1000ng/ml
Gonadotropin(hCG)		

СИСТЕМА СЕРТИФИКАЦИИ ГОСТ Р
ФЕДЕРАЛЬНОЕ АГЕНТСТВО ПО ТЕХНИЧЕСКОМУ РЕГУЛИРОВАНИЮ И МЕТРОЛОГИИ



СЕРТИФИКАТ СООТВЕТСТВИЯ

№ РОСС RU.ИМ02.Н18061

Срок действия с 24.06.2019г.

по 24.06.2020г.

№ 0405450

ОРГАН ПО СЕРТИФИКАЦИИ № RA.RU.11ИМ02

МЕДИЦИНСКИХ ИЗДЕЛИЙ АНО «ВНИИИМТ»

129301, г. Москва, ул. Касаткина, д.3 тел. (495) 683-97-92, факс (499)187-89-54,
e-mail: im02@bk.ru

ПРОДУКЦИЯ Индикаторы паровой стерилизации химические
одноразовые «ИНТЕСТ-П-«ВИНАР» по ТУ 9398-041-11764404-2003

Серийный выпуск.

код ОК

034-2014 (КПЕС 2008)
32.50.50.190

СООТВЕТСТВУЕТ ТРЕБОВАНИЯМ НОРМАТИВНЫХ ДОКУМЕНТОВ

ГОСТ ISO 11140-1-2011 (класс 4),

ГОСТ Р 50444-92 (разделы 3,5,8)

код ТН ВЭД

3822 00 000 0

ИЗГОТОВИТЕЛЬ Общество с ограниченной ответственностью «Научно-производственная фирма «ВИНАР» (ООО «НПФ «ВИНАР»), Россия, 105094, г. Москва, Госпитальный вал, д.5, стр.7А, пом. VIII ИНН 5023001024
Место производства-141009, Московская обл., г.Мытищи, ул.Колонцова,д.17/2

СЕРТИФИКАТ ВЫДАН Общество с ограниченной ответственностью «Научно-производственная фирма «ВИНАР» (ООО «НПФ «ВИНАР»), Россия, 105094, г. Москва, Госпитальный вал, д.5, стр.7А, пом. VIII
тел./факс (495) 988-76-67

НА ОСНОВАНИИ протокола испытаний № 16-853 от 20.06.2016г. ИЦ МИ АНО «ВНИИИМТ» (№ RA.RU.21ИМ04).

Регистрационные удостоверения № РЗН 2013/39 от 08.02.2013г. Федеральной службы по надзору в сфере здравоохранения (РОСЗДРАВНАДЗОР)

ДОПОЛНИТЕЛЬНАЯ ИНФОРМАЦИЯ

Маркирование продукции производится знаком соответствия Системы сертификации ГОСТ Р при добровольной сертификации продукции



Руководитель органа

подпись

Эксперт

подпись

Е. И. Полянская

инициалы, фамилия

В.В. Русова

инициалы, фамилия

Сертификат не применяется при обязательной сертификации



TÜVRheinland

EC Certificate

Directive 98/79/EC Annex IV, excluding Sections 4 and 6

Full Quality Assurance System

In Vitro Diagnostic Medical Devices

Registration No.: HL 60119814 0001

Report No.: 21265422 001

Manufacturer:

Macherey-Nagel GmbH & Co. KG
Neumann-Neander-Str. 6-8
52355 Düren
Deutschland

Products:

Products for self-testing

(see Attachment for products and sites included)

Replaces Certificate, Registration No.: HL 60076697 0001

Expiry Date:

2022-05-28

The Notified Body hereby declares that the requirements of Annex IV, excluding section 4 and 6 of the directive 98/79/EC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex IV, section 5 of the aforementioned directive. For placing on the market of List A devices covered by this certificate an EC design-examination certificate according to Annex IV, section 4 and a verification of manufactured products according to section 6 is required.

Effective Date:

2017-05-29

Date:

2017-05-29



Notified Body

Dipl.-Ing. Sven Hofmann

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 98/79/EC concerning in vitro diagnostic medical devices with the identification number 0197.



Notified Body

Dipl.-Ing. Sven Hofmann

Date: 2017-05-29



TÜVRheinland

Doc. 1/1, Rev. 0

TÜV Rheinland

LGA Products GmbH

Tillystraße 2, 90431 Nürnberg

Attachment to

Certificate

Registration No.: HL 60119814 0001

Report No.: 21265422 001

Manufacturer:

Macherey-Nagel GmbH & Co. KG
Neumann-Neander-Str. 6-8
52355 Düren
Deutschland

Products for self-testing:

- Single and multi-parameter disposable test strips for urine analysis
- Indicator test strips and papers for measurement of pH in urine

Additional site for warehousing and logistics:

Bahnstr. 120
52355 Düren, Germany